The attached table will provide monitoring organizations a running tally of planned edits to the 2017 QA Handbook document. This list should not be construed as comments that have been received but as changes that will be made to the current document (EPA-454/B-17-001).

2017 QA Handbook Planned Edits			
Section/subsection	Change Type	Current Language (2017 Document)	Change
Sec 5 Statute of Limitations	Editorial	All original documents an records be kept for the statute of limitation.	All original documents and records will be kept for the statute of limitation
Section 7/ 7.2.2	Additon	Analyzers must be operated within the range for which they were designated, in order for the data produced by the analyzers to be considered FRM/FEM.	The following will be added after sentence in current language column. "Data from monitors should be considered valid that is outside the 20-30 <sup>o</sup> C shelter temperature guidance if the federal reference method (FRM) or federal equivalent method (FEM) monitor has been approved at a wider temperature range"
Section 17	Addition	Not applicable	Suggested including guidance on warning limits vs. control limits for acceptance criteria in verification/validation section.