

# DATA VALIDATION REPORT

# Gold King Mine Release Incident

# SAMPLE DELIVERY GROUP: 680-117013-7

Prepared by

MEC<sup>x</sup> 12269 East Vassar Drive Aurora, CO 80014



## I. INTRODUCTION

Task Order Title: Project No.:	Gold King Mine Release Incident 20408.012.001.0274.00 20408.012.001.0267.00
Sample Delivery Group:	680-117013-7
EPA Project Manager:	Steve Way
Weston Project Manager:	Dave Robinson
TDD No.:	0001/1508-04
Matrix:	Water
QC Level:	Stage 2A
No. of Samples:	1
No. of Reanalyses/Dilutions:	0
Laboratory:	TestAmerica- St. Louis

### Table 1. Sample Identification


Location ID	Lab Sample Name	Matrix Type	Collection Date	Method
CC06_09212015_1300	680-117013-3	Water	9/21/15 1:00 PM	6020A

### II. Sample Management

Anomalies regarding sample management are noted below. The sample was received at the laboratories below the temperature limits of  $4^{\circ}C \pm 2^{\circ}C$ , at 0.6 at TestAmerica-Savannah and at 1.3°C at TestAmerica-St. Louis. As the sample was not noted to be frozen or damaged, no qualifications were required. The samples were received intact, on ice, and properly preserved. The chains-of-custody (COCs) were appropriately signed and dated by field and laboratory personnel. The presence or absence of custody seals on the cooler was not specifically noted.

The following issues were noted:

- The COCs listing the sample requested a number of other analyses. Results of these analyses were reported in related SDGs 117013-1, 117013-3, 117013-4, 117013-6, and 117013-5.
- Some corrections made to one COC were made by overwriting the original entry. These corrections were not initialed or dated.
- The COCs did not list CLP sample IDs, and none were provided. The laboratory logged the samples per the location IDs on the COCs.
- The presence or absence of sample tags was not noted in the case narrative, and sample tags were not listed on the COCs.





Qualifier	Organics	Inorganics
U	The analyte was analyzed for, but was not detected above the reported sample quantitation limit. The associated value is the quantitation limit or the estimated detection limit for dioxins or PCB congeners.	The material was analyzed for, but was not detected above the level of the associated value. The associated value is either the sample quantitation limit or the sample detection limit. The associated value is the sample detection limit or the quantitation limit for perchlorate only.
UB	The analyte was detected in the sample and in either the associated laboratory blank or field blank. If detected below the reporting limit (RL) the analyte result was reported as non- detected at the RL due to blank contamination. If detected above the RL, the analyte result was reported as non-detected at the reported result due to blank contamination.	The analyte was detected in the sample and in either the associated laboratory blank or field blank. If detected below the reporting limit (RL) the analyte result was reported as non-detected at the RL due to blank contamination. If detected above the RL, the analyte result was reported as non-detected at the reported result due to blank contamination.
J	The analyte was positively identified; the associated numerical value is the approximate concentration of the analyte in the sample.	The analyte was positively identified; the associated numerical value is the approximate concentration of the analyte in the sample.
J+	Not applicable	The analyte was positively identified; the associated numerical value is the approximate concentration of the analyte in the sample, and may have a potential positive bias.
J-	Not applicable	The analyte was positively identified; the associated numerical value is the approximate concentration of the analyte in the sample, and may have a potential negative bias.



#### DATA VALIDATION REPORT

Qualifier	Organics	Inorganics
UJ	The analyte was not deemed above the reported sample quantitation limit. However, the reported quantitation limit is approximate and may or may not represent the actual limit of quantitation necessary to accurately and precisely measure the analyte in the sample.	The material was analyzed for, but was not detected. The associated value is an estimate and may be inaccurate or imprecise.
UJB	The analyte was detected in the sample and in either the associated laboratory blank or field blank; the analyte result was reported as non-detected at either the RL or the reported result. The reported quantitation limit is approximate and may or may not represent the actual limit of quantitation necessary to accurately and precisely measure the analyte in the sample.	The analyte was detected in the sample and in either the associated laboratory blank or field blank; the analyte result was reported as non- detected at either the RL or the reported result. The reported quantitation limit is approximate and may or may not represent the actual limit of quantitation necessary to accurately and precisely measure the analyte in the sample.
Ν	The analysis indicates the presence of an analyte for which there is presumptive evidence to make a "tentative identification."	Not applicable.
NJ	The analysis indicates the presence of an analyte that has been "tentatively identified" and the associated numerical value represents its approximate concentration.	Not applicable.
R	The data are unusable. The sample results are rejected due to serious deficiencies in the ability to analyze the sample and to meet quality control criteria. The presence or absence of the analyte cannot be verified.	The data are unusable. The sample results are rejected due to serious deficiencies in the ability to analyze the sample and to meet quality control criteria. The presence or absence of the analyte cannot be verified.



# **Qualification Code Reference Table**

Qualifier	Organics	Inorganics			
Н	Holding times were exceeded.	Holding times were exceeded.			
S	Surrogate recovery was outside QC limits.	The sequence or number of standards used for the calibration was incorrect			
С	Calibration %RSD or %D was noncompliant.	Correlation coefficient is <0.995 or calibration was noncompliant.			
R	Calibration RRF was <0.05.	%R for calibration is not within control limits.			
В	Presumed contamination as indicated by the preparation (method) blank results.	Presumed contamination as indicated by the preparation (method) or calibration blank results.			
L	Laboratory Blank Spike/Blank Spike Duplicate %R was not within control limits.	Laboratory Control Sample %R was not within control limits.			
L1	LCS/LCSD RPD was outside control limits.	LCS/LCSD RPD was outside control limits.			
Q	MS/MSD recovery was poor.	MS recovery was poor.			
Q1	MS/MSD RPD was outside control limits.	MS/MSD RPD was outside control limits.			
Е	Not applicable.	Duplicates showed poor agreement.			
Ι	Internal standard performance was unsatisfactory.	ICP ICS results were unsatisfactory.			
А	Not applicable.	ICP Serial Dilution %D were not within control limits.			
М	Tuning (BFB or DFTPP) was noncompliant.	ICPMS tune was not compliant.			
Т	Presumed contamination as indicated by the trip blank results.	Not applicable.			
+	False positive – reported compound was not present.	Not applicable.			
-	False negative – compound was present but not reported.	Not applicable.			
F	Presumed contamination as indicated by the FB or ER results.	Presumed contamination as indicated by the FB or ER results.			
F1	Field duplicate results were outside the control limit.	Field duplicate results were outside the control limit.			
\$	Reported result or other information was incorrect.	Reported result or other information was incorrect.			





#### DATA VALIDATION REPORT

Qualifier	Organics	Inorganics		
?	TIC identity or reported retention time has been changed.	Not applicable.		
D	The analysis with this flag should not be used because another more technically sound analysis is available.	The analysis with this flag should not be used because another more technically sound analysis is available.		
Р	Instrument performance for pesticides was poor.	Post Digestion Spike recovery was not within control limits.		
*11, *111	Unusual problems found with the data that have been described in Section II, "Sample Management," or Section III, "Method Analyses." The number following the asterisk (*) will indicate the report section where a description of the problem can be found.	Unusual problems found with the data that have been described in Section II, "Sample Management," or Section III, "Method Analyses." The number following the asterisk (*) will indicate the report section where a description of the problem can be found.		



# III. Method Analyses

## A. Contract Laboratory Program Statement of Work for Inorganic Superfund Method 6020A—Metals and Mercury

Reviewed By: P. Meeks Date Reviewed: October 1, 2015

The sample listed in Table 1 for this analysis was validated based on the guidelines outlined in the Sampling and Analysis Plan/Quality Assurance Project Plan for Gold King Mine Release, Silverton, San Juan County, Colorado (2015), United States Environmental Protection Agency Contract Laboratory Program Statement of Work for Inorganic Superfund Methods, EPA Method 6020A, and the National Functional Guidelines for Inorganic Superfund Data Review (2010).

- Holding Times: The analytical holding time, six months, was met.
- Analytical Method Blanks: Uranium was not detected in the method blank.
- Laboratory Control Samples (LCS): The recovery was within method control limits of 80-120%.
- Laboratory Duplicates: A laboratory duplicate analysis was performed on the sample in this SDG. The parent sample result was less than 5× the reporting limit (RL); therefore, the QAPP relative percent difference (RPD) control limit was not applied. Based on professional judgment, the reasonable control limit of ±RL was utilized. The results were within this control limit.
- Matrix Spike/Matrix Spike Duplicate (MS/MSD): MS/MSD analyses were performed on the sample in this SDG. The recoveries were within the method control limits of 75-125% and the RPD was within the QAPP control limit of ≤20%.
- Post Digestion Spike (PDS): There were no PDS analyses performed on the sample in this SDG.
- Serial Dilution: There were no serial dilution analyses performed in this SDG.
- Field QC Samples: Field QC samples were evaluated, and if necessary, qualified based on method blanks and other laboratory QC results affecting the usability of the field QC data. Any remaining detects were used to evaluate the associated site samples. Following are findings associated with field QC samples:



- Field Blanks and Equipment Rinsates: No field blank or equipment rinsate samples were identified in this SDG.
- Field Duplicates: There were no field duplicate samples identified in this SDG.

# Validated Sample Result Forms: 680-117013-7

Analysis Method 6020A									
Sample Name CC06_09212015_1300				Matrix Type	: Water				
Lab Sample N	ame: 680-117013	-3 Sam	ple Date:	9/21/2015 1:00:00	PM				
Analyte	Total/Dissolved	CAS No	Result Value	Reporting Limit	MDL	Result Units	Lab Qualifier	Validation Qualifier	Validation Notes
Uranium	Т	7440-61-1	8.8	2.5	0.58	ug/L			