



U.S. EPA Ambient Air Monitoring Protocol Gas Verification Program

Annual Report CY 2020

EPA-454/R-21-006
July 2021

U.S. EPA Ambient Air Monitoring Protocol Gas Verification Program
Annual Report CY 2020

U.S. Environmental Protection Agency
Office of Air Quality Planning and Standards
Air Quality Assessment Division
Research Triangle Park, NC

Table of Contents

Acknowledgements	5
Acronyms and Abbreviations	6
1.0 Introduction	7
2.0 Implementation Summary	9
3.0 Survey and Verification Results	12
4.0 Summary and Conclusions	16
Appendix A QA Reports from Measurement Data Worksheets for 2020	19
Tables	
TABLE 1. RAVL VERIFICATION DATES	10
TABLE 2. GAS STANDARDS SENT TO RAVLS	13
TABLE 3. MQOs FOR THE AA-PGVP	14
TABLE 4. AA-PGVP CO AND SO ₂ VERIFICATIONS	15
TABLE 5. AA-PGVP NO AND NO _x VERIFICATIONS	15
Figures	
FIGURE 1. AA-PGVP FLOW CHART	11
Figure 2 Annual Survey	12

Acknowledgements

The following individuals and organizations are acknowledged for their contributions to this project:

US EPA, Office of Air Quality Planning and Standards
Douglas Jager

US EPA, Office of Research and Development
Bob Wright

US EPA Region 7

Michael Davis

Thien Bui

James Regehr

Monitoring Organizations

EPA acknowledges the monitoring organizations that supported the AA-PGVP Annual Survey.

They include:

Alaska Department of Environmental Conservation
Canton City Health Department Air Pollution Control
Cherokee Nation, Oklahoma
Clark County, NV DAQEM
FDEP Ambient Monitoring Section
Forsyth County Environmental Affairs Department
Idaho Department of Health & Welfare-Env. Div.
Indiana Depart of Environ Management
Kentucky Division for Air Quality
Louisville, KY Metro Air Pollution Control District
Maricopa County Air Quality
Maryland Department of the Environment
Michigan Dept of Environment, Air Quality Division
Minnesota Pollution Control Agency, Div. of Air Quality

Missouri Laboratory Services Program
MT Dept of Environmental Quality, Air Quality Div.
New Jersey State Department of Environmental Protection
North Carolina Dept of Environmental Quality
Ohio EPA, Central District Office
Ohio EPA, Northeast District Office
Ohio EPA, Southeast District Office
Oklahoma Dept. of Environmental Quality, Air Qual. Div.
Oregon Department of Environmental Quality
South Carolina Department Health & Env. Control
South Coast Air Quality Management District
Washington State Department of Ecology
Wisconsin Dept of Natural Resources

Acronyms and Abbreviations

AA-PGVP	Ambient Air Protocol Gas Verification Program
AMTIC	Ambient Monitoring Technology Information Center
AQS	Air Quality System
CFR	Code of Federal Regulations
CO	Carbon Monoxide
COC	Chain of Custody
CONC	Concentration
COVID-19	Coronavirus Disease 2019
EPA	Environmental Protection Agency
GMIS	Gas Manufacturer's Internal Standard
ID	Identification
MFC	Mass Flow Controller
MQO	Measurement Quality Objective
NIST	National Institute of Standards and Technology
NO ₂	Nitrogen Dioxide
NO _x	Nitrogen Oxides
NTRM	NIST Traceable Reference Material
NVLAP	National Voluntary Laboratory Accreditation Program
OAQPS	Office of Air Quality Planning and Standards
OAR	Office of Air and Radiation
OIG	Office of the Inspector General
ORD	Office of Research and Development
PQAO	Primary Quality Assurance Organization
PSI	Pounds per Square Inch
PSIG	Pounds per Square Inch Gauge
QA	Quality Assurance
QAPP	Quality Assurance Project Plan
QC	Quality Control
QTR	Quarter
RAVL	Regional Analytical Verification Laboratory
RD	Relative Difference
RPD	Relative Percent Difference
SO ₂	Sulfur Dioxide
SOP	Standard Operating Procedure
SRM	Standard Reference Material
URL	Upper Range Limit

1.0 Introduction

Background and Program Goals

The basic principles of the U.S. Environmental Protection Agency's (EPA) *Traceability Protocol for the Assay and Certification of Gaseous Calibration Standards* (EPA, 2012)¹ were developed jointly by EPA, the National Bureau of Standards (now National Institute of Standards and Technology [NIST]), and specialty gas producers over 40 years ago. At the time, commercially prepared calibration gases were perceived as being too inaccurate and too unstable for use in calibrations and audits of continuous source emission monitors and ambient air quality monitors². The protocol was developed to improve the quality of the gases by establishing their traceability to NIST Standard Reference Materials (SRMs) and to provide reasonably priced products. This protocol established the gas metrological procedures for measurement and certification of these calibration gases for EPA's Acid Rain Program under 40 Code of Federal Regulations (CFR) Part 75, for the Ambient Air Quality Monitoring Program under 40 CFR Part 58, and for the Source Testing Program under 40 CFR Parts 60, 61, and 68. EPA required monitoring organizations implementing these programs ("the regulated community") to use EPA Protocol Gases as their calibration gases. EPA revised the protocol to establish detailed statistical procedures for estimating the total uncertainty of these gases. EPA's Acid Rain Program developed acceptance criteria for the uncertainty estimate³.

Specialty gas producers prepare and analyze EPA Protocol Gases without direct governmental oversight. In the 1980s and 1990s, EPA conducted a series of EPA-funded accuracy assessments of EPA Protocol Gases sold by producers. The intent of these audits was to:

- increase the acceptance and use of EPA Protocol Gases as calibration gases;
- provide a quality assurance (QA) check for the producers of these gases; and
- help users identify producers who can consistently provide accurately certified gases.

Either directly or through third parties, EPA procured EPA Protocol Gases from the producers, assessed the accuracy of the gases' certified concentrations through independent analyses, and inspected the

¹ EPA-600/R-12/531

² Decker, C.E. et al., 1981. "Analysis of Commercial Cylinder Gases of Nitric Oxide, Sulfur Dioxide, and Carbon Monoxide at Source Concentrations," *Proceedings of the APCA Specialty Conference on Continuous Emission Monitoring-Design, Operation, and Experience*, APCA Publication No. SP-43.

³ "Continuous Emission Monitoring," *Code of Federal Regulations*, Title 40, Part 75

accompanying certificates of analysis for completeness and accuracy. The producers were not aware that EPA had procured the gases for these audits.

The accuracy of the EPA Protocol Gases' certified concentrations was assessed using SRMs as the analytical reference standards. If the difference between the audit's measured concentration and the producer's certified concentration was more than ± 2.0 percent or if the documentation was incomplete or inaccurate, EPA notified the producer to resolve and correct the problem. The results of the accuracy assessments were published in peer-reviewed journals and were posted on EPA's Technology Transfer Network website. The accuracy assessments were discontinued in 1998.

In 2009, the Office of the Inspector General (OIG) published the report *EPA Needs an Oversight Program for Protocol Gases*⁴. One of the report's findings suggested that EPA "does not have reasonable assurance that the gases that are used to calibrate emissions monitors for the Acid Rain Program and continuous ambient monitors for the nation's air monitoring network are accurate". OIG recommended that the Office of Air and Radiation (OAR) implement oversight programs to assure the quality of the EPA Protocol Gases that are used to calibrate these monitors. It also recommended that EPA's Office of Research and Development (ORD) update and maintain the document *Traceability Protocol for Assay and Certification of Gaseous Calibration Standards* to ensure that the monitoring programs' objectives are met.

In order to address the OIG findings for ambient air monitoring, the Office of Air Quality Planning and Standards (OAQPS), in cooperation with two EPA Regional Offices, developed an Ambient Air Protocol Gas Verification Program (AA-PGVP). The program established two gas metrology laboratories to verify the certified concentrations of EPA Protocol Gases used to calibrate ambient air quality monitors. The program is expected to ensure that producers selling EPA Protocol Gases participate in the AA-PGVP and provides end users with information about participating producers and verification results.

The EPA Ambient Air Quality Monitoring Program's QA requirements, as described in Section 2.6.1 of 40 CFR Part 58, Appendix A, include:

Gaseous pollutant concentration standards (permeation devices or cylinders of compressed gas) used to obtain test concentrations for CO, SO₂, NO, and NO₂ must be traceable to either a National Institute of Standards and Technology (NIST) Traceable Reference Material (NTRM) or a NIST-certified Gas Manufacturer's Internal Standard (GMIS), certified in accordance with one of the procedures given in reference 4 of this appendix. Vendors advertising certification with the procedures provided in reference 4 of this appendix and distributing gases as "EPA Protocol Gas" for ambient air monitoring purposes must participate in the EPA Ambient Air Protocol Gas Verification Program or not use "EPA" in any form of advertising. Monitoring organizations must provide information to the EPA on the gas producers they use on an annual basis and those PQAOs purchasing standards will be obligated, at the request of the EPA, to participate in the program at least once every 5 years by sending a new unused standard to a designated verification laboratory.

⁴ <https://www.epa.gov/office-inspector-general/report-epa-needs-oversight-program-protocol-gases-09-P-0235.pdf>

This program is considered a verification program because its current level of evaluation does not allow for a large enough sample of EPA Protocol Gases from any one specialty gas producer to yield a statistically rigorous assessment of the accuracy of the producer's gases. It will not provide end users with a scientifically defensible estimate of whether gases of acceptable quality can be purchased from a specific producer. Rather, the results provide information to end users that the specialty gas producer is participating in the program and with information that may be helpful when selecting a producer.

Purpose of This Document

The purpose of this document is to report the activities that occurred in 2020 and provide the results of the verifications performed.

This document will not explain the implementation of the AA-PGVP, the quality system or the verification procedure. That information has been documented in the Implementation Plan, Quality Assurance Project Plan (QAPP) and standard operating procedures (SOPs) that can be found on the AA-PGVP Web Page on the Ambient Monitoring Technology Information Center (AMTIC)⁵.

2.0 Implementation Summary

Since the program implementation started in 2010, when most of the initial preparation work took place, no major “new” implementation activities took place. However, EPA regional realignments and aging infrastructure reduced the capabilities of this program. Due to these constraints, the EPA Region 2 Regional Analytical Verification Laboratory (RAVL) ceased its active participation in the AA-PGVP in calendar year 2019. During 2020 the AA-PGVP began transitioning Region 2 operations to the Region 4 laboratory. However, during 2020 the AA-PGVP continued to operate with only the Region 7 RAVL. Operations with only a single RAVL resulted in the AA-PGVP unable to swap internal quality control samples and cylinders needing confirmatory assay between two independent RAVLs. During 2020 EPA began reengineering the AA-PGVP and as part of that process EPA began assisting the EPA Region 4 laboratory to serve as a replacement RAVL.

The following provides a brief explanation of the 2020 implementation process.

Producer Information Data Collection – In 2010 EPA sent out an Excel spreadsheet to each monitoring organization to obtain information on the gas standard producers being used by the monitoring organization and to determine their interest in participating in the program. In 2011, EPA began work with Research Triangle Institute to develop a web-based survey that one point of contact for each monitoring organization

⁵ www.epa.gov/amtic/ambient-air-protocol-gas-verification-program

could access. The intent was to make recording and evaluation of the survey information easier for the monitoring organizations and EPA. This contracted survey work has since migrated to Battelle. Based on the information obtained from monitoring organization surveys, EPA would develop a list of the specialty gas producers being used by the monitoring organizations. From this list, EPA would attempt to perform representative sampling of the standards from protocol gas production facilities by identifying regulatory monitoring agencies that use standards from each of these producers. However, only 27 agencies participated in the survey for 2020. With only limited survey results, a systematic selection of producers could not be performed. The AA-PGVP performed assays on all cylinders submitted by regulatory monitoring agencies. Representative sampling was not attempted for CY-2020. OAQPS is in the process of developing an Air Quality System (AQS) database solution to upgrade and replace the specialty gas usage information that is currently acquired through the contractor based annual questionnaire.

AA-PGVP Verification Dates – OAQPS worked with the Region 7 Regional Analytical Verification Laboratory (RAVL) to establish verification dates as indicated in Table 1. However, due to the COVID-19 pandemic the EPA Region 7 laboratory was shutdown during quarters 1 and 2 of calendar year 2020.

Table 1. RAVL Verification Dates

Quarter	Region 7	
	Cylinder Receipt	Analysis
1*	No later than Mar 20	Mar 30 – April 10
2*	No later than June 17	June 22 – July 3
3	No later than Sept 4	Sept 14 – Sept 25
4	No later than Nov 27	Dec 7 – Dec 18
Open House	December 16, 2020	

TABLE 1. RAVL VERIFICATION DATES

Note: * No assays were performed in Quarters 1 & 2 due to the COVID-19 pandemic.

Table 1 RAVL Open House – During Open House the RAVL allows specialty gas producers to visit and ask questions regarding the laboratory processes and operations. During 2020 no specialty gas producers visited the Region 7 RAVL.

Flow of the AA-PGVP

Figure 1 provides a flow diagram of the implementation activities of the AA-PGVP. The major activities in these steps are explained below. More details of these steps are found in the AA-PGVP Implementation Plan, QAPP and SOPs.




FIGURE 1. AA-PGVP FLOW CHART

1. EPA sends e-mails to the monitoring organization's points of contact to complete the AA-PGVP Survey. EPA compiles information on specialty gas producers and the monitoring organizations that plan to participate. EPA tries to schedule the monitoring organization in an appropriate verification quarter based on delivery of standards from the specialty gas producer.
2. The monitoring organizations order gas standards from specialty gas producers during the normal course of business. If EPA cannot get a cylinder from the monitoring organization, and that producer is being used, EPA will invite the producer to send a cylinder directly to an RAVL.
3. The monitoring organizations send a new/unused standard, specialty gas certification and chain of custody form to the RAVLs.
4. The RAVLS analyze the cylinders and provide the validated results to OAQPS and the monitoring organizations.
5. OAQPS reviews the data and sends verification results to the specialty gas vendors.
6. At the end of the year, OAQPS compiles final results into a report, sends the report out to the specialty gas vendors and posts it on the AA-PGVP AMTIC web page.

3.0 Survey and Verification Results

Monitoring Organization Survey

Based upon the maximum capability of 40 gas cylinders per RAVL per year, the AA-PGVP selection goal, in the following order, is:

- 1) At least one gas standard from every specialty gas producer being used by the monitoring community.
- 2) If all specialty gas producers have been assessed at least once, then attempt to verify three standards per specialty gas producer.
- 3) If all specialty gas producers have been assessed three times, weigh additional verifications by producer market share in the ambient air monitoring community.

In order to assess which specialty gas producers are used by the monitoring organizations, EPA annually uses a web-based survey that each monitoring organization completes. Since 2016, EPA regulations found in 40 CFR Part 58 Appendix A 2.6.1 require monitoring organizations to annually provide this information. However, as can be seen from Figure 2, participation in the annual survey has not improved since the 2016 monitoring rule revisions.




FIGURE 2 ANNUAL SURVEY

Table 2. Gas Standards Sent to RAVLs in Calendar Year 2020

Qtr	Cylinder ID	Pollutant	Lab	Producer	Facility	Agency
3	CC93605	CO	7	Air Gas	Chicago, IL	Missouri Laboratory Services Program
4	FF55927	SO ₂	7	Air Gas	Chicago, IL	Kentucky Div. for Air Quality
4	CC178285	NO, NO _x	7	Air Gas	Chicago, IL	Polk County Air Quality
4	FF36562	NO, NO _x	7	Air Gas	Chicago, IL	Kentucky Div. for Air Quality
4	BAL4148	CO, NO, NO _x	7	Air Gas	Los Angeles, CA	South Coast Air Quality Management District
3	SD15080	NO, NO _x ^Ω	7	Matheson	Waverly, TN	Kansas Department of Health and Environment
3	DT0011078*	NO, NO _x ^Ω	7	PraxAir	Los Angeles, CA	(Sent by Producer)
3	DT0029223*	NO, NO _x ^Ω	7	PraxAir	Toledo, OH	(Sent by Producer)
3	LL127859	NO, NO _x ^Ω	7	PraxAir	Toledo, OH	State Hygienic Lab (Univ of Iowa)
4	LL197504	SO ₂	7	Tier 5	Indianapolis, IN	Kentucky Div. for Air Quality
4	LL197501	NO, NO _x ^Ω	7	Tier 5	Indianapolis, IN	Kentucky Div. for Air Quality
3	CC460552**	NO, NO _x	7	WestAir	Anaheim, CA	San Diego Air Pollution Control District

TABLE 2. GAS STANDARDS SENT TO RAVLS

Notes: * Direct submission by producer, assay results not blind to producer

** Cylinder opened prior to submission to EPA

Ω NO_x concentration provided by Producer as “informational only”; concentration not certified by Producer.

Verification Results

As indicated in 40 CFR Part 75 Appendix A, EPA Protocol Gases must have a certified uncertainty (95 percent confidence interval) that must not be greater than plus or minus 2 percent ($\pm 2.0\%$) of the certified concentration (tag value) of the gas mixture. This acceptance criterion is for the Acid Rain Program. The AA-PGVP adopted the criteria as its data quality objective and developed a quality system to allow the RAVLs to determine whether or not an individual protocol gas standard concentration was within $\pm 2\%$ of the certified value. The Ambient Air Program has never identified an acceptance criterion for the protocol gases. Since the AA-PGVP has not been established to provide a statistically rigorous assessment of any specialty gas producer, the RAVLs report all valid results as analyzed, but it is suggested that any difference greater than $\pm 4\%$ is cause for concern. Information related to the analytical reference standards, analytical instruments and methods used, the data reduction procedures and the data assessment procedures are all found in the AA-PGVP QAPP and SOP and are not repeated in this report. Table 3 provides the measurement quality objectives (MQOs) that are included in the AA-PGVP QAPP (Table 7-1 of the QAPP). The acceptance criteria in Table 3 were met for each day of verification. In addition, conformance to these requirements can be found in the measurement data worksheets that are generated for each comparison run and are available upon request. Appendix A provides a report of the quality control (QC) checks associated with each verification run. Table 4 provides the verification results for CO and SO₂, and Table 5 provides the NO_x results.

Table 3. MQOs for the AA-PGVP

Requirement	Frequency	Acceptance Criteria	Protocol Gas Doc. Reference	Comments
Completeness	All standards analyzed	95%		Based on an anticipated 40 cylinders per lab per year.
Quarterly Flow Calibration	Quarterly -no more than 1 mo. before verification	Calibration flow accuracy within $\pm 1\%$	2.3.7	Using flow primary standard
Calibrator Dilution Check	Quarterly -within 2 weeks of assay	$\pm 1\%$ RD	2.3.5.1	Second SRM. Three or more discrete measurements
Analyzer Calibration	Quarterly - within 2 weeks of assay	$\pm 1\%$ RPD (each point) Slope 0.89 – 1.02	2.1.7.2	5 points between 50-90% of upper range limit of analyzer + zero point
Zero & Span Verifications	Each day of verification	SE mean $\leq 1\%$ and accuracy $\pm 5\%$ RD	2.1.7.3, 2.3.5.4	Drift accountability. 3 discrete measurements of zero and span
Precision Test ¹	Day of Verification	$\pm 1\%$ RD standard error of the mean	2.3.5.4	SRM at conc. >80% of analyzer URL
Routine Data Check	Any Standard with Value >2% Tag Value	NA		Sample run three times to verify value.
Lab Comparability	2/year	$\pm 2\%$ RPD	NA	Sample run three average value used.
Standards Certification				
Primary flow standard	Annually-Certified by NVLAP certified lab	1.0 %	NA	Compared to NIST Traceable
NIST SRMs	Expiration date SRM pressure > 150 psig			Will follow NIST recertification requirements

TABLE 3. MQOs FOR THE AA-PGVP

¹ The precision test does not need to be accomplished if analyzer calibrated on same day as analysis.

Table 4. 2020 AA-PGVP CO and SO₂ Verifications[‡]

Qtr	Lab	Producer	Facility	Cylinder ID	Pollutant	Assay Conc	Producer Conc	% Bias	95% Uncertainty
4	7	Air Gas	Los Angeles, CA	BAL4148	CO	909.5	907.7	0.19	0.27
3	7	Air Gas	Chicago, IL	CC93605	CO	220.4	220.1	0.14	0.26
4	7	Air Gas	Chicago, IL	FF55927	SO ₂	5.000	5.042	-0.82	0.42
4	7	Tier 5	Indianapolis, IN	LL197504	SO ₂	49.42	49.5	-0.15	0.21

TABLE 4. AA-PGVP CO AND SO₂ VERIFICATIONS

Notes: ‡ An Estimate for the national usage for specific protocol gas producers cannot be determined due to lack of participation in annual survey

Table 5. 2020 AA-PGVP NO and NO_x Verifications[‡]

Qtr	Lab	Producer	Facility	Cylinder ID	Pollutant	Assay Conc	Producer Conc	% Bias	95% Uncertainty
3	7	PraxAir	Toledo, OH	DT0029223*	NO	25.54	25.4	0.55	0.18
3	7	PraxAir	Los Angeles, CA	DT0011078*	NO	49.69	49.7	-0.02	0.18
3	7	PraxAir	Toledo, OH	LL127859	NO	25.2	25.1	0.38	0.18
3	7	WestAir	Anaheim, CA	CC460552**	NO	15	14.88	0.82	0.19
4	7	Air Gas	Los Angeles, CA	BAL4148	NO	44.88	45	-0.27	0.17
3	7	Matheson	Waverly, TN	SD15080	NO	19.55	19.96	-2.03 ^μ	0.19
4	7	Air Gas	Chicago, IL	CC178285	NO	10.024	9.886	1.4	0.5
4	7	Air Gas	Chicago, IL	FF36562	NO	9.644	9.439	2.17 ^μ	0.52
4	7	Tier 5	Indianapolis, IN	LL197501	NO	48.14	49.5	-2.75	0.16
3	7	PraxAir	Toledo, OH	DT0029223*	NO _x ^Ω	25.54	25.7	-0.62	0.21
3	7	PraxAir	Los Angeles, CA	DT0011078*	NO _x ^Ω	49.75	49.9	-0.3	0.21
3	7	PraxAir	Toledo, OH	LL127859	NO _x ^Ω	25.19	25.4	-0.83	0.22
3	7	WestAir	Anaheim, CA	CC460552**	NO _x	15.06	14.97	0.63	0.22
4	7	Air Gas	Los Angeles, CA	BAL4148	NO _x	45.21	45.11	0.23	0.17
3	7	Matheson	Waverly, TN	SD15080	NO _x ^Ω	19.57	20.01	-2.22 ^μ	0.22
4	7	Air Gas	Chicago, IL	CC178285	NO _x	10.063	9.886	1.79	0.41
4	7	Air Gas	Chicago, IL	FF36562	NO _x	9.986	9.56	4.46	0.41
4	7	Tier 5	Indianapolis, IN	LL197501	NO _x ^Ω	48.25	50	-3.49	0.17

TABLE 5. AA-PGVP NO AND NO_x VERIFICATIONS

Notes: * Direct submission by producer, assay results not blind to producer
 ** Cylinder opened prior to submission to EPA
 ‡ An Estimate for the national usage for specific protocol gas producers cannot be determined due to lack of participation in annual survey
^μ % Bias within the ±2% acid rain criteria acceptance criterion when RAVL 95% Uncertainty is considered
^Ω NO_x concentration provided by Producer as “informational only”; concentration not certified by Producer.

Twelve cylinders were received by the AA-PGVP during calendar year 2020. Ten of these cylinders were received from regulatory monitoring agencies to support the AA-PGVP. Two of the cylinders were directly submitted from protocol gas producers. The assay results for all cylinders are included in Tables 4 and 5 but some are qualified with footnotes to denote that they do not fully meet the assessment requirements for the AA-PGVP. The NO cylinders directly submitted by Praxair were not blind to the producer. The NO cylinder provided by San Diego County Air Pollution Control District (SDAPCD) was opened by the agency prior to shipment to EPA. AA-PGVP requires that cylinders be unopened prior to receipt at the RAVL. While opened prior to receipt at the RAVL, the WestAir cylinder from SDAPCD is considered a valid sample based on the good agreement between EPA's independent assay and the assay results of the producer certified concentrations.

All results for the CO and SO₂ standards were within the ±2% acid rain criteria acceptance criterion. Two of the nine protocol gas cylinder standards assayed for NO_x (Cylinder IDs: FF36562 and LL197501) did not meet the ±2% acid rain criteria acceptance criterion. Cylinder ID FF36562 was also above the ±4% AA-PGVP action level criteria for NO_x. Additionally, Cylinder ID FF36562 also approached the ±2% acid rain criteria acceptance criterion for NO, but based on the RAVL's 95% uncertainty value of 0.52% it was assessed in this annual report as meeting the ±2% acid rain criteria acceptance criterion for NO. Cylinder ID LL197501 exceeded the ±2% acid rain criteria acceptance criterion for NO in addition to NO_x. The assay for Cylinder ID SD15080 was found to approach the ±2% acid rain criteria acceptance criterion for NO_x and NO but was assessed in this annual report as meeting these criteria when the RAVL's 95% uncertainty was considered. The NO_x concentration for five cylinders assayed were provided as "informational only" by the Producers and were not certified. While not certified, these NO_x concentrations were assayed by the AA-PGVP to provide context to for the certified NO component in the protocol gas cylinder standard.

In 2020 the AA-PGVP operated with a single RAVL. As such, the quality assurance designated for the laboratory intercomparison of the internal standards could not be performed. Protocol gas producers not meeting the ±2% acceptance criteria were offered, and accepted, an opportunity to perform their own confirmatory assay of the cylinders. The results of these producer independent assays were not provided to EPA to include in this report.

4.0 Summary and Conclusions

General –

The AA-PGVP is implementing a verification process that is blind to the specialty gas producers. One of goals of the ambient air monitoring rule (published March 28, 2016) was for the verifications performed by the RAVLs to be focused more on our ambient air monitoring organizations rather than as a resource to be utilized by specialty gas producers for their own quality assurance. The purpose of the program (blind verification of gas cylinders provided by monitoring organizations) cannot be accomplished if EPA relies on

the specialty gas producers to submit cylinders for assessment. Of the 12 protocol gas cylinder standards submitted for analysis only two cylinders were directly submitted by gas producers.

While the program is successfully implementing a blind verification process, only 12 cylinders were analyzed in 2020 or 15% of the AA-PGVP goal of 80 cylinders annually. EPA's single RAVL was only able to operate for the last half of 2020 due to the COVID-19 pandemic. Two of the 12 cylinders were sent direct from the producers and another cylinder was opened by the submitting agency before sent to EPA. These twelve-cylinder submissions resulted in only 22 verifications performed in 2020 (some cylinders are a blend of multiple gas standards). Results show that 3 of the 22 verifications (14%) failed the $\pm 2\%$ Acid Rain Program criteria. It is difficult to assess the extent to which this issue impacts our ambient air monitoring networks in 2020 due to the low utilization of the RAVLs by our monitoring programs and low participation rate in the annual protocol gas questionnaire. In 2020 there were 25 EPA protocol gas production facilities in operation nationally. It is uncertain how many of these facilities were used in the ambient air monitoring networks in 2020. Additionally, of the 25 protocol gas production facilities operating, only 7 were verified by our ambient air monitoring program in calendar year 2020.

Survey Participation Improvement –

Since its inception, the AA-PGVP has relied on an annual survey to determine which gas producers and facilities are used for generating CO, SO₂, and NO₂ test atmospheres from protocol gas cylinder standards. Participation in the annual survey was initially voluntary. To improve the participation rate and to more completely document which protocol gas producers are utilized by our ambient air monitoring organizations, in 2016 states using protocol gases were required to complete the survey every year. While it was thought at the time that this regulatory requirement would increase the participation and create a comprehensive list of the protocol gas producers used in the national network, the survey participation rate has not improved. In calendar year 2020 participation in the annual questionnaire was about 17% of the monitoring agencies that operate CO, SO₂, NO₂ ambient air analyzers. OAQPS is actively assessing EPA's AQS database as an alternative solution to gather this information. See Data Management Improvement section below for further details.

RAVL Participation Improvement –

Since the monitoring rule was revised in 2016, the AA-PGVP has made progress in achieving blind verifications of the protocol gas cylinders used in our ambient air monitoring networks. However, the program continues to not achieve its goal of having every Primary Quality Assurance Organization (PQAO) submit an unused cylinder at least once every five years for verification. The AA-PGVP's goal to perform 80 protocol gas verifications each year and to strategically select these protocol cylinders to represent the national ambient air monitoring networks was not achieved in calendar year 2020. Only twelve protocol gas cylinder standards were submitted by seven PQAOs in 2020 to support this national program. Region 7 assayed all the cylinders received in 2020. Four of the seven monitoring programs submitting protocol gas cylinders for the AA-PGVP were clustered in proximity to the Region 7 laboratory. A better national sampling of monitoring programs and protocol gas producers continues to be needed. Diminished assay capacity was due to only having a single RAVL in 2020 and operations being impacted by the COVID-19 pandemic. OAQPS is working to add assay capacity in 2021 and 2022 by using the Region 4 laboratory as an additional RAVL.

Quality System Improvement –

The Quality Assurance Project Plan (QAPP) has not been updated since the inception of the program in 2010. Since calendar year 2010, changes to the program have occurred, including regulatory changes in 2016. These documents need to be reconciled with current program practices and regulatory requirements. OAQPS began reviewing and revising the QAPP for this program in 2020 with a goal of having a draft revised QAPP by the end of 2021.

In 2020, the AA-PGVP operated with a single RAVL. As such, the quality assurance designated for the laboratory intercomparison of the internal standards could not be performed. OAQPS is currently working with EPA Region 4 to begin using their laboratory as a second RAVL. The Region 4 RAVL is scheduled to become operational the end of calendar year 2021.

Data Management Improvement –

The AA-PGVP has relied solely on the annual survey for determining which protocol gas standard producers are used in the national ambient air monitoring networks. The annual survey was originally a voluntary program and later in 2016 it became a regulatory requirement. Neither implementation of this process has proven to be fully effective. The data management practices for conducting the annual survey and storing its results are not optimized to be readily reconciled with the data produced by the RAVLs. Additionally, data validation and data entry business rules are needed to ensure the accuracy of the data submitted for both portions of this program (protocol gas survey and RAVL analytical results). Once accomplished this will enable both datasets to be readily assessed with respect to monitoring organization, PQAQO, and producer production facility. Data entry errors on the annual survey and chain of custody forms and the lack of key fields impede analysis of the information collected for this program. A revised spreadsheet based COC form with data validation will be used in calendar 2021 to help address so these issues.

OAQPS is actively investigating an AQS database solution to replace some of the data management practices historically performed in the program. This includes assessing the feasibility of making modifications to the current AQS “QA-Transaction” file format for the single point quality control checks and annual performance audits. The proposed modifications being investigated would allow for documenting the protocol gas producer and facility of the protocol gas cylinder used for generating the test atmospheres for each of these checks. Utilizing this modified AQS data submission process would allow EPA to document 100% of the protocol gas production facilities used in the ambient air monitoring networks as opposed to the current process which has only been 44% effective between 2015-2020.

Appendix A QA Reports from Measurement Data Worksheets for 2020

Ambient Air Protocol Gas Verification Program QA Reports from Measurement Data Worksheets for 2020

During the verification process, the Regional Air Verification Laboratories perform a number of quality control checks that are recorded on the Measurement Data Worksheets. This information is reported and saved along with the verification reports. The following sheets represent the quality control for all verifications that were implemented in 2020.

QA Requirements Summary, Region 7 - 3rd Quarter of 2020

	QA Requirement	Result	Status
SRM Gas Standards	Primary SRM Cylinder Expiration Date	13-Apr-24	Primary SRM Gas Standard OK
	Primary SRM Cylinder Pressure >150 psi	2000	Primary SRM cylinder pressure is OK
	SRM Dilution Check Cylinder Expiration Date	20-Sep-21	Dilution Check SRM Gas Standard OK
	Dilution Check SRM Cylinder Pressure >150 psi	675	Dilution check SRM cylinder pressure is OK
Laboratory Flow Standard	High Flow Standard Expiration Date	8-Jan-21	Standard OK
	Low Flow Standard Expiration Date	8-Jan-21	Standard OK
	Ultra Low Flow Standard Expiration Date	21-Jan-21	Standard OK
Calibrator (mass flow controllers)	Calibrator Flow Calibration within 2 weeks of assay	13-Sep-20	Calibrator flow calibration within 2 weeks of assay
	Calibrated High Flow MFC Slope Range = 0.99 - 1.0	0.9999990	High MFC OK
	Calibrated Low Flow MFC Slope Range = 0.99 - 1.0	0.9999907	Low MFC OK
Carbon Monoxide Gas Analyzer	Analyzer Calibration within 2 week of assay	13-Sep-20	Analyzer calibration within 2 weeks of assay
	Estimate of Uncertainty < 1% at point #1 (>80% URL	0.25%	Assay may be conducted at this concentration
	Estimate of Uncertainty < 1% at point #2	0.26%	Assay may be conducted at this concentration
	Estimate of Uncertainty < 1% at point #3	0.27%	Assay may be conducted at this concentration
	Estimate of Uncertainty < 1% at point #4	0.28%	Assay may be conducted at this concentration
	Estimate of Uncertainty < 1% at point #5 (~50% URL	0.31%	Assay may be conducted at this concentration
	Analyzer slope is within 0.98-1.02	1.0012	Analyzer Slope is acceptable
Dilution Check	Dilution Check Date within 2 weeks of assay	14-Sep-20	Dilution check within 2 weeks of assay
	Dilution Check Relative % Difference < 1%	-0.410%	Dilution Check RSD is OK
Day of Assay Zero/Span Check	Day of Assay Zero Check - Std. Error < 1%	Std. Error is okay.	Zero Gas Std. Error is OK
	Day of Assay Zero Check - Relative Difference < 5%	RD is okay.	Zero Gas RD is OK
	Day of Assay Span Check - Std. Error < 1%	Std. Error is okay.	Span Gas Std. Error is OK
	Day of Assay Span Check - Relative Difference <5%	RD is okay.	Span Gas RD is OK

QA Requirements Summary, Region 7 - 3rd Quarter of 2020

	QA Requirement	Result	Status
SRM Gas Standards	Primary SRM Cylinder Expiration Date	8-Aug-23	Primary SRM Gas Standard OK
	Primary SRM Cylinder Pressure >150 psi	550	Primary SRM cylinder pressure is OK
	SRM Dilution Check Cylinder Expiration Date	1-Feb-24	Dilution Check SRM Gas Standard OK
	Dilution Check SRM Cylinder Pressure >150 psi	800	Dilution check SRM cylinder pressure is OK
Laboratory Flow Standard	High Flow Standard Expiration Date	8-Jan-21	Standard OK
	Low Flow Standard Expiration Date	8-Jan-21	Standard OK
	Ultra Low Flow Expiration Date	21-Jan-21	Standard OK
Calibrator (mass flow controllers)	Calibrator Flow Calibration within 2 weeks of assay	13-Sep-20	Calibrator flow calibration within 2 weeks of assay
	Calibrated High Flow MFC Slope Range = 0.99 - 1.0	0.9999990	High MFC OK
	Calibrated Low Flow MFC Slope Range = 0.99 - 1.0	0.9999915	Low MFC OK
Oxides of Nitrogen Gas Analyzer NO Portion	Analyzer Calibration within 2 weeks of assay	15-Sep-20	Analyzer calibration within 2 weeks of assay
	Estimate of Uncertainty < 1% at point #1 (>80% URL	0.27%	Assay may be conducted at this concentration
	Estimate of Uncertainty < 1% at point #2	0.28%	Assay may be conducted at this concentration
	Estimate of Uncertainty < 1% at point #3	0.29%	Assay may be conducted at this concentration
	Estimate of Uncertainty < 1% at point #4	0.31%	Assay may be conducted at this concentration
	Estimate of Uncertainty < 1% at point #5 (~50% URL	0.34%	Assay may be conducted at this concentration
	Analyzer slope is within 0.98-1.02	1.0005	Analyzer Slope is acceptable
Oxides of Nitrogen Gas Analyzer NOx Portion	Analyzer Calibration within 2 week of assay	15-Sep-20	Analyzer calibration within 2 weeks of assay
	Estimate of Uncertainty < 1% at point #1 (>80% URL	0.32%	Assay may be conducted at this concentration
	Estimate of Uncertainty < 1% at point #2	0.33%	Assay may be conducted at this concentration
	Estimate of Uncertainty < 1% at point #3	0.34%	Assay may be conducted at this concentration
	Estimate of Uncertainty < 1% at point #4	0.36%	Assay may be conducted at this concentration
	Estimate of Uncertainty < 1% at point #5 (~50% URL	0.39%	Assay may be conducted at this concentration
	Analyzer slope is within 0.98-1.02	0.9967	Analyzer Slope is acceptable
Dilution Check	Dilution Check Date within 2 weeks of assay	14-Sep-20	Dilution check within 2 weeks of assay
	Dilution Check Relative % Difference < 1%	-0.407%	Dilution Check RSD is OK
Day of Assay Zero/Span Check NO Portion	Day of Assay Zero Check - Std. Error < 1%	Std. Error is okay.	Zero Gas Std. Error is OK
	Day of Assay Zero Check - Relative Difference < 5% RD is okay.		Zero Gas RD is OK
	Day of Assay Span Check - Std. Error < 1%	Std. Error is okay.	Span Gas Std. Error is OK
	Day of Assay Span Check - Relative Difference <5% RD is okay.		Span Gas RD is OK
Day of Assay Zero/Span Check NOx Portion	Day of Assay Zero Check - Std. Error < 1%	Std. Error is okay.	Zero Gas Std. Error is OK
	Day of Assay Zero Check - Relative Difference < 5% RD is okay.		Zero Gas RD is OK
	Day of Assay Span Check - Std. Error < 1%	Std. Error is okay.	Span Gas Std. Error is OK
	Day of Assay Span Check - Relative Difference <5% RD is okay.		Span Gas RD is OK

QA Requirements Summary, Region 7 - 4th Quarter of 2020

	QA Requirement	Result	Status
SRM Gas Standards	Primary SRM Cylinder Expiration Date	20-Sep-21	Primary SRM Gas Standard OK
	Primary SRM Cylinder Pressure >150 psi	1900	Primary SRM cylinder pressure is OK
	SRM Dilution Check Cylinder Expiration Date	20-Sep-21	Dilution Check SRM Gas Standard OK
	Dilution Check SRM Cylinder Pressure >150 psi	675	Dilution check SRM cylinder pressure is OK
Laboratory Flow Standard	High Flow Standard Expiration Date	8-Jan-21	Standard OK
	Low Flow Standard Expiration Date	8-Jan-21	Standard OK
	Ultra Low Flow Standard Expiration Date	21-Jan-21	Standard OK
Calibrator (mass flow controllers)	Calibrator Flow Calibration within 2 weeks of assay	6-Dec-20	Calibrator flow calibration within 2 weeks of assay
	Calibrated High Flow MFC Slope Range = 0.99 - 1.0	0.9999993	High MFC OK
	Calibrated Low Flow MFC Slope Range = 0.99 - 1.0	0.9999844	Low MFC OK
Carbon Monoxide Gas Analyzer	Analyzer Calibration within 2 week of assay	7-Dec-20	Analyzer calibration within 2 weeks of assay
	Estimate of Uncertainty < 1% at point #1 (>80% URL	0.38%	Assay may be conducted at this concentration
	Estimate of Uncertainty < 1% at point #2	0.39%	Assay may be conducted at this concentration
	Estimate of Uncertainty < 1% at point #3	0.41%	Assay may be conducted at this concentration
	Estimate of Uncertainty < 1% at point #4	0.44%	Assay may be conducted at this concentration
	Estimate of Uncertainty < 1% at point #5 (~50% URL	0.48%	Assay may be conducted at this concentration
	Analyzer slope is within 0.98-1.02	1.0029	Analyzer Slope is acceptable
Dilution Check	Dilution Check Date within 2 weeks of assay	7-Dec-20	Dilution check within 2 weeks of assay
	Dilution Check Relative % Difference < 1%	-0.107%	Dilution Check RSD is OK
Day of Assay Zero/Span Check	Day of Assay Zero Check - Std. Error < 1%	Std. Error is okay.	Zero Gas Std. Error is OK
	Day of Assay Zero Check - Relative Difference < 5%	RD is okay.	Zero Gas RD is OK
	Day of Assay Span Check - Std. Error < 1%	Std. Error is okay.	Span Gas Std. Error is OK
	Day of Assay Span Check - Relative Difference <5%	RD is okay.	Span Gas RD is OK

QA Requirements Summary, Region 7 - 4th Quarter of 2020

	QA Requirement	Result	Status
SRM Gas Standards	Primary SRM Cylinder Expiration Date	8-Aug-23	Primary SRM Gas Standard OK
	Primary SRM Cylinder Pressure >150 psi	475	Primary SRM cylinder pressure is OK
	SRM Dilution Check Cylinder Expiration Date	1-Feb-24	Dilution Check SRM Gas Standard OK
	Dilution Check SRM Cylinder Pressure >150 psi	800	Dilution check SRM cylinder pressure is OK
Laboratory Flow Standard	High Flow Standard Expiration Date	8-Jan-21	Standard OK
	Low Flow Standard Expiration Date	8-Jan-21	Standard OK
	Ultra Low Flow Expiration Date	21-Jan-21	Standard OK
Calibrator (mass flow controllers)	Calibrator Flow Calibration within 2 weeks of assay	6-Dec-20	Calibrator flow calibration within 2 weeks of assay
	Calibrated High Flow MFC Slope Range = 0.99 - 1.0	0.9999993	High MFC OK
	Calibrated Low Flow MFC Slope Range = 0.99 - 1.0	0.9999844	Low MFC OK
Oxides of Nitrogen Gas Analyzer NO Portion	Analyzer Calibration within 2 weeks of assay	10-Dec-20	Analyzer calibration within 2 weeks of assay
	Estimate of Uncertainty < 1% at point #1 (>80% URL	0.24%	Assay may be conducted at this concentration
	Estimate of Uncertainty < 1% at point #2	0.25%	Assay may be conducted at this concentration
	Estimate of Uncertainty < 1% at point #3	0.26%	Assay may be conducted at this concentration
	Estimate of Uncertainty < 1% at point #4	0.27%	Assay may be conducted at this concentration
	Estimate of Uncertainty < 1% at point #5 (~50% URL	0.30%	Assay may be conducted at this concentration
	Analyzer slope is within 0.98-1.02	1.0006	Analyzer Slope is acceptable
Oxides of Nitrogen Gas Analyzer NOx Portion	Analyzer Calibration within 2 week of assay	10-Dec-20	Analyzer calibration within 2 weeks of assay
	Estimate of Uncertainty < 1% at point #1 (>80% URL	0.24%	Assay may be conducted at this concentration
	Estimate of Uncertainty < 1% at point #2	0.25%	Assay may be conducted at this concentration
	Estimate of Uncertainty < 1% at point #3	0.26%	Assay may be conducted at this concentration
	Estimate of Uncertainty < 1% at point #4	0.28%	Assay may be conducted at this concentration
	Estimate of Uncertainty < 1% at point #5 (~50% URL	0.31%	Assay may be conducted at this concentration
	Analyzer slope is within 0.98-1.02	1.0002	Analyzer Slope is acceptable
Dilution Check	Dilution Check Date within 2 weeks of assay	7-Dec-20	Dilution check within 2 weeks of assay
	Dilution Check Relative % Difference < 1%	0.000%	Dilution Check RSD is OK
Day of Assay Zero/Span Check NO Portion	Day of Assay Zero Check - Std. Error < 1%	Std. Error is okay.	Zero Gas Std. Error is OK
	Day of Assay Zero Check - Relative Difference < 5% RD is okay.		Zero Gas RD is OK
	Day of Assay Span Check - Std. Error < 1%	Std. Error is okay.	Span Gas Std. Error is OK
	Day of Assay Span Check - Relative Difference <5% RD is okay.		Span Gas RD is OK
Day of Assay Zero/Span Check NOx Portion	Day of Assay Zero Check - Std. Error < 1%	Std. Error is okay.	Zero Gas Std. Error is OK
	Day of Assay Zero Check - Relative Difference < 5% RD is okay.		Zero Gas RD is OK
	Day of Assay Span Check - Std. Error < 1%	Std. Error is okay.	Span Gas Std. Error is OK
	Day of Assay Span Check - Relative Difference <5% RD is okay.		Span Gas RD is OK

QA Requirements Summary, Region 7 - 4th Quarter of 2020

	QA Requirement	Result	Status
SRM Gas Standards	Primary SRM Cylinder Expiration Date	27-Jun-23	Primary SRM Gas Standard OK
	Primary SRM Cylinder Pressure >150 psi	1350	Primary SRM cylinder pressure is OK
	SRM Dilution Check Cylinder Expiration Date	5-Apr-22	Dilution Check SRM Gas Standard OK
	Dilution Check SRM Cylinder Pressure >150 psi	1625	Dilution check SRM cylinder pressure is OK
Laboratory Flow Standard	High Flow Standard Expiration Date	8-Jan-21	Standard OK
	Low Flow Standard Expiration Date	8-Jan-21	Standard OK
	Ultra Low Flow Standard Expiration Date	21-Jan-21	Standard OK
Calibrator (mass flow controllers)	Calibrator Flow Calibration within 2 weeks of assay	6-Dec-20	Calibrator flow calibration within 2 weeks of assay
	Calibrated High Flow MFC Slope Range = 0.99 - 1.0	0.9999993	High MFC OK
	Calibrated Low Flow MFC Slope Range = 0.99 - 1.0	0.9999929	Low MFC OK
Sulfur Dioxide Gas Analyzer	Analyzer Calibration within 2 weeks of assay	8-Dec-20	Analyzer calibration within 2 weeks of assay
	Estimate of Uncertainty < 1% at point #1 (>80% URL	0.31%	Assay may be conducted at this concentration
	Estimate of Uncertainty < 1% at point #2	0.32%	Assay may be conducted at this concentration
	Estimate of Uncertainty < 1% at point #3	0.35%	Assay may be conducted at this concentration
	Estimate of Uncertainty < 1% at point #4	0.41%	Assay may be conducted at this concentration
	Estimate of Uncertainty < 1% at point #5 (~50% URL	0.74%	Assay may be conducted at this concentration
	Analyzer slope is within 0.98-1.02	1.0007	Analyzer Slope is acceptable
Dilution Check	Dilution Check Date within 2 weeks of assay	7-Dec-20	Dilution check within 2 weeks of assay
	Dilution Check Relative % Difference < 1%	-0.107%	Dilution Check RSD is OK
Day of Assay Zero/Span Check	Day of Assay Zero Check - Std. Error < 1%	Std. Error is okay.	Zero Gas Std. Error is OK
	Day of Assay Zero Check - Relative Difference < 5%	RD is okay.	Zero Gas RD is OK
	Day of Assay Span Check - Std. Error < 1%	Std. Error is okay.	Span Gas Std. Error is OK
	Day of Assay Span Check - Relative Difference <5%	RD is okay.	Span Gas RD is OK

United States
Environmental Protection
Agency

Office of Air Quality Planning and Standards
Air Quality Assessment Division
Research Triangle Park, NC

Publication No. EPA-454/R-21-006
July 2021
