



Pace Analytical Services, LLC
1700 Elm St. SE, Suite 200
Minneapolis, MN 55414

Phone: 612-607-1700
Fax: 612-607-6444

Narrative Summary:
Pace Analytical Atlanta WO#
10579319
Client project ID Conbraco
Pace MN WO#10579319

The laboratory received 32 air samples for analysis by method EPA TO15.

At client request, all data associated with Pace workorder 10579319 with no discrepancies to report. All cans were individually certified. Previous media usage was reviewed (canister and flow controller data) to ensure no potential carryover or contamination could have occurred. Additionally, the instrument conditions were reviewed for potential sources of carryover or contamination. No elevated trichloroethene results were found during the investigation.

Lastly, canisters in the run log were matched to the Chain of Custody to ensure that there was no labeling errors had occurred. After review, all canisters were labeled correctly.

Based on the laboratory review, there is no indication that the trichloroethene detections in these samples were impacted by the media or laboratory conditions.

Prepared 10/5/2021
David Randall
Laboratory Services Manager

September 27, 2021

Wes Barfield
Shield Engineering
4301 Taggart Creek Rd
Charlotte, NC 28203

RE: Project: CONBRACO
Pace Project No.: 92561801

Dear Wes Barfield:

Enclosed are the analytical results for sample(s) received by the laboratory on September 16, 2021. The results relate only to the samples included in this report. Results reported herein conform to the applicable TNI/NELAC Standards and the laboratory's Quality Manual, where applicable, unless otherwise noted in the body of the report.

The test results provided in this final report were generated by each of the following laboratories within the Pace Network:

- Pace Analytical Services - Minneapolis

If you have any questions concerning this report, please feel free to contact me.

Sincerely,

Ryan Brumfield
ryan.brumfield@pacelabs.com
(770)734-4200
Project Manager

Enclosures



REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.

CERTIFICATIONS

Project: CONBRACO

Pace Project No.: 92561801

Pace Analytical Services, LLC - Minneapolis MN

1700 Elm Street SE, Minneapolis, MN 55414
1800 Elm Street SE, Minneapolis, MN 55414--Satellite Air Lab

A2LA Certification #: 2926.01*
Alabama Certification #: 40770
Alaska Contaminated Sites Certification #: 17-009*
Alaska DW Certification #: MN00064
Arizona Certification #: AZ0014*
Arkansas DW Certification #: MN00064
Arkansas WW Certification #: 88-0680
California Certification #: 2929
Colorado Certification #: MN00064
Connecticut Certification #: PH-0256
EPA Region 8 Tribal Water Systems+Wyoming DW Certification #: via MN 027-053-137
Florida Certification #: E87605*
Georgia Certification #: 959
Hawaii Certification #: MN00064
Idaho Certification #: MN00064
Illinois Certification #: 200011
Indiana Certification #: C-MN-01
Iowa Certification #: 368
Kansas Certification #: E-10167
Kentucky DW Certification #: 90062
Kentucky WW Certification #: 90062
Louisiana DEQ Certification #: AI-03086*
Louisiana DW Certification #: MN00064
Maine Certification #: MN00064*
Maryland Certification #: 322
Michigan Certification #: 9909
Minnesota Certification #: 027-053-137*
Minnesota Dept of Ag Approval: via MN 027-053-137
Minnesota Petrofund Registration #: 1240*
Mississippi Certification #: MN00064

Missouri Certification #: 10100
Montana Certification #: CERT0092
Nebraska Certification #: NE-OS-18-06
Nevada Certification #: MN00064
New Hampshire Certification #: 2081*
New Jersey Certification #: MN002
New York Certification #: 11647*
North Carolina DW Certification #: 27700
North Carolina WW Certification #: 530
North Dakota Certification #: R-036
Ohio DW Certification #: 41244
Ohio VAP Certification (1700) #: CL101
Ohio VAP Certification (1800) #: CL110*
Oklahoma Certification #: 9507*
Oregon Primary Certification #: MN300001
Oregon Secondary Certification #: MN200001*
Pennsylvania Certification #: 68-00563*
Puerto Rico Certification #: MN00064
South Carolina Certification #:74003001
Tennessee Certification #: TN02818
Texas Certification #: T104704192*
Utah Certification #: MN00064*
Vermont Certification #: VT-027053137
Virginia Certification #: 460163*
Washington Certification #: C486*
West Virginia DEP Certification #: 382
West Virginia DW Certification #: 9952 C
Wisconsin Certification #: 999407970
Wyoming UST Certification #: via A2LA 2926.01
USDA Permit #: P330-19-00208
Please Note: Applicable air certifications are denoted with an asterisk ().

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.

SAMPLE SUMMARY

Project: CONBRACO

Pace Project No.: 92561801

Lab ID	Sample ID	Matrix	Date Collected	Date Received
92561801001	IA-14	Air	09/14/21 17:32	09/16/21 12:00
92561801002	IA-6	Air	09/14/21 16:39	09/16/21 12:00
92561801003	IA-1	Air	09/14/21 18:15	09/16/21 12:00
92561801004	IA-12	Air	09/14/21 17:35	09/16/21 12:00
92561801005	IA-2	Air	09/14/21 18:16	09/16/21 12:00
92561801006	IA-8	Air	09/14/21 16:52	09/16/21 12:00
92561801007	IA-5	Air	09/14/21 16:15	09/16/21 12:00
92561801008	DUP-1	Air	09/14/21 17:04	09/16/21 12:00
92561801009	IA-13	Air	09/14/21 17:04	09/16/21 12:00
92561801010	IA-7	Air	09/14/21 16:56	09/16/21 12:00
92561801011	IA-3	Air	09/14/21 16:48	09/16/21 12:00
92561801012	IA-10	Air	09/14/21 17:25	09/16/21 12:00
92561801013	IA-11	Air	09/14/21 18:45	09/16/21 12:00
92561801014	IA-4	Air	09/14/21 16:16	09/16/21 12:00
92561801015	IA-9	Air	09/14/21 18:04	09/16/21 12:00
92561801016	BACKGROUND	Air	09/14/21 18:08	09/16/21 12:00
92561801017	IA-14 Cert# 1604	Air	09/14/21 17:32	09/16/21 13:15
92561801018	IA-6 Cert# 2319	Air	09/14/21 16:29	09/16/21 13:15
92561801019	IA-1 Cert #0799	Air	09/14/21 18:15	09/16/21 13:15
92561801020	IA-12 Cert# 0145	Air	09/14/21 17:35	09/16/21 13:15
92561801021	IA-2 Cert# 0796	Air	09/14/21 18:16	09/16/21 13:15
92561801022	IA-8 Cert# 2109	Air	09/14/21 16:52	09/16/21 13:15
92561801023	IA-5 Cert# 2811	Air	09/14/21 16:15	09/16/21 13:15
92561801024	DUP-1 Cert#1195	Air	09/14/21 17:09	09/16/21 13:15
92561801025	IA-13 Cert# 2803	Air	09/14/21 17:09	09/16/21 13:15
92561801026	IA-7 Cert# 3629	Air	09/14/21 16:56	09/16/21 13:15
92561801027	IA-3 Cert # 0629	Air	09/14/21 16:48	09/16/21 13:15
92561801028	IA-10 Cert# 3475	Air	09/14/21 17:35	09/16/21 13:15
92561801029	IA-11 Cert# 3408	Air	09/14/21 18:45	09/16/21 13:15
92561801030	IA-4 Cert# 2049	Air	09/14/21 16:16	09/16/21 13:15
92561801031	IA-9 Cert# 0857	Air	09/14/21 18:04	09/16/21 13:15
92561801032	Background Cert# 3891	Air	09/14/21 18:08	09/16/21 13:15
92561801033	Unused Can #2703	Air	09/14/21 00:00	09/16/21 13:15

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.

SAMPLE ANALYTE COUNT

Project: CONBRACO
Pace Project No.: 92561801

Lab ID	Sample ID	Method	Analysts	Analytes Reported	Laboratory
92561801001	IA-14	TO-15	SW	61	PASI-M
92561801002	IA-6	TO-15	SW	61	PASI-M
92561801003	IA-1	TO-15	SW	61	PASI-M
92561801004	IA-12	TO-15	SW	61	PASI-M
92561801005	IA-2	TO-15	SW	61	PASI-M
92561801006	IA-8	TO-15	SW	61	PASI-M
92561801007	IA-5	TO-15	SW	61	PASI-M
92561801008	DUP-1	TO-15	SW	61	PASI-M
92561801009	IA-13	TO-15	SW	61	PASI-M
92561801010	IA-7	TO-15	SW	61	PASI-M
92561801011	IA-3	TO-15	SW	61	PASI-M
92561801012	IA-10	TO-15	SW	61	PASI-M
92561801013	IA-11	TO-15	SW	61	PASI-M
92561801014	IA-4	TO-15	SW	61	PASI-M
92561801015	IA-9	TO-15	SW	61	PASI-M
92561801016	BACKGROUND	TO-15	SW	61	PASI-M
92561801017	IA-14 Cert# 1604	TO-15	SW	61	PASI-M
92561801018	IA-6 Cert# 2319	TO-15	SW	61	PASI-M
92561801019	IA-1 Cert #0799	TO-15	SW	61	PASI-M
92561801020	IA-12 Cert# 0145	TO-15	GT	61	PASI-M
92561801021	IA-2 Cert# 0796	TO-15	HMH	61	PASI-M
92561801022	IA-8 Cert# 2109	TO-15	AFV	61	PASI-M
92561801023	IA-5 Cert# 2811	TO-15	MJL	61	PASI-M
92561801024	DUP-1 Cert#1195	TO-15	SW	61	PASI-M
92561801025	IA-13 Cert# 2803	TO-15	SW	61	PASI-M
92561801026	IA-7 Cert# 3629	TO-15	SW	61	PASI-M
92561801027	IA-3 Cert # 0629	TO-15	HMH	61	PASI-M
92561801028	IA-10 Cert# 3475	TO-15	SW	61	PASI-M
92561801029	IA-11 Cert# 3408	TO-15	AFV	61	PASI-M
92561801030	IA-4 Cert# 2049	TO-15	SW	61	PASI-M
92561801031	IA-9 Cert# 0857	TO-15	AFV	61	PASI-M
92561801032	Background Cert# 3891	TO-15	HMH	61	PASI-M

PASI-M = Pace Analytical Services - Minneapolis

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.

ANALYTICAL RESULTS

Project: CONBRACO

Pace Project No.: 92561801

Sample: IA-14 Lab ID: 92561801001 Collected: 09/14/21 17:32 Received: 09/16/21 12:00 Matrix: Air

Parameters	Results	Units	Report			Prepared	Analyzed	CAS No.	Qual
			Limit	MDL	DF				
TO15 MSV AIR									
Analytical Method: TO-15									
Pace Analytical Services - Minneapolis									
Acetone	616	ug/m3	9.5	2.9	1.58		09/23/21 16:49	67-64-1	
Benzene	0.34J	ug/m3	0.51	0.18	1.58		09/23/21 16:49	71-43-2	
Benzyl chloride	<1.4	ug/m3	4.2	1.4	1.58		09/23/21 16:49	100-44-7	
Bromodichloromethane	<0.37	ug/m3	2.1	0.37	1.58		09/23/21 16:49	75-27-4	
Bromoform	<2.6	ug/m3	8.3	2.6	1.58		09/23/21 16:49	75-25-2	
Bromomethane	<0.24	ug/m3	1.2	0.24	1.58		09/23/21 16:49	74-83-9	
1,3-Butadiene	<0.19	ug/m3	0.71	0.19	1.58		09/23/21 16:49	106-99-0	
2-Butanone (MEK)	2.2J	ug/m3	4.7	0.73	1.58		09/23/21 16:49	78-93-3	
Carbon disulfide	0.73J	ug/m3	1.0	0.20	1.58		09/23/21 16:49	75-15-0	
Carbon tetrachloride	<0.44	ug/m3	2.0	0.44	1.58		09/23/21 16:49	56-23-5	
Chlorobenzene	<0.24	ug/m3	1.5	0.24	1.58		09/23/21 16:49	108-90-7	
Chloroethane	<0.35	ug/m3	0.85	0.35	1.58		09/23/21 16:49	75-00-3	
Chloroform	<0.29	ug/m3	0.78	0.29	1.58		09/23/21 16:49	67-66-3	
Chloromethane	1.0	ug/m3	0.66	0.13	1.58		09/23/21 16:49	74-87-3	
Cyclohexane	0.74J	ug/m3	2.8	0.35	1.58		09/23/21 16:49	110-82-7	
Dibromochloromethane	<0.81	ug/m3	2.7	0.81	1.58		09/23/21 16:49	124-48-1	
1,2-Dibromoethane (EDB)	<0.47	ug/m3	1.2	0.47	1.58		09/23/21 16:49	106-93-4	
1,2-Dichlorobenzene	<0.64	ug/m3	4.8	0.64	1.58		09/23/21 16:49	95-50-1	
1,3-Dichlorobenzene	<0.80	ug/m3	4.8	0.80	1.58		09/23/21 16:49	541-73-1	
1,4-Dichlorobenzene	<1.4	ug/m3	4.8	1.4	1.58		09/23/21 16:49	106-46-7	
Dichlorodifluoromethane	2.1	ug/m3	1.6	0.30	1.58		09/23/21 16:49	75-71-8	
1,1-Dichloroethane	<0.26	ug/m3	1.3	0.26	1.58		09/23/21 16:49	75-34-3	
1,2-Dichloroethane	<0.31	ug/m3	1.3	0.31	1.58		09/23/21 16:49	107-06-2	
1,1-Dichloroethene	<0.22	ug/m3	1.3	0.22	1.58		09/23/21 16:49	75-35-4	
cis-1,2-Dichloroethene	<0.31	ug/m3	1.3	0.31	1.58		09/23/21 16:49	156-59-2	
trans-1,2-Dichloroethene	<0.27	ug/m3	1.3	0.27	1.58		09/23/21 16:49	156-60-5	
1,2-Dichloropropane	<0.43	ug/m3	1.5	0.43	1.58		09/23/21 16:49	78-87-5	
cis-1,3-Dichloropropene	<0.40	ug/m3	3.6	0.40	1.58		09/23/21 16:49	10061-01-5	
trans-1,3-Dichloropropene	<0.86	ug/m3	3.6	0.86	1.58		09/23/21 16:49	10061-02-6	
Dichlorotetrafluoroethane	<0.32	ug/m3	2.2	0.32	1.58		09/23/21 16:49	76-14-2	
Ethanol	144	ug/m3	3.0	0.94	1.58		09/23/21 16:49	64-17-5	
Ethyl acetate	<0.21	ug/m3	1.2	0.21	1.58		09/23/21 16:49	141-78-6	
Ethylbenzene	0.91J	ug/m3	1.4	0.49	1.58		09/23/21 16:49	100-41-4	
4-Ethyltoluene	<0.75	ug/m3	4.0	0.75	1.58		09/23/21 16:49	622-96-8	
n-Heptane	<0.29	ug/m3	1.3	0.29	1.58		09/23/21 16:49	142-82-5	
Hexachloro-1,3-butadiene	<1.9	ug/m3	8.6	1.9	1.58		09/23/21 16:49	87-68-3	
n-Hexane	0.82J	ug/m3	1.1	0.30	1.58		09/23/21 16:49	110-54-3	
2-Hexanone	<0.70	ug/m3	6.6	0.70	1.58		09/23/21 16:49	591-78-6	
Methylene Chloride	<0.94	ug/m3	5.6	0.94	1.58		09/23/21 16:49	75-09-2	
4-Methyl-2-pentanone (MIBK)	<0.51	ug/m3	6.6	0.51	1.58		09/23/21 16:49	108-10-1	
Methyl-tert-butyl ether	<0.20	ug/m3	5.8	0.20	1.58		09/23/21 16:49	1634-04-4	
Naphthalene	<3.4	ug/m3	4.2	3.4	1.58		09/23/21 16:49	91-20-3	
2-Propanol	36.6	ug/m3	4.0	0.80	1.58		09/23/21 16:49	67-63-0	
Propylene	<0.21	ug/m3	1.4	0.21	1.58		09/23/21 16:49	115-07-1	
Styrene	<0.61	ug/m3	3.4	0.61	1.58		09/23/21 16:49	100-42-5	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.

ANALYTICAL RESULTS

Project: CONBRACO
Pace Project No.: 92561801

Sample: IA-14 Lab ID: 92561801001 Collected: 09/14/21 17:32 Received: 09/16/21 12:00 Matrix: Air

Parameters	Results	Units	Report			Prepared	Analyzed	CAS No.	Qual
			Limit	MDL	DF				
TO15 MSV AIR									
Analytical Method: TO-15									
Pace Analytical Services - Minneapolis									
1,1,2,2-Tetrachloroethane	<0.59	ug/m3	2.2	0.59	1.58		09/23/21 16:49	79-34-5	
Tetrachloroethene	<0.46	ug/m3	1.1	0.46	1.58		09/23/21 16:49	127-18-4	
Tetrahydrofuran	<0.28	ug/m3	0.95	0.28	1.58		09/23/21 16:49	109-99-9	
Toluene	1.6	ug/m3	1.2	0.39	1.58		09/23/21 16:49	108-88-3	
1,2,4-Trichlorobenzene	<7.7	ug/m3	11.9	7.7	1.58		09/23/21 16:49	120-82-1	
1,1,1-Trichloroethane	<0.29	ug/m3	1.8	0.29	1.58		09/23/21 16:49	71-55-6	
1,1,2-Trichloroethane	<0.31	ug/m3	0.88	0.31	1.58		09/23/21 16:49	79-00-5	
Trichloroethene	7.6	ug/m3	0.86	0.31	1.58		09/23/21 16:49	79-01-6	
Trichlorofluoromethane	3.6	ug/m3	1.8	0.37	1.58		09/23/21 16:49	75-69-4	
1,1,2-Trichlorotrifluoroethane	0.50J	ug/m3	2.5	0.46	1.58		09/23/21 16:49	76-13-1	
1,2,4-Trimethylbenzene	<0.56	ug/m3	1.6	0.56	1.58		09/23/21 16:49	95-63-6	
1,3,5-Trimethylbenzene	<0.46	ug/m3	1.6	0.46	1.58		09/23/21 16:49	108-67-8	
Vinyl acetate	<0.33	ug/m3	1.1	0.33	1.58		09/23/21 16:49	108-05-4	
Vinyl chloride	<0.14	ug/m3	0.41	0.14	1.58		09/23/21 16:49	75-01-4	
m&p-Xylene	<1.0	ug/m3	2.8	1.0	1.58		09/23/21 16:49	179601-23-1	
o-Xylene	<0.43	ug/m3	1.4	0.43	1.58		09/23/21 16:49	95-47-6	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.

ANALYTICAL RESULTS

Project: CONBRACO
Pace Project No.: 92561801

Sample: IA-6 Lab ID: 92561801002 Collected: 09/14/21 16:39 Received: 09/16/21 12:00 Matrix: Air

Parameters	Results	Units	Report			Prepared	Analyzed	CAS No.	Qual
			Limit	MDL	DF				
TO15 MSV AIR									
Analytical Method: TO-15									
Pace Analytical Services - Minneapolis									
Acetone	23.6	ug/m3	9.7	2.9	1.61		09/23/21 22:10	67-64-1	
Benzene	0.68	ug/m3	0.52	0.18	1.61		09/23/21 22:10	71-43-2	
Benzyl chloride	<1.4	ug/m3	4.2	1.4	1.61		09/23/21 22:10	100-44-7	
Bromodichloromethane	<0.38	ug/m3	2.2	0.38	1.61		09/23/21 22:10	75-27-4	
Bromoform	<2.6	ug/m3	8.5	2.6	1.61		09/23/21 22:10	75-25-2	
Bromomethane	<0.24	ug/m3	1.3	0.24	1.61		09/23/21 22:10	74-83-9	
1,3-Butadiene	<0.19	ug/m3	0.72	0.19	1.61		09/23/21 22:10	106-99-0	
2-Butanone (MEK)	1.7J	ug/m3	4.8	0.75	1.61		09/23/21 22:10	78-93-3	
Carbon disulfide	<0.21	ug/m3	1.0	0.21	1.61		09/23/21 22:10	75-15-0	
Carbon tetrachloride	<0.45	ug/m3	2.1	0.45	1.61		09/23/21 22:10	56-23-5	
Chlorobenzene	<0.25	ug/m3	1.5	0.25	1.61		09/23/21 22:10	108-90-7	
Chloroethane	<0.36	ug/m3	0.86	0.36	1.61		09/23/21 22:10	75-00-3	
Chloroform	1.4	ug/m3	0.80	0.29	1.61		09/23/21 22:10	67-66-3	
Chloromethane	43.8	ug/m3	0.68	0.14	1.61		09/23/21 22:10	74-87-3	
Cyclohexane	<0.36	ug/m3	2.8	0.36	1.61		09/23/21 22:10	110-82-7	
Dibromochloromethane	<0.83	ug/m3	2.8	0.83	1.61		09/23/21 22:10	124-48-1	
1,2-Dibromoethane (EDB)	<0.48	ug/m3	1.3	0.48	1.61		09/23/21 22:10	106-93-4	
1,2-Dichlorobenzene	<0.65	ug/m3	4.9	0.65	1.61		09/23/21 22:10	95-50-1	
1,3-Dichlorobenzene	<0.82	ug/m3	4.9	0.82	1.61		09/23/21 22:10	541-73-1	
1,4-Dichlorobenzene	<1.4	ug/m3	4.9	1.4	1.61		09/23/21 22:10	106-46-7	
Dichlorodifluoromethane	2.5	ug/m3	1.6	0.30	1.61		09/23/21 22:10	75-71-8	
1,1-Dichloroethane	<0.27	ug/m3	1.3	0.27	1.61		09/23/21 22:10	75-34-3	
1,2-Dichloroethane	<0.31	ug/m3	1.3	0.31	1.61		09/23/21 22:10	107-06-2	
1,1-Dichloroethene	<0.22	ug/m3	1.3	0.22	1.61		09/23/21 22:10	75-35-4	
cis-1,2-Dichloroethene	3.3	ug/m3	1.3	0.31	1.61		09/23/21 22:10	156-59-2	
trans-1,2-Dichloroethene	<0.27	ug/m3	1.3	0.27	1.61		09/23/21 22:10	156-60-5	
1,2-Dichloropropane	<0.43	ug/m3	1.5	0.43	1.61		09/23/21 22:10	78-87-5	
cis-1,3-Dichloropropene	<0.41	ug/m3	3.7	0.41	1.61		09/23/21 22:10	10061-01-5	
trans-1,3-Dichloropropene	<0.88	ug/m3	3.7	0.88	1.61		09/23/21 22:10	10061-02-6	
Dichlorotetrafluoroethane	<0.33	ug/m3	2.3	0.33	1.61		09/23/21 22:10	76-14-2	
Ethanol	232	ug/m3	3.1	0.95	1.61		09/23/21 22:10	64-17-5	
Ethyl acetate	12.4	ug/m3	1.2	0.21	1.61		09/23/21 22:10	141-78-6	
Ethylbenzene	<0.50	ug/m3	1.4	0.50	1.61		09/23/21 22:10	100-41-4	
4-Ethyltoluene	<0.76	ug/m3	4.0	0.76	1.61		09/23/21 22:10	622-96-8	
n-Heptane	<0.29	ug/m3	1.3	0.29	1.61		09/23/21 22:10	142-82-5	
Hexachloro-1,3-butadiene	<2.0	ug/m3	8.7	2.0	1.61		09/23/21 22:10	87-68-3	
n-Hexane	0.55J	ug/m3	1.2	0.31	1.61		09/23/21 22:10	110-54-3	
2-Hexanone	<0.71	ug/m3	6.7	0.71	1.61		09/23/21 22:10	591-78-6	
Methylene Chloride	<0.95	ug/m3	5.7	0.95	1.61		09/23/21 22:10	75-09-2	
4-Methyl-2-pentanone (MIBK)	<0.52	ug/m3	6.7	0.52	1.61		09/23/21 22:10	108-10-1	
Methyl-tert-butyl ether	<0.20	ug/m3	5.9	0.20	1.61		09/23/21 22:10	1634-04-4	
Naphthalene	<3.5	ug/m3	4.3	3.5	1.61		09/23/21 22:10	91-20-3	
2-Propanol	4.9	ug/m3	4.0	0.82	1.61		09/23/21 22:10	67-63-0	
Propylene	<0.21	ug/m3	1.4	0.21	1.61		09/23/21 22:10	115-07-1	
Styrene	<0.62	ug/m3	3.5	0.62	1.61		09/23/21 22:10	100-42-5	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.

ANALYTICAL RESULTS

Project: CONBRACO
Pace Project No.: 92561801

Sample: IA-6 **Lab ID: 92561801002** Collected: 09/14/21 16:39 Received: 09/16/21 12:00 Matrix: Air

Parameters	Results	Units	Report			Prepared	Analyzed	CAS No.	Qual
			Limit	MDL	DF				
TO15 MSV AIR									
Analytical Method: TO-15									
Pace Analytical Services - Minneapolis									
1,1,2,2-Tetrachloroethane	<0.60	ug/m3	2.3	0.60	1.61		09/23/21 22:10	79-34-5	
Tetrachloroethene	1.2	ug/m3	1.1	0.47	1.61		09/23/21 22:10	127-18-4	
Tetrahydrofuran	<0.29	ug/m3	0.97	0.29	1.61		09/23/21 22:10	109-99-9	
Toluene	2.9	ug/m3	1.2	0.39	1.61		09/23/21 22:10	108-88-3	
1,2,4-Trichlorobenzene	<7.9	ug/m3	12.1	7.9	1.61		09/23/21 22:10	120-82-1	
1,1,1-Trichloroethane	<0.30	ug/m3	1.8	0.30	1.61		09/23/21 22:10	71-55-6	
1,1,2-Trichloroethane	<0.32	ug/m3	0.89	0.32	1.61		09/23/21 22:10	79-00-5	
Trichloroethene	101	ug/m3	0.88	0.32	1.61		09/23/21 22:10	79-01-6	
Trichlorofluoromethane	7.3	ug/m3	1.8	0.38	1.61		09/23/21 22:10	75-69-4	
1,1,2-Trichlorotrifluoroethane	0.63J	ug/m3	2.5	0.47	1.61		09/23/21 22:10	76-13-1	
1,2,4-Trimethylbenzene	<0.57	ug/m3	1.6	0.57	1.61		09/23/21 22:10	95-63-6	
1,3,5-Trimethylbenzene	<0.47	ug/m3	1.6	0.47	1.61		09/23/21 22:10	108-67-8	
Vinyl acetate	<0.33	ug/m3	1.2	0.33	1.61		09/23/21 22:10	108-05-4	
Vinyl chloride	<0.14	ug/m3	0.42	0.14	1.61		09/23/21 22:10	75-01-4	
m&p-Xylene	<1.0	ug/m3	2.8	1.0	1.61		09/23/21 22:10	179601-23-1	
o-Xylene	<0.44	ug/m3	1.4	0.44	1.61		09/23/21 22:10	95-47-6	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.

ANALYTICAL RESULTS

Project: CONBRACO
Pace Project No.: 92561801

Sample: IA-1 Lab ID: 92561801003 Collected: 09/14/21 18:15 Received: 09/16/21 12:00 Matrix: Air

Parameters	Results	Units	Report			Prepared	Analyzed	CAS No.	Qual
			Limit	MDL	DF				
TO15 MSV AIR									
Analytical Method: TO-15									
Pace Analytical Services - Minneapolis									
Acetone	48.9	ug/m3	9.7	2.9	1.61		09/23/21 16:20	67-64-1	
Benzene	0.22J	ug/m3	0.52	0.18	1.61		09/23/21 16:20	71-43-2	
Benzyl chloride	<1.4	ug/m3	4.2	1.4	1.61		09/23/21 16:20	100-44-7	
Bromodichloromethane	<0.38	ug/m3	2.2	0.38	1.61		09/23/21 16:20	75-27-4	
Bromoform	<2.6	ug/m3	8.5	2.6	1.61		09/23/21 16:20	75-25-2	
Bromomethane	<0.24	ug/m3	1.3	0.24	1.61		09/23/21 16:20	74-83-9	
1,3-Butadiene	<0.19	ug/m3	0.72	0.19	1.61		09/23/21 16:20	106-99-0	
2-Butanone (MEK)	5.9	ug/m3	4.8	0.75	1.61		09/23/21 16:20	78-93-3	
Carbon disulfide	<0.21	ug/m3	1.0	0.21	1.61		09/23/21 16:20	75-15-0	
Carbon tetrachloride	<0.45	ug/m3	2.1	0.45	1.61		09/23/21 16:20	56-23-5	
Chlorobenzene	<0.25	ug/m3	1.5	0.25	1.61		09/23/21 16:20	108-90-7	
Chloroethane	<0.36	ug/m3	0.86	0.36	1.61		09/23/21 16:20	75-00-3	
Chloroform	<0.29	ug/m3	0.80	0.29	1.61		09/23/21 16:20	67-66-3	
Chloromethane	2.4	ug/m3	0.68	0.14	1.61		09/23/21 16:20	74-87-3	
Cyclohexane	<0.36	ug/m3	2.8	0.36	1.61		09/23/21 16:20	110-82-7	
Dibromochloromethane	<0.83	ug/m3	2.8	0.83	1.61		09/23/21 16:20	124-48-1	
1,2-Dibromoethane (EDB)	<0.48	ug/m3	1.3	0.48	1.61		09/23/21 16:20	106-93-4	
1,2-Dichlorobenzene	<0.65	ug/m3	4.9	0.65	1.61		09/23/21 16:20	95-50-1	
1,3-Dichlorobenzene	<0.82	ug/m3	4.9	0.82	1.61		09/23/21 16:20	541-73-1	
1,4-Dichlorobenzene	<1.4	ug/m3	4.9	1.4	1.61		09/23/21 16:20	106-46-7	
Dichlorodifluoromethane	2.5	ug/m3	1.6	0.30	1.61		09/23/21 16:20	75-71-8	
1,1-Dichloroethane	<0.27	ug/m3	1.3	0.27	1.61		09/23/21 16:20	75-34-3	
1,2-Dichloroethane	<0.31	ug/m3	1.3	0.31	1.61		09/23/21 16:20	107-06-2	
1,1-Dichloroethene	<0.22	ug/m3	1.3	0.22	1.61		09/23/21 16:20	75-35-4	
cis-1,2-Dichloroethene	1.6	ug/m3	1.3	0.31	1.61		09/23/21 16:20	156-59-2	
trans-1,2-Dichloroethene	<0.27	ug/m3	1.3	0.27	1.61		09/23/21 16:20	156-60-5	
1,2-Dichloropropane	<0.43	ug/m3	1.5	0.43	1.61		09/23/21 16:20	78-87-5	
cis-1,3-Dichloropropene	<0.41	ug/m3	3.7	0.41	1.61		09/23/21 16:20	10061-01-5	
trans-1,3-Dichloropropene	<0.88	ug/m3	3.7	0.88	1.61		09/23/21 16:20	10061-02-6	
Dichlorotetrafluoroethane	<0.33	ug/m3	2.3	0.33	1.61		09/23/21 16:20	76-14-2	
Ethanol	22.5	ug/m3	3.1	0.95	1.61		09/23/21 16:20	64-17-5	
Ethyl acetate	2.1	ug/m3	1.2	0.21	1.61		09/23/21 16:20	141-78-6	
Ethylbenzene	<0.50	ug/m3	1.4	0.50	1.61		09/23/21 16:20	100-41-4	
4-Ethyltoluene	<0.76	ug/m3	4.0	0.76	1.61		09/23/21 16:20	622-96-8	
n-Heptane	<0.29	ug/m3	1.3	0.29	1.61		09/23/21 16:20	142-82-5	
Hexachloro-1,3-butadiene	<2.0	ug/m3	8.7	2.0	1.61		09/23/21 16:20	87-68-3	
n-Hexane	0.95J	ug/m3	1.2	0.31	1.61		09/23/21 16:20	110-54-3	
2-Hexanone	3.0J	ug/m3	6.7	0.71	1.61		09/23/21 16:20	591-78-6	
Methylene Chloride	<0.95	ug/m3	5.7	0.95	1.61		09/23/21 16:20	75-09-2	
4-Methyl-2-pentanone (MIBK)	6.2J	ug/m3	6.7	0.52	1.61		09/23/21 16:20	108-10-1	
Methyl-tert-butyl ether	<0.20	ug/m3	5.9	0.20	1.61		09/23/21 16:20	1634-04-4	
Naphthalene	<3.5	ug/m3	4.3	3.5	1.61		09/23/21 16:20	91-20-3	
2-Propanol	13.9	ug/m3	4.0	0.82	1.61		09/23/21 16:20	67-63-0	
Propylene	<0.21	ug/m3	1.4	0.21	1.61		09/23/21 16:20	115-07-1	
Styrene	<0.62	ug/m3	3.5	0.62	1.61		09/23/21 16:20	100-42-5	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.

ANALYTICAL RESULTS

Project: CONBRACO

Pace Project No.: 92561801

Sample: IA-1 **Lab ID: 92561801003** Collected: 09/14/21 18:15 Received: 09/16/21 12:00 Matrix: Air

Parameters	Results	Units	Report			Prepared	Analyzed	CAS No.	Qual
			Limit	MDL	DF				
TO15 MSV AIR									
Analytical Method: TO-15									
Pace Analytical Services - Minneapolis									
1,1,2,2-Tetrachloroethane	<0.60	ug/m3	2.3	0.60	1.61		09/23/21 16:20	79-34-5	
Tetrachloroethene	2.2	ug/m3	1.1	0.47	1.61		09/23/21 16:20	127-18-4	
Tetrahydrofuran	<0.29	ug/m3	0.97	0.29	1.61		09/23/21 16:20	109-99-9	
Toluene	1.4	ug/m3	1.2	0.39	1.61		09/23/21 16:20	108-88-3	
1,2,4-Trichlorobenzene	<7.9	ug/m3	12.1	7.9	1.61		09/23/21 16:20	120-82-1	
1,1,1-Trichloroethane	<0.30	ug/m3	1.8	0.30	1.61		09/23/21 16:20	71-55-6	
1,1,2-Trichloroethane	<0.32	ug/m3	0.89	0.32	1.61		09/23/21 16:20	79-00-5	
Trichloroethene	159	ug/m3	0.88	0.32	1.61		09/23/21 16:20	79-01-6	
Trichlorofluoromethane	3.6	ug/m3	1.8	0.38	1.61		09/23/21 16:20	75-69-4	
1,1,2-Trichlorotrifluoroethane	1.9J	ug/m3	2.5	0.47	1.61		09/23/21 16:20	76-13-1	
1,2,4-Trimethylbenzene	<0.57	ug/m3	1.6	0.57	1.61		09/23/21 16:20	95-63-6	
1,3,5-Trimethylbenzene	<0.47	ug/m3	1.6	0.47	1.61		09/23/21 16:20	108-67-8	
Vinyl acetate	<0.33	ug/m3	1.2	0.33	1.61		09/23/21 16:20	108-05-4	
Vinyl chloride	<0.14	ug/m3	0.42	0.14	1.61		09/23/21 16:20	75-01-4	
m&p-Xylene	<1.0	ug/m3	2.8	1.0	1.61		09/23/21 16:20	179601-23-1	
o-Xylene	<0.44	ug/m3	1.4	0.44	1.61		09/23/21 16:20	95-47-6	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.

ANALYTICAL RESULTS

Project: CONBRACO

Pace Project No.: 92561801

Sample: IA-12 Lab ID: 92561801004 Collected: 09/14/21 17:35 Received: 09/16/21 12:00 Matrix: Air

Parameters	Results	Units	Report			Prepared	Analyzed	CAS No.	Qual
			Limit	MDL	DF				
TO15 MSV AIR									
Analytical Method: TO-15									
Pace Analytical Services - Minneapolis									
Acetone	511	ug/m3	9.4	2.8	1.55		09/23/21 17:17	67-64-1	
Benzene	0.40J	ug/m3	0.50	0.18	1.55		09/23/21 17:17	71-43-2	
Benzyl chloride	<1.4	ug/m3	4.1	1.4	1.55		09/23/21 17:17	100-44-7	
Bromodichloromethane	<0.37	ug/m3	2.1	0.37	1.55		09/23/21 17:17	75-27-4	
Bromoform	<2.5	ug/m3	8.1	2.5	1.55		09/23/21 17:17	75-25-2	
Bromomethane	<0.23	ug/m3	1.2	0.23	1.55		09/23/21 17:17	74-83-9	
1,3-Butadiene	<0.19	ug/m3	0.70	0.19	1.55		09/23/21 17:17	106-99-0	
2-Butanone (MEK)	4.5J	ug/m3	4.6	0.72	1.55		09/23/21 17:17	78-93-3	
Carbon disulfide	<0.20	ug/m3	0.98	0.20	1.55		09/23/21 17:17	75-15-0	
Carbon tetrachloride	<0.43	ug/m3	2.0	0.43	1.55		09/23/21 17:17	56-23-5	
Chlorobenzene	<0.24	ug/m3	1.5	0.24	1.55		09/23/21 17:17	108-90-7	
Chloroethane	<0.35	ug/m3	0.83	0.35	1.55		09/23/21 17:17	75-00-3	
Chloroform	<0.28	ug/m3	0.77	0.28	1.55		09/23/21 17:17	67-66-3	
Chloromethane	1.2	ug/m3	0.65	0.13	1.55		09/23/21 17:17	74-87-3	
Cyclohexane	0.57J	ug/m3	2.7	0.34	1.55		09/23/21 17:17	110-82-7	
Dibromochloromethane	<0.80	ug/m3	2.7	0.80	1.55		09/23/21 17:17	124-48-1	
1,2-Dibromoethane (EDB)	<0.46	ug/m3	1.2	0.46	1.55		09/23/21 17:17	106-93-4	
1,2-Dichlorobenzene	<0.63	ug/m3	4.7	0.63	1.55		09/23/21 17:17	95-50-1	
1,3-Dichlorobenzene	<0.79	ug/m3	4.7	0.79	1.55		09/23/21 17:17	541-73-1	
1,4-Dichlorobenzene	<1.4	ug/m3	4.7	1.4	1.55		09/23/21 17:17	106-46-7	
Dichlorodifluoromethane	2.6	ug/m3	1.6	0.29	1.55		09/23/21 17:17	75-71-8	
1,1-Dichloroethane	<0.26	ug/m3	1.3	0.26	1.55		09/23/21 17:17	75-34-3	
1,2-Dichloroethane	<0.30	ug/m3	1.3	0.30	1.55		09/23/21 17:17	107-06-2	
1,1-Dichloroethene	<0.21	ug/m3	1.2	0.21	1.55		09/23/21 17:17	75-35-4	
cis-1,2-Dichloroethene	<0.30	ug/m3	1.2	0.30	1.55		09/23/21 17:17	156-59-2	
trans-1,2-Dichloroethene	<0.26	ug/m3	1.2	0.26	1.55		09/23/21 17:17	156-60-5	
1,2-Dichloropropane	<0.42	ug/m3	1.5	0.42	1.55		09/23/21 17:17	78-87-5	
cis-1,3-Dichloropropene	<0.40	ug/m3	3.6	0.40	1.55		09/23/21 17:17	10061-01-5	
trans-1,3-Dichloropropene	<0.84	ug/m3	3.6	0.84	1.55		09/23/21 17:17	10061-02-6	
Dichlorotetrafluoroethane	<0.31	ug/m3	2.2	0.31	1.55		09/23/21 17:17	76-14-2	
Ethanol	299	ug/m3	3.0	0.92	1.55		09/23/21 17:17	64-17-5	
Ethyl acetate	1.4	ug/m3	1.1	0.20	1.55		09/23/21 17:17	141-78-6	
Ethylbenzene	0.65J	ug/m3	1.4	0.48	1.55		09/23/21 17:17	100-41-4	
4-Ethyltoluene	<0.73	ug/m3	3.9	0.73	1.55		09/23/21 17:17	622-96-8	
n-Heptane	<0.28	ug/m3	1.3	0.28	1.55		09/23/21 17:17	142-82-5	
Hexachloro-1,3-butadiene	<1.9	ug/m3	8.4	1.9	1.55		09/23/21 17:17	87-68-3	
n-Hexane	0.98J	ug/m3	1.1	0.30	1.55		09/23/21 17:17	110-54-3	
2-Hexanone	<0.69	ug/m3	6.4	0.69	1.55		09/23/21 17:17	591-78-6	
Methylene Chloride	0.92J	ug/m3	5.5	0.92	1.55		09/23/21 17:17	75-09-2	
4-Methyl-2-pentanone (MIBK)	<0.50	ug/m3	6.4	0.50	1.55		09/23/21 17:17	108-10-1	
Methyl-tert-butyl ether	<0.20	ug/m3	5.7	0.20	1.55		09/23/21 17:17	1634-04-4	
Naphthalene	<3.4	ug/m3	4.1	3.4	1.55		09/23/21 17:17	91-20-3	
2-Propanol	207	ug/m3	3.9	0.79	1.55		09/23/21 17:17	67-63-0	
Propylene	<0.20	ug/m3	1.4	0.20	1.55		09/23/21 17:17	115-07-1	
Styrene	<0.60	ug/m3	3.4	0.60	1.55		09/23/21 17:17	100-42-5	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.

ANALYTICAL RESULTS

Project: CONBRACO
Pace Project No.: 92561801

Sample: IA-12 Lab ID: 92561801004 Collected: 09/14/21 17:35 Received: 09/16/21 12:00 Matrix: Air

Parameters	Results	Units	Report			Prepared	Analyzed	CAS No.	Qual
			Limit	MDL	DF				
TO15 MSV AIR									
Analytical Method: TO-15									
Pace Analytical Services - Minneapolis									
1,1,2,2-Tetrachloroethane	<0.58	ug/m3	2.2	0.58	1.55		09/23/21 17:17	79-34-5	
Tetrachloroethene	1.7	ug/m3	1.1	0.45	1.55		09/23/21 17:17	127-18-4	
Tetrahydrofuran	<0.28	ug/m3	0.93	0.28	1.55		09/23/21 17:17	109-99-9	
Toluene	1.9	ug/m3	1.2	0.38	1.55		09/23/21 17:17	108-88-3	
1,2,4-Trichlorobenzene	<7.6	ug/m3	11.7	7.6	1.55		09/23/21 17:17	120-82-1	
1,1,1-Trichloroethane	<0.29	ug/m3	1.7	0.29	1.55		09/23/21 17:17	71-55-6	
1,1,2-Trichloroethane	<0.31	ug/m3	0.86	0.31	1.55		09/23/21 17:17	79-00-5	
Trichloroethene	69.0	ug/m3	0.85	0.30	1.55		09/23/21 17:17	79-01-6	
Trichlorofluoromethane	13.6	ug/m3	1.8	0.36	1.55		09/23/21 17:17	75-69-4	
1,1,2-Trichlorotrifluoroethane	1.4J	ug/m3	2.4	0.45	1.55		09/23/21 17:17	76-13-1	
1,2,4-Trimethylbenzene	<0.55	ug/m3	1.5	0.55	1.55		09/23/21 17:17	95-63-6	
1,3,5-Trimethylbenzene	<0.45	ug/m3	1.5	0.45	1.55		09/23/21 17:17	108-67-8	
Vinyl acetate	<0.32	ug/m3	1.1	0.32	1.55		09/23/21 17:17	108-05-4	
Vinyl chloride	<0.13	ug/m3	0.40	0.13	1.55		09/23/21 17:17	75-01-4	
m&p-Xylene	2.0J	ug/m3	2.7	1.0	1.55		09/23/21 17:17	179601-23-1	
o-Xylene	0.75J	ug/m3	1.4	0.42	1.55		09/23/21 17:17	95-47-6	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.

ANALYTICAL RESULTS

Project: CONBRACO

Pace Project No.: 92561801

Sample: IA-2 Lab ID: 92561801005 Collected: 09/14/21 18:16 Received: 09/16/21 12:00 Matrix: Air

Parameters	Results	Units	Report			Prepared	Analyzed	CAS No.	Qual
			Limit	MDL	DF				
TO15 MSV AIR									
Analytical Method: TO-15									
Pace Analytical Services - Minneapolis									
Acetone	48.2	ug/m3	9.9	3.0	1.64		09/23/21 15:21	67-64-1	
Benzene	<0.19	ug/m3	0.53	0.19	1.64		09/23/21 15:21	71-43-2	
Benzyl chloride	<1.5	ug/m3	4.3	1.5	1.64		09/23/21 15:21	100-44-7	
Bromodichloromethane	<0.39	ug/m3	2.2	0.39	1.64		09/23/21 15:21	75-27-4	
Bromoform	<2.7	ug/m3	8.6	2.7	1.64		09/23/21 15:21	75-25-2	
Bromomethane	<0.25	ug/m3	1.3	0.25	1.64		09/23/21 15:21	74-83-9	
1,3-Butadiene	<0.20	ug/m3	0.74	0.20	1.64		09/23/21 15:21	106-99-0	
2-Butanone (MEK)	6.4	ug/m3	4.9	0.76	1.64		09/23/21 15:21	78-93-3	
Carbon disulfide	<0.21	ug/m3	1.0	0.21	1.64		09/23/21 15:21	75-15-0	
Carbon tetrachloride	<0.46	ug/m3	2.1	0.46	1.64		09/23/21 15:21	56-23-5	
Chlorobenzene	<0.25	ug/m3	1.5	0.25	1.64		09/23/21 15:21	108-90-7	
Chloroethane	<0.37	ug/m3	0.88	0.37	1.64		09/23/21 15:21	75-00-3	
Chloroform	<0.30	ug/m3	0.81	0.30	1.64		09/23/21 15:21	67-66-3	
Chloromethane	2.0	ug/m3	0.69	0.14	1.64		09/23/21 15:21	74-87-3	
Cyclohexane	<0.36	ug/m3	2.9	0.36	1.64		09/23/21 15:21	110-82-7	
Dibromochloromethane	<0.84	ug/m3	2.8	0.84	1.64		09/23/21 15:21	124-48-1	
1,2-Dibromoethane (EDB)	<0.49	ug/m3	1.3	0.49	1.64		09/23/21 15:21	106-93-4	
1,2-Dichlorobenzene	<0.66	ug/m3	5.0	0.66	1.64		09/23/21 15:21	95-50-1	
1,3-Dichlorobenzene	<0.83	ug/m3	5.0	0.83	1.64		09/23/21 15:21	541-73-1	
1,4-Dichlorobenzene	<1.4	ug/m3	5.0	1.4	1.64		09/23/21 15:21	106-46-7	
Dichlorodifluoromethane	2.4	ug/m3	1.7	0.31	1.64		09/23/21 15:21	75-71-8	
1,1-Dichloroethane	<0.27	ug/m3	1.3	0.27	1.64		09/23/21 15:21	75-34-3	
1,2-Dichloroethane	<0.32	ug/m3	1.3	0.32	1.64		09/23/21 15:21	107-06-2	
1,1-Dichloroethene	<0.23	ug/m3	1.3	0.23	1.64		09/23/21 15:21	75-35-4	
cis-1,2-Dichloroethene	1.8	ug/m3	1.3	0.32	1.64		09/23/21 15:21	156-59-2	
trans-1,2-Dichloroethene	<0.28	ug/m3	1.3	0.28	1.64		09/23/21 15:21	156-60-5	
1,2-Dichloropropane	<0.44	ug/m3	1.5	0.44	1.64		09/23/21 15:21	78-87-5	
cis-1,3-Dichloropropene	<0.42	ug/m3	3.8	0.42	1.64		09/23/21 15:21	10061-01-5	
trans-1,3-Dichloropropene	<0.89	ug/m3	3.8	0.89	1.64		09/23/21 15:21	10061-02-6	
Dichlorotetrafluoroethane	<0.33	ug/m3	2.3	0.33	1.64		09/23/21 15:21	76-14-2	
Ethanol	22.2	ug/m3	3.1	0.97	1.64		09/23/21 15:21	64-17-5	
Ethyl acetate	2.0	ug/m3	1.2	0.21	1.64		09/23/21 15:21	141-78-6	
Ethylbenzene	<0.51	ug/m3	1.4	0.51	1.64		09/23/21 15:21	100-41-4	
4-Ethyltoluene	<0.77	ug/m3	4.1	0.77	1.64		09/23/21 15:21	622-96-8	
n-Heptane	<0.30	ug/m3	1.4	0.30	1.64		09/23/21 15:21	142-82-5	
Hexachloro-1,3-butadiene	<2.0	ug/m3	8.9	2.0	1.64		09/23/21 15:21	87-68-3	
n-Hexane	0.96J	ug/m3	1.2	0.31	1.64		09/23/21 15:21	110-54-3	
2-Hexanone	<0.72	ug/m3	6.8	0.72	1.64		09/23/21 15:21	591-78-6	
Methylene Chloride	<0.97	ug/m3	5.8	0.97	1.64		09/23/21 15:21	75-09-2	
4-Methyl-2-pentanone (MIBK)	9.3	ug/m3	6.8	0.53	1.64		09/23/21 15:21	108-10-1	
Methyl-tert-butyl ether	<0.21	ug/m3	6.0	0.21	1.64		09/23/21 15:21	1634-04-4	
Naphthalene	3.6J	ug/m3	4.4	3.6	1.64		09/23/21 15:21	91-20-3	
2-Propanol	10.7	ug/m3	4.1	0.83	1.64		09/23/21 15:21	67-63-0	
Propylene	<0.21	ug/m3	1.4	0.21	1.64		09/23/21 15:21	115-07-1	
Styrene	<0.63	ug/m3	3.6	0.63	1.64		09/23/21 15:21	100-42-5	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.

ANALYTICAL RESULTS

Project: CONBRACO
Pace Project No.: 92561801

Sample: IA-2 Lab ID: 92561801005 Collected: 09/14/21 18:16 Received: 09/16/21 12:00 Matrix: Air

Parameters	Results	Units	Report			Prepared	Analyzed	CAS No.	Qual
			Limit	MDL	DF				
TO15 MSV AIR									
Analytical Method: TO-15									
Pace Analytical Services - Minneapolis									
1,1,2,2-Tetrachloroethane	<0.61	ug/m3	2.3	0.61	1.64		09/23/21 15:21	79-34-5	
Tetrachloroethene	2.1	ug/m3	1.1	0.48	1.64		09/23/21 15:21	127-18-4	
Tetrahydrofuran	<0.30	ug/m3	0.98	0.30	1.64		09/23/21 15:21	109-99-9	
Toluene	1.5	ug/m3	1.3	0.40	1.64		09/23/21 15:21	108-88-3	
1,2,4-Trichlorobenzene	<8.0	ug/m3	12.4	8.0	1.64		09/23/21 15:21	120-82-1	
1,1,1-Trichloroethane	<0.31	ug/m3	1.8	0.31	1.64		09/23/21 15:21	71-55-6	
1,1,2-Trichloroethane	<0.32	ug/m3	0.91	0.32	1.64		09/23/21 15:21	79-00-5	
Trichloroethene	165	ug/m3	0.90	0.32	1.64		09/23/21 15:21	79-01-6	
Trichlorofluoromethane	4.4	ug/m3	1.9	0.38	1.64		09/23/21 15:21	75-69-4	
1,1,2-Trichlorotrifluoroethane	1.6J	ug/m3	2.6	0.47	1.64		09/23/21 15:21	76-13-1	
1,2,4-Trimethylbenzene	<0.58	ug/m3	1.6	0.58	1.64		09/23/21 15:21	95-63-6	
1,3,5-Trimethylbenzene	<0.48	ug/m3	1.6	0.48	1.64		09/23/21 15:21	108-67-8	
Vinyl acetate	<0.34	ug/m3	1.2	0.34	1.64		09/23/21 15:21	108-05-4	
Vinyl chloride	<0.14	ug/m3	0.43	0.14	1.64		09/23/21 15:21	75-01-4	
m&p-Xylene	<1.1	ug/m3	2.9	1.1	1.64		09/23/21 15:21	179601-23-1	
o-Xylene	<0.44	ug/m3	1.4	0.44	1.64		09/23/21 15:21	95-47-6	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.

ANALYTICAL RESULTS

Project: CONBRACO

Pace Project No.: 92561801

Sample: IA-8 **Lab ID: 92561801006** Collected: 09/14/21 16:52 Received: 09/16/21 12:00 Matrix: Air

Parameters	Results	Units	Report			Prepared	Analyzed	CAS No.	Qual
			Limit	MDL	DF				
TO15 MSV AIR									
Analytical Method: TO-15									
Pace Analytical Services - Minneapolis									
Acetone	65.1	ug/m3	9.7	2.9	1.61		09/23/21 18:47	67-64-1	
Benzene	0.28J	ug/m3	0.52	0.18	1.61		09/23/21 18:47	71-43-2	
Benzyl chloride	<1.4	ug/m3	4.2	1.4	1.61		09/23/21 18:47	100-44-7	
Bromodichloromethane	<0.38	ug/m3	2.2	0.38	1.61		09/23/21 18:47	75-27-4	
Bromoform	<2.6	ug/m3	8.5	2.6	1.61		09/23/21 18:47	75-25-2	
Bromomethane	<0.24	ug/m3	1.3	0.24	1.61		09/23/21 18:47	74-83-9	
1,3-Butadiene	<0.19	ug/m3	0.72	0.19	1.61		09/23/21 18:47	106-99-0	
2-Butanone (MEK)	5.3	ug/m3	4.8	0.75	1.61		09/23/21 18:47	78-93-3	
Carbon disulfide	<0.21	ug/m3	1.0	0.21	1.61		09/23/21 18:47	75-15-0	
Carbon tetrachloride	<0.45	ug/m3	2.1	0.45	1.61		09/23/21 18:47	56-23-5	
Chlorobenzene	<0.25	ug/m3	1.5	0.25	1.61		09/23/21 18:47	108-90-7	
Chloroethane	<0.36	ug/m3	0.86	0.36	1.61		09/23/21 18:47	75-00-3	
Chloroform	<0.29	ug/m3	0.80	0.29	1.61		09/23/21 18:47	67-66-3	
Chloromethane	0.70	ug/m3	0.68	0.14	1.61		09/23/21 18:47	74-87-3	
Cyclohexane	<0.36	ug/m3	2.8	0.36	1.61		09/23/21 18:47	110-82-7	
Dibromochloromethane	<0.83	ug/m3	2.8	0.83	1.61		09/23/21 18:47	124-48-1	
1,2-Dibromoethane (EDB)	<0.48	ug/m3	1.3	0.48	1.61		09/23/21 18:47	106-93-4	
1,2-Dichlorobenzene	<0.65	ug/m3	4.9	0.65	1.61		09/23/21 18:47	95-50-1	
1,3-Dichlorobenzene	<0.82	ug/m3	4.9	0.82	1.61		09/23/21 18:47	541-73-1	
1,4-Dichlorobenzene	<1.4	ug/m3	4.9	1.4	1.61		09/23/21 18:47	106-46-7	
Dichlorodifluoromethane	2.0	ug/m3	1.6	0.30	1.61		09/23/21 18:47	75-71-8	
1,1-Dichloroethane	<0.27	ug/m3	1.3	0.27	1.61		09/23/21 18:47	75-34-3	
1,2-Dichloroethane	<0.31	ug/m3	1.3	0.31	1.61		09/23/21 18:47	107-06-2	
1,1-Dichloroethene	<0.22	ug/m3	1.3	0.22	1.61		09/23/21 18:47	75-35-4	
cis-1,2-Dichloroethene	0.74J	ug/m3	1.3	0.31	1.61		09/23/21 18:47	156-59-2	
trans-1,2-Dichloroethene	<0.27	ug/m3	1.3	0.27	1.61		09/23/21 18:47	156-60-5	
1,2-Dichloropropane	<0.43	ug/m3	1.5	0.43	1.61		09/23/21 18:47	78-87-5	
cis-1,3-Dichloropropene	<0.41	ug/m3	3.7	0.41	1.61		09/23/21 18:47	10061-01-5	
trans-1,3-Dichloropropene	<0.88	ug/m3	3.7	0.88	1.61		09/23/21 18:47	10061-02-6	
Dichlorotetrafluoroethane	<0.33	ug/m3	2.3	0.33	1.61		09/23/21 18:47	76-14-2	
Ethanol	20.3	ug/m3	3.1	0.95	1.61		09/23/21 18:47	64-17-5	
Ethyl acetate	6.3	ug/m3	1.2	0.21	1.61		09/23/21 18:47	141-78-6	
Ethylbenzene	<0.50	ug/m3	1.4	0.50	1.61		09/23/21 18:47	100-41-4	
4-Ethyltoluene	<0.76	ug/m3	4.0	0.76	1.61		09/23/21 18:47	622-96-8	
n-Heptane	<0.29	ug/m3	1.3	0.29	1.61		09/23/21 18:47	142-82-5	
Hexachloro-1,3-butadiene	<2.0	ug/m3	8.7	2.0	1.61		09/23/21 18:47	87-68-3	
n-Hexane	0.64J	ug/m3	1.2	0.31	1.61		09/23/21 18:47	110-54-3	
2-Hexanone	<0.71	ug/m3	6.7	0.71	1.61		09/23/21 18:47	591-78-6	
Methylene Chloride	<0.95	ug/m3	5.7	0.95	1.61		09/23/21 18:47	75-09-2	
4-Methyl-2-pentanone (MIBK)	<0.52	ug/m3	6.7	0.52	1.61		09/23/21 18:47	108-10-1	
Methyl-tert-butyl ether	<0.20	ug/m3	5.9	0.20	1.61		09/23/21 18:47	1634-04-4	
Naphthalene	<3.5	ug/m3	4.3	3.5	1.61		09/23/21 18:47	91-20-3	
2-Propanol	36.3	ug/m3	4.0	0.82	1.61		09/23/21 18:47	67-63-0	
Propylene	<0.21	ug/m3	1.4	0.21	1.61		09/23/21 18:47	115-07-1	
Styrene	<0.62	ug/m3	3.5	0.62	1.61		09/23/21 18:47	100-42-5	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.

ANALYTICAL RESULTS

Project: CONBRACO
Pace Project No.: 92561801

Sample: IA-8 **Lab ID: 92561801006** Collected: 09/14/21 16:52 Received: 09/16/21 12:00 Matrix: Air

Parameters	Results	Units	Report			Prepared	Analyzed	CAS No.	Qual
			Limit	MDL	DF				
TO15 MSV AIR									
Analytical Method: TO-15									
Pace Analytical Services - Minneapolis									
1,1,2,2-Tetrachloroethane	<0.60	ug/m3	2.3	0.60	1.61		09/23/21 18:47	79-34-5	
Tetrachloroethene	3.4	ug/m3	1.1	0.47	1.61		09/23/21 18:47	127-18-4	
Tetrahydrofuran	<0.29	ug/m3	0.97	0.29	1.61		09/23/21 18:47	109-99-9	
Toluene	1.3	ug/m3	1.2	0.39	1.61		09/23/21 18:47	108-88-3	
1,2,4-Trichlorobenzene	<7.9	ug/m3	12.1	7.9	1.61		09/23/21 18:47	120-82-1	
1,1,1-Trichloroethane	<0.30	ug/m3	1.8	0.30	1.61		09/23/21 18:47	71-55-6	
1,1,2-Trichloroethane	<0.32	ug/m3	0.89	0.32	1.61		09/23/21 18:47	79-00-5	
Trichloroethene	256	ug/m3	0.88	0.32	1.61		09/23/21 18:47	79-01-6	
Trichlorofluoromethane	2.7	ug/m3	1.8	0.38	1.61		09/23/21 18:47	75-69-4	
1,1,2-Trichlorotrifluoroethane	1.5J	ug/m3	2.5	0.47	1.61		09/23/21 18:47	76-13-1	
1,2,4-Trimethylbenzene	0.66J	ug/m3	1.6	0.57	1.61		09/23/21 18:47	95-63-6	
1,3,5-Trimethylbenzene	<0.47	ug/m3	1.6	0.47	1.61		09/23/21 18:47	108-67-8	
Vinyl acetate	<0.33	ug/m3	1.2	0.33	1.61		09/23/21 18:47	108-05-4	
Vinyl chloride	<0.14	ug/m3	0.42	0.14	1.61		09/23/21 18:47	75-01-4	
m&p-Xylene	<1.0	ug/m3	2.8	1.0	1.61		09/23/21 18:47	179601-23-1	
o-Xylene	<0.44	ug/m3	1.4	0.44	1.61		09/23/21 18:47	95-47-6	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.

ANALYTICAL RESULTS

Project: CONBRACO

Pace Project No.: 92561801

Sample: IA-5 Lab ID: 92561801007 Collected: 09/14/21 16:15 Received: 09/16/21 12:00 Matrix: Air

Parameters	Results	Units	Report			Prepared	Analyzed	CAS No.	Qual
			Limit	MDL	DF				
TO15 MSV AIR									
Analytical Method: TO-15									
Pace Analytical Services - Minneapolis									
Acetone	24.6	ug/m3	10.6	3.2	1.75		09/23/21 18:19	67-64-1	
Benzene	<0.20	ug/m3	0.57	0.20	1.75		09/23/21 18:19	71-43-2	
Benzyl chloride	<1.6	ug/m3	4.6	1.6	1.75		09/23/21 18:19	100-44-7	
Bromodichloromethane	<0.41	ug/m3	2.4	0.41	1.75		09/23/21 18:19	75-27-4	
Bromoform	<2.8	ug/m3	9.2	2.8	1.75		09/23/21 18:19	75-25-2	
Bromomethane	<0.26	ug/m3	1.4	0.26	1.75		09/23/21 18:19	74-83-9	
1,3-Butadiene	<0.21	ug/m3	0.79	0.21	1.75		09/23/21 18:19	106-99-0	
2-Butanone (MEK)	3.4J	ug/m3	5.2	0.81	1.75		09/23/21 18:19	78-93-3	
Carbon disulfide	<0.23	ug/m3	1.1	0.23	1.75		09/23/21 18:19	75-15-0	
Carbon tetrachloride	<0.49	ug/m3	2.2	0.49	1.75		09/23/21 18:19	56-23-5	
Chlorobenzene	<0.27	ug/m3	1.6	0.27	1.75		09/23/21 18:19	108-90-7	
Chloroethane	<0.39	ug/m3	0.94	0.39	1.75		09/23/21 18:19	75-00-3	
Chloroform	0.77J	ug/m3	0.87	0.32	1.75		09/23/21 18:19	67-66-3	
Chloromethane	5.2	ug/m3	0.74	0.15	1.75		09/23/21 18:19	74-87-3	
Cyclohexane	<0.39	ug/m3	3.1	0.39	1.75		09/23/21 18:19	110-82-7	
Dibromochloromethane	<0.90	ug/m3	3.0	0.90	1.75		09/23/21 18:19	124-48-1	
1,2-Dibromoethane (EDB)	<0.52	ug/m3	1.4	0.52	1.75		09/23/21 18:19	106-93-4	
1,2-Dichlorobenzene	<0.71	ug/m3	5.4	0.71	1.75		09/23/21 18:19	95-50-1	
1,3-Dichlorobenzene	<0.89	ug/m3	5.4	0.89	1.75		09/23/21 18:19	541-73-1	
1,4-Dichlorobenzene	<1.5	ug/m3	5.4	1.5	1.75		09/23/21 18:19	106-46-7	
Dichlorodifluoromethane	2.1	ug/m3	1.8	0.33	1.75		09/23/21 18:19	75-71-8	
1,1-Dichloroethane	<0.29	ug/m3	1.4	0.29	1.75		09/23/21 18:19	75-34-3	
1,2-Dichloroethane	<0.34	ug/m3	1.4	0.34	1.75		09/23/21 18:19	107-06-2	
1,1-Dichloroethene	<0.24	ug/m3	1.4	0.24	1.75		09/23/21 18:19	75-35-4	
cis-1,2-Dichloroethene	3.9	ug/m3	1.4	0.34	1.75		09/23/21 18:19	156-59-2	
trans-1,2-Dichloroethene	<0.29	ug/m3	1.4	0.29	1.75		09/23/21 18:19	156-60-5	
1,2-Dichloropropane	<0.47	ug/m3	1.6	0.47	1.75		09/23/21 18:19	78-87-5	
cis-1,3-Dichloropropene	<0.45	ug/m3	4.0	0.45	1.75		09/23/21 18:19	10061-01-5	
trans-1,3-Dichloropropene	<0.95	ug/m3	4.0	0.95	1.75		09/23/21 18:19	10061-02-6	
Dichlorotetrafluoroethane	<0.35	ug/m3	2.5	0.35	1.75		09/23/21 18:19	76-14-2	
Ethanol	171	ug/m3	3.4	1.0	1.75		09/23/21 18:19	64-17-5	
Ethyl acetate	4.6	ug/m3	1.3	0.23	1.75		09/23/21 18:19	141-78-6	
Ethylbenzene	<0.54	ug/m3	1.5	0.54	1.75		09/23/21 18:19	100-41-4	
4-Ethyltoluene	<0.83	ug/m3	4.4	0.83	1.75		09/23/21 18:19	622-96-8	
n-Heptane	<0.32	ug/m3	1.5	0.32	1.75		09/23/21 18:19	142-82-5	
Hexachloro-1,3-butadiene	<2.2	ug/m3	9.5	2.2	1.75		09/23/21 18:19	87-68-3	
n-Hexane	0.57J	ug/m3	1.3	0.33	1.75		09/23/21 18:19	110-54-3	
2-Hexanone	<0.77	ug/m3	7.3	0.77	1.75		09/23/21 18:19	591-78-6	
Methylene Chloride	<1.0	ug/m3	6.2	1.0	1.75		09/23/21 18:19	75-09-2	
4-Methyl-2-pentanone (MIBK)	<0.56	ug/m3	7.3	0.56	1.75		09/23/21 18:19	108-10-1	
Methyl-tert-butyl ether	<0.22	ug/m3	6.4	0.22	1.75		09/23/21 18:19	1634-04-4	
Naphthalene	<3.8	ug/m3	4.7	3.8	1.75		09/23/21 18:19	91-20-3	
2-Propanol	6.0	ug/m3	4.4	0.89	1.75		09/23/21 18:19	67-63-0	
Propylene	<0.23	ug/m3	1.5	0.23	1.75		09/23/21 18:19	115-07-1	
Styrene	<0.67	ug/m3	3.8	0.67	1.75		09/23/21 18:19	100-42-5	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.

ANALYTICAL RESULTS

Project: CONBRACO
Pace Project No.: 92561801

Sample: IA-5 Lab ID: 92561801007 Collected: 09/14/21 16:15 Received: 09/16/21 12:00 Matrix: Air

Parameters	Results	Units	Report			Prepared	Analyzed	CAS No.	Qual
			Limit	MDL	DF				
TO15 MSV AIR									
Analytical Method: TO-15									
Pace Analytical Services - Minneapolis									
1,1,2,2-Tetrachloroethane	<0.65	ug/m3	2.4	0.65	1.75		09/23/21 18:19	79-34-5	
Tetrachloroethene	1.3	ug/m3	1.2	0.51	1.75		09/23/21 18:19	127-18-4	
Tetrahydrofuran	<0.32	ug/m3	1.0	0.32	1.75		09/23/21 18:19	109-99-9	
Toluene	2.0	ug/m3	1.3	0.43	1.75		09/23/21 18:19	108-88-3	
1,2,4-Trichlorobenzene	<8.5	ug/m3	13.2	8.5	1.75		09/23/21 18:19	120-82-1	
1,1,1-Trichloroethane	<0.33	ug/m3	1.9	0.33	1.75		09/23/21 18:19	71-55-6	
1,1,2-Trichloroethane	<0.34	ug/m3	0.97	0.34	1.75		09/23/21 18:19	79-00-5	
Trichloroethene	131	ug/m3	0.96	0.34	1.75		09/23/21 18:19	79-01-6	
Trichlorofluoromethane	7.6	ug/m3	2.0	0.41	1.75		09/23/21 18:19	75-69-4	
1,1,2-Trichlorotrifluoroethane	0.71J	ug/m3	2.7	0.51	1.75		09/23/21 18:19	76-13-1	
1,2,4-Trimethylbenzene	<0.62	ug/m3	1.7	0.62	1.75		09/23/21 18:19	95-63-6	
1,3,5-Trimethylbenzene	<0.51	ug/m3	1.7	0.51	1.75		09/23/21 18:19	108-67-8	
Vinyl acetate	<0.36	ug/m3	1.3	0.36	1.75		09/23/21 18:19	108-05-4	
Vinyl chloride	<0.15	ug/m3	0.46	0.15	1.75		09/23/21 18:19	75-01-4	
m&p-Xylene	<1.1	ug/m3	3.1	1.1	1.75		09/23/21 18:19	179601-23-1	
o-Xylene	<0.47	ug/m3	1.5	0.47	1.75		09/23/21 18:19	95-47-6	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.

ANALYTICAL RESULTS

Project: CONBRACO

Pace Project No.: 92561801

Sample: DUP-1 Lab ID: 92561801008 Collected: 09/14/21 17:04 Received: 09/16/21 12:00 Matrix: Air

Parameters	Results	Units	Report			Prepared	Analyzed	CAS No.	Qual
			Limit	MDL	DF				
TO15 MSV AIR									
Analytical Method: TO-15									
Pace Analytical Services - Minneapolis									
Acetone	539	ug/m3	9.9	3.0	1.64		09/23/21 20:45	67-64-1	
Benzene	0.33J	ug/m3	0.53	0.19	1.64		09/23/21 20:45	71-43-2	
Benzyl chloride	<1.5	ug/m3	4.3	1.5	1.64		09/23/21 20:45	100-44-7	
Bromodichloromethane	<0.39	ug/m3	2.2	0.39	1.64		09/23/21 20:45	75-27-4	
Bromoform	<2.7	ug/m3	8.6	2.7	1.64		09/23/21 20:45	75-25-2	
Bromomethane	<0.25	ug/m3	1.3	0.25	1.64		09/23/21 20:45	74-83-9	
1,3-Butadiene	<0.20	ug/m3	0.74	0.20	1.64		09/23/21 20:45	106-99-0	
2-Butanone (MEK)	1.2J	ug/m3	4.9	0.76	1.64		09/23/21 20:45	78-93-3	
Carbon disulfide	<0.21	ug/m3	1.0	0.21	1.64		09/23/21 20:45	75-15-0	
Carbon tetrachloride	<0.46	ug/m3	2.1	0.46	1.64		09/23/21 20:45	56-23-5	
Chlorobenzene	<0.25	ug/m3	1.5	0.25	1.64		09/23/21 20:45	108-90-7	
Chloroethane	<0.37	ug/m3	0.88	0.37	1.64		09/23/21 20:45	75-00-3	
Chloroform	<0.30	ug/m3	0.81	0.30	1.64		09/23/21 20:45	67-66-3	
Chloromethane	1.1	ug/m3	0.69	0.14	1.64		09/23/21 20:45	74-87-3	
Cyclohexane	0.47J	ug/m3	2.9	0.36	1.64		09/23/21 20:45	110-82-7	
Dibromochloromethane	<0.84	ug/m3	2.8	0.84	1.64		09/23/21 20:45	124-48-1	
1,2-Dibromoethane (EDB)	<0.49	ug/m3	1.3	0.49	1.64		09/23/21 20:45	106-93-4	
1,2-Dichlorobenzene	<0.66	ug/m3	5.0	0.66	1.64		09/23/21 20:45	95-50-1	
1,3-Dichlorobenzene	<0.83	ug/m3	5.0	0.83	1.64		09/23/21 20:45	541-73-1	
1,4-Dichlorobenzene	<1.4	ug/m3	5.0	1.4	1.64		09/23/21 20:45	106-46-7	
Dichlorodifluoromethane	2.3	ug/m3	1.7	0.31	1.64		09/23/21 20:45	75-71-8	
1,1-Dichloroethane	<0.27	ug/m3	1.3	0.27	1.64		09/23/21 20:45	75-34-3	
1,2-Dichloroethane	<0.32	ug/m3	1.3	0.32	1.64		09/23/21 20:45	107-06-2	
1,1-Dichloroethene	<0.23	ug/m3	1.3	0.23	1.64		09/23/21 20:45	75-35-4	
cis-1,2-Dichloroethene	<0.32	ug/m3	1.3	0.32	1.64		09/23/21 20:45	156-59-2	
trans-1,2-Dichloroethene	<0.28	ug/m3	1.3	0.28	1.64		09/23/21 20:45	156-60-5	
1,2-Dichloropropane	<0.44	ug/m3	1.5	0.44	1.64		09/23/21 20:45	78-87-5	
cis-1,3-Dichloropropene	<0.42	ug/m3	3.8	0.42	1.64		09/23/21 20:45	10061-01-5	
trans-1,3-Dichloropropene	<0.89	ug/m3	3.8	0.89	1.64		09/23/21 20:45	10061-02-6	
Dichlorotetrafluoroethane	<0.33	ug/m3	2.3	0.33	1.64		09/23/21 20:45	76-14-2	
Ethanol	217	ug/m3	3.1	0.97	1.64		09/23/21 20:45	64-17-5	
Ethyl acetate	1.0J	ug/m3	1.2	0.21	1.64		09/23/21 20:45	141-78-6	
Ethylbenzene	<0.51	ug/m3	1.4	0.51	1.64		09/23/21 20:45	100-41-4	
4-Ethyltoluene	<0.77	ug/m3	4.1	0.77	1.64		09/23/21 20:45	622-96-8	
n-Heptane	<0.30	ug/m3	1.4	0.30	1.64		09/23/21 20:45	142-82-5	
Hexachloro-1,3-butadiene	<2.0	ug/m3	8.9	2.0	1.64		09/23/21 20:45	87-68-3	
n-Hexane	0.84J	ug/m3	1.2	0.31	1.64		09/23/21 20:45	110-54-3	
2-Hexanone	<0.72	ug/m3	6.8	0.72	1.64		09/23/21 20:45	591-78-6	
Methylene Chloride	<0.97	ug/m3	5.8	0.97	1.64		09/23/21 20:45	75-09-2	
4-Methyl-2-pentanone (MIBK)	<0.53	ug/m3	6.8	0.53	1.64		09/23/21 20:45	108-10-1	
Methyl-tert-butyl ether	<0.21	ug/m3	6.0	0.21	1.64		09/23/21 20:45	1634-04-4	
Naphthalene	<3.6	ug/m3	4.4	3.6	1.64		09/23/21 20:45	91-20-3	
2-Propanol	135	ug/m3	4.1	0.83	1.64		09/23/21 20:45	67-63-0	
Propylene	<0.21	ug/m3	1.4	0.21	1.64		09/23/21 20:45	115-07-1	
Styrene	1.5J	ug/m3	3.6	0.63	1.64		09/23/21 20:45	100-42-5	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.

ANALYTICAL RESULTS

Project: CONBRACO
Pace Project No.: 92561801

Sample: DUP-1 **Lab ID: 92561801008** Collected: 09/14/21 17:04 Received: 09/16/21 12:00 Matrix: Air

Parameters	Results	Units	Report			Prepared	Analyzed	CAS No.	Qual
			Limit	MDL	DF				
TO15 MSV AIR									
Analytical Method: TO-15									
Pace Analytical Services - Minneapolis									
1,1,2,2-Tetrachloroethane	<0.61	ug/m3	2.3	0.61	1.64		09/23/21 20:45	79-34-5	
Tetrachloroethene	0.62J	ug/m3	1.1	0.48	1.64		09/23/21 20:45	127-18-4	
Tetrahydrofuran	<0.30	ug/m3	0.98	0.30	1.64		09/23/21 20:45	109-99-9	
Toluene	1.8	ug/m3	1.3	0.40	1.64		09/23/21 20:45	108-88-3	
1,2,4-Trichlorobenzene	<8.0	ug/m3	12.4	8.0	1.64		09/23/21 20:45	120-82-1	
1,1,1-Trichloroethane	<0.31	ug/m3	1.8	0.31	1.64		09/23/21 20:45	71-55-6	
1,1,2-Trichloroethane	<0.32	ug/m3	0.91	0.32	1.64		09/23/21 20:45	79-00-5	
Trichloroethene	49.7	ug/m3	0.90	0.32	1.64		09/23/21 20:45	79-01-6	
Trichlorofluoromethane	8.3	ug/m3	1.9	0.38	1.64		09/23/21 20:45	75-69-4	
1,1,2-Trichlorotrifluoroethane	0.61J	ug/m3	2.6	0.47	1.64		09/23/21 20:45	76-13-1	
1,2,4-Trimethylbenzene	<0.58	ug/m3	1.6	0.58	1.64		09/23/21 20:45	95-63-6	
1,3,5-Trimethylbenzene	<0.48	ug/m3	1.6	0.48	1.64		09/23/21 20:45	108-67-8	
Vinyl acetate	<0.34	ug/m3	1.2	0.34	1.64		09/23/21 20:45	108-05-4	
Vinyl chloride	<0.14	ug/m3	0.43	0.14	1.64		09/23/21 20:45	75-01-4	
m&p-Xylene	1.5J	ug/m3	2.9	1.1	1.64		09/23/21 20:45	179601-23-1	
o-Xylene	0.55J	ug/m3	1.4	0.44	1.64		09/23/21 20:45	95-47-6	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.

ANALYTICAL RESULTS

Project: CONBRACO

Pace Project No.: 92561801

Sample: IA-13 Lab ID: 92561801009 Collected: 09/14/21 17:04 Received: 09/16/21 12:00 Matrix: Air

Parameters	Results	Units	Report			Prepared	Analyzed	CAS No.	Qual
			Limit	MDL	DF				
TO15 MSV AIR									
Analytical Method: TO-15									
Pace Analytical Services - Minneapolis									
Acetone	580	ug/m3	9.7	2.9	1.61		09/23/21 17:46	67-64-1	
Benzene	0.36J	ug/m3	0.52	0.18	1.61		09/23/21 17:46	71-43-2	
Benzyl chloride	<1.4	ug/m3	4.2	1.4	1.61		09/23/21 17:46	100-44-7	
Bromodichloromethane	<0.38	ug/m3	2.2	0.38	1.61		09/23/21 17:46	75-27-4	
Bromoform	<2.6	ug/m3	8.5	2.6	1.61		09/23/21 17:46	75-25-2	
Bromomethane	<0.24	ug/m3	1.3	0.24	1.61		09/23/21 17:46	74-83-9	
1,3-Butadiene	<0.19	ug/m3	0.72	0.19	1.61		09/23/21 17:46	106-99-0	
2-Butanone (MEK)	3.9J	ug/m3	4.8	0.75	1.61		09/23/21 17:46	78-93-3	
Carbon disulfide	<0.21	ug/m3	1.0	0.21	1.61		09/23/21 17:46	75-15-0	
Carbon tetrachloride	<0.45	ug/m3	2.1	0.45	1.61		09/23/21 17:46	56-23-5	
Chlorobenzene	<0.25	ug/m3	1.5	0.25	1.61		09/23/21 17:46	108-90-7	
Chloroethane	<0.36	ug/m3	0.86	0.36	1.61		09/23/21 17:46	75-00-3	
Chloroform	<0.29	ug/m3	0.80	0.29	1.61		09/23/21 17:46	67-66-3	
Chloromethane	1.1	ug/m3	0.68	0.14	1.61		09/23/21 17:46	74-87-3	
Cyclohexane	0.52J	ug/m3	2.8	0.36	1.61		09/23/21 17:46	110-82-7	
Dibromochloromethane	<0.83	ug/m3	2.8	0.83	1.61		09/23/21 17:46	124-48-1	
1,2-Dibromoethane (EDB)	<0.48	ug/m3	1.3	0.48	1.61		09/23/21 17:46	106-93-4	
1,2-Dichlorobenzene	<0.65	ug/m3	4.9	0.65	1.61		09/23/21 17:46	95-50-1	
1,3-Dichlorobenzene	<0.82	ug/m3	4.9	0.82	1.61		09/23/21 17:46	541-73-1	
1,4-Dichlorobenzene	<1.4	ug/m3	4.9	1.4	1.61		09/23/21 17:46	106-46-7	
Dichlorodifluoromethane	2.2	ug/m3	1.6	0.30	1.61		09/23/21 17:46	75-71-8	
1,1-Dichloroethane	<0.27	ug/m3	1.3	0.27	1.61		09/23/21 17:46	75-34-3	
1,2-Dichloroethane	<0.31	ug/m3	1.3	0.31	1.61		09/23/21 17:46	107-06-2	
1,1-Dichloroethene	<0.22	ug/m3	1.3	0.22	1.61		09/23/21 17:46	75-35-4	
cis-1,2-Dichloroethene	<0.31	ug/m3	1.3	0.31	1.61		09/23/21 17:46	156-59-2	
trans-1,2-Dichloroethene	<0.27	ug/m3	1.3	0.27	1.61		09/23/21 17:46	156-60-5	
1,2-Dichloropropane	<0.43	ug/m3	1.5	0.43	1.61		09/23/21 17:46	78-87-5	
cis-1,3-Dichloropropene	<0.41	ug/m3	3.7	0.41	1.61		09/23/21 17:46	10061-01-5	
trans-1,3-Dichloropropene	<0.88	ug/m3	3.7	0.88	1.61		09/23/21 17:46	10061-02-6	
Dichlorotetrafluoroethane	<0.33	ug/m3	2.3	0.33	1.61		09/23/21 17:46	76-14-2	
Ethanol	225	ug/m3	3.1	0.95	1.61		09/23/21 17:46	64-17-5	
Ethyl acetate	0.91J	ug/m3	1.2	0.21	1.61		09/23/21 17:46	141-78-6	
Ethylbenzene	0.50J	ug/m3	1.4	0.50	1.61		09/23/21 17:46	100-41-4	
4-Ethyltoluene	<0.76	ug/m3	4.0	0.76	1.61		09/23/21 17:46	622-96-8	
n-Heptane	<0.29	ug/m3	1.3	0.29	1.61		09/23/21 17:46	142-82-5	
Hexachloro-1,3-butadiene	<2.0	ug/m3	8.7	2.0	1.61		09/23/21 17:46	87-68-3	
n-Hexane	1.2	ug/m3	1.2	0.31	1.61		09/23/21 17:46	110-54-3	
2-Hexanone	<0.71	ug/m3	6.7	0.71	1.61		09/23/21 17:46	591-78-6	
Methylene Chloride	<0.95	ug/m3	5.7	0.95	1.61		09/23/21 17:46	75-09-2	
4-Methyl-2-pentanone (MIBK)	<0.52	ug/m3	6.7	0.52	1.61		09/23/21 17:46	108-10-1	
Methyl-tert-butyl ether	<0.20	ug/m3	5.9	0.20	1.61		09/23/21 17:46	1634-04-4	
Naphthalene	<3.5	ug/m3	4.3	3.5	1.61		09/23/21 17:46	91-20-3	
2-Propanol	144	ug/m3	4.0	0.82	1.61		09/23/21 17:46	67-63-0	
Propylene	<0.21	ug/m3	1.4	0.21	1.61		09/23/21 17:46	115-07-1	
Styrene	<0.62	ug/m3	3.5	0.62	1.61		09/23/21 17:46	100-42-5	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.

ANALYTICAL RESULTS

Project: CONBRACO
Pace Project No.: 92561801

Sample: IA-13 **Lab ID:** 92561801009 Collected: 09/14/21 17:04 Received: 09/16/21 12:00 Matrix: Air

Parameters	Results	Units	Report			Prepared	Analyzed	CAS No.	Qual
			Limit	MDL	DF				
TO15 MSV AIR									
Analytical Method: TO-15									
Pace Analytical Services - Minneapolis									
1,1,2,2-Tetrachloroethane	<0.60	ug/m3	2.3	0.60	1.61		09/23/21 17:46	79-34-5	
Tetrachloroethene	0.64J	ug/m3	1.1	0.47	1.61		09/23/21 17:46	127-18-4	
Tetrahydrofuran	<0.29	ug/m3	0.97	0.29	1.61		09/23/21 17:46	109-99-9	
Toluene	2.1	ug/m3	1.2	0.39	1.61		09/23/21 17:46	108-88-3	
1,2,4-Trichlorobenzene	<7.9	ug/m3	12.1	7.9	1.61		09/23/21 17:46	120-82-1	
1,1,1-Trichloroethane	<0.30	ug/m3	1.8	0.30	1.61		09/23/21 17:46	71-55-6	
1,1,2-Trichloroethane	<0.32	ug/m3	0.89	0.32	1.61		09/23/21 17:46	79-00-5	
Trichloroethene	52.1	ug/m3	0.88	0.32	1.61		09/23/21 17:46	79-01-6	
Trichlorofluoromethane	9.2	ug/m3	1.8	0.38	1.61		09/23/21 17:46	75-69-4	
1,1,2-Trichlorotrifluoroethane	0.70J	ug/m3	2.5	0.47	1.61		09/23/21 17:46	76-13-1	
1,2,4-Trimethylbenzene	<0.57	ug/m3	1.6	0.57	1.61		09/23/21 17:46	95-63-6	
1,3,5-Trimethylbenzene	<0.47	ug/m3	1.6	0.47	1.61		09/23/21 17:46	108-67-8	
Vinyl acetate	<0.33	ug/m3	1.2	0.33	1.61		09/23/21 17:46	108-05-4	
Vinyl chloride	<0.14	ug/m3	0.42	0.14	1.61		09/23/21 17:46	75-01-4	
m&p-Xylene	1.6J	ug/m3	2.8	1.0	1.61		09/23/21 17:46	179601-23-1	
o-Xylene	0.56J	ug/m3	1.4	0.44	1.61		09/23/21 17:46	95-47-6	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.

ANALYTICAL RESULTS

Project: CONBRACO

Pace Project No.: 92561801

Sample: IA-7 Lab ID: 92561801010 Collected: 09/14/21 16:56 Received: 09/16/21 12:00 Matrix: Air

Parameters	Results	Units	Report			Prepared	Analyzed	CAS No.	Qual
			Limit	MDL	DF				
TO15 MSV AIR									
Analytical Method: TO-15									
Pace Analytical Services - Minneapolis									
Acetone	38.7	ug/m3	9.5	2.9	1.58		09/23/21 19:16	67-64-1	
Benzene	0.29J	ug/m3	0.51	0.18	1.58		09/23/21 19:16	71-43-2	
Benzyl chloride	<1.4	ug/m3	4.2	1.4	1.58		09/23/21 19:16	100-44-7	
Bromodichloromethane	<0.37	ug/m3	2.1	0.37	1.58		09/23/21 19:16	75-27-4	
Bromoform	<2.6	ug/m3	8.3	2.6	1.58		09/23/21 19:16	75-25-2	
Bromomethane	<0.24	ug/m3	1.2	0.24	1.58		09/23/21 19:16	74-83-9	
1,3-Butadiene	<0.19	ug/m3	0.71	0.19	1.58		09/23/21 19:16	106-99-0	
2-Butanone (MEK)	4.8	ug/m3	4.7	0.73	1.58		09/23/21 19:16	78-93-3	
Carbon disulfide	<0.20	ug/m3	1.0	0.20	1.58		09/23/21 19:16	75-15-0	
Carbon tetrachloride	<0.44	ug/m3	2.0	0.44	1.58		09/23/21 19:16	56-23-5	
Chlorobenzene	<0.24	ug/m3	1.5	0.24	1.58		09/23/21 19:16	108-90-7	
Chloroethane	<0.35	ug/m3	0.85	0.35	1.58		09/23/21 19:16	75-00-3	
Chloroform	0.47J	ug/m3	0.78	0.29	1.58		09/23/21 19:16	67-66-3	
Chloromethane	0.72	ug/m3	0.66	0.13	1.58		09/23/21 19:16	74-87-3	
Cyclohexane	<0.35	ug/m3	2.8	0.35	1.58		09/23/21 19:16	110-82-7	
Dibromochloromethane	<0.81	ug/m3	2.7	0.81	1.58		09/23/21 19:16	124-48-1	
1,2-Dibromoethane (EDB)	<0.47	ug/m3	1.2	0.47	1.58		09/23/21 19:16	106-93-4	
1,2-Dichlorobenzene	<0.64	ug/m3	4.8	0.64	1.58		09/23/21 19:16	95-50-1	
1,3-Dichlorobenzene	<0.80	ug/m3	4.8	0.80	1.58		09/23/21 19:16	541-73-1	
1,4-Dichlorobenzene	<1.4	ug/m3	4.8	1.4	1.58		09/23/21 19:16	106-46-7	
Dichlorodifluoromethane	2.0	ug/m3	1.6	0.30	1.58		09/23/21 19:16	75-71-8	
1,1-Dichloroethane	<0.26	ug/m3	1.3	0.26	1.58		09/23/21 19:16	75-34-3	
1,2-Dichloroethane	0.50J	ug/m3	1.3	0.31	1.58		09/23/21 19:16	107-06-2	
1,1-Dichloroethene	<0.22	ug/m3	1.3	0.22	1.58		09/23/21 19:16	75-35-4	
cis-1,2-Dichloroethene	0.66J	ug/m3	1.3	0.31	1.58		09/23/21 19:16	156-59-2	
trans-1,2-Dichloroethene	<0.27	ug/m3	1.3	0.27	1.58		09/23/21 19:16	156-60-5	
1,2-Dichloropropane	<0.43	ug/m3	1.5	0.43	1.58		09/23/21 19:16	78-87-5	
cis-1,3-Dichloropropene	<0.40	ug/m3	3.6	0.40	1.58		09/23/21 19:16	10061-01-5	
trans-1,3-Dichloropropene	<0.86	ug/m3	3.6	0.86	1.58		09/23/21 19:16	10061-02-6	
Dichlorotetrafluoroethane	<0.32	ug/m3	2.2	0.32	1.58		09/23/21 19:16	76-14-2	
Ethanol	94.2	ug/m3	3.0	0.94	1.58		09/23/21 19:16	64-17-5	
Ethyl acetate	3.4	ug/m3	1.2	0.21	1.58		09/23/21 19:16	141-78-6	
Ethylbenzene	<0.49	ug/m3	1.4	0.49	1.58		09/23/21 19:16	100-41-4	
4-Ethyltoluene	<0.75	ug/m3	4.0	0.75	1.58		09/23/21 19:16	622-96-8	
n-Heptane	<0.29	ug/m3	1.3	0.29	1.58		09/23/21 19:16	142-82-5	
Hexachloro-1,3-butadiene	<1.9	ug/m3	8.6	1.9	1.58		09/23/21 19:16	87-68-3	
n-Hexane	0.83J	ug/m3	1.1	0.30	1.58		09/23/21 19:16	110-54-3	
2-Hexanone	<0.70	ug/m3	6.6	0.70	1.58		09/23/21 19:16	591-78-6	
Methylene Chloride	<0.94	ug/m3	5.6	0.94	1.58		09/23/21 19:16	75-09-2	
4-Methyl-2-pentanone (MIBK)	<0.51	ug/m3	6.6	0.51	1.58		09/23/21 19:16	108-10-1	
Methyl-tert-butyl ether	<0.20	ug/m3	5.8	0.20	1.58		09/23/21 19:16	1634-04-4	
Naphthalene	<3.4	ug/m3	4.2	3.4	1.58		09/23/21 19:16	91-20-3	
2-Propanol	9.8	ug/m3	4.0	0.80	1.58		09/23/21 19:16	67-63-0	
Propylene	<0.21	ug/m3	1.4	0.21	1.58		09/23/21 19:16	115-07-1	
Styrene	<0.61	ug/m3	3.4	0.61	1.58		09/23/21 19:16	100-42-5	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.

ANALYTICAL RESULTS

Project: CONBRACO
Pace Project No.: 92561801

Sample: IA-7 Lab ID: 92561801010 Collected: 09/14/21 16:56 Received: 09/16/21 12:00 Matrix: Air

Parameters	Results	Units	Report			Prepared	Analyzed	CAS No.	Qual
			Limit	MDL	DF				
TO15 MSV AIR									
Analytical Method: TO-15									
Pace Analytical Services - Minneapolis									
1,1,2,2-Tetrachloroethane	<0.59	ug/m3	2.2	0.59	1.58		09/23/21 19:16	79-34-5	
Tetrachloroethene	2.9	ug/m3	1.1	0.46	1.58		09/23/21 19:16	127-18-4	
Tetrahydrofuran	<0.28	ug/m3	0.95	0.28	1.58		09/23/21 19:16	109-99-9	
Toluene	1.3	ug/m3	1.2	0.39	1.58		09/23/21 19:16	108-88-3	
1,2,4-Trichlorobenzene	<7.7	ug/m3	11.9	7.7	1.58		09/23/21 19:16	120-82-1	
1,1,1-Trichloroethane	<0.29	ug/m3	1.8	0.29	1.58		09/23/21 19:16	71-55-6	
1,1,2-Trichloroethane	<0.31	ug/m3	0.88	0.31	1.58		09/23/21 19:16	79-00-5	
Trichloroethene	188	ug/m3	0.86	0.31	1.58		09/23/21 19:16	79-01-6	
Trichlorofluoromethane	2.1	ug/m3	1.8	0.37	1.58		09/23/21 19:16	75-69-4	
1,1,2-Trichlorotrifluoroethane	1.9J	ug/m3	2.5	0.46	1.58		09/23/21 19:16	76-13-1	
1,2,4-Trimethylbenzene	0.81J	ug/m3	1.6	0.56	1.58		09/23/21 19:16	95-63-6	
1,3,5-Trimethylbenzene	<0.46	ug/m3	1.6	0.46	1.58		09/23/21 19:16	108-67-8	
Vinyl acetate	<0.33	ug/m3	1.1	0.33	1.58		09/23/21 19:16	108-05-4	
Vinyl chloride	<0.14	ug/m3	0.41	0.14	1.58		09/23/21 19:16	75-01-4	
m&p-Xylene	<1.0	ug/m3	2.8	1.0	1.58		09/23/21 19:16	179601-23-1	
o-Xylene	<0.43	ug/m3	1.4	0.43	1.58		09/23/21 19:16	95-47-6	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.

ANALYTICAL RESULTS

Project: CONBRACO

Pace Project No.: 92561801

Sample: IA-3 Lab ID: 92561801011 Collected: 09/14/21 16:48 Received: 09/16/21 12:00 Matrix: Air

Parameters	Results	Units	Report			Prepared	Analyzed	CAS No.	Qual
			Limit	MDL	DF				
TO15 MSV AIR									
Analytical Method: TO-15									
Pace Analytical Services - Minneapolis									
Acetone	41.2	ug/m3	9.5	2.9	1.58		09/23/21 14:23	67-64-1	
Benzene	0.27J	ug/m3	0.51	0.18	1.58		09/23/21 14:23	71-43-2	
Benzyl chloride	<1.4	ug/m3	4.2	1.4	1.58		09/23/21 14:23	100-44-7	
Bromodichloromethane	<0.37	ug/m3	2.1	0.37	1.58		09/23/21 14:23	75-27-4	
Bromoform	<2.6	ug/m3	8.3	2.6	1.58		09/23/21 14:23	75-25-2	
Bromomethane	<0.24	ug/m3	1.2	0.24	1.58		09/23/21 14:23	74-83-9	
1,3-Butadiene	<0.19	ug/m3	0.71	0.19	1.58		09/23/21 14:23	106-99-0	
2-Butanone (MEK)	1.6J	ug/m3	4.7	0.73	1.58		09/23/21 14:23	78-93-3	
Carbon disulfide	<0.20	ug/m3	1.0	0.20	1.58		09/23/21 14:23	75-15-0	
Carbon tetrachloride	<0.44	ug/m3	2.0	0.44	1.58		09/23/21 14:23	56-23-5	
Chlorobenzene	<0.24	ug/m3	1.5	0.24	1.58		09/23/21 14:23	108-90-7	
Chloroethane	<0.35	ug/m3	0.85	0.35	1.58		09/23/21 14:23	75-00-3	
Chloroform	<0.29	ug/m3	0.78	0.29	1.58		09/23/21 14:23	67-66-3	
Chloromethane	1.4	ug/m3	0.66	0.13	1.58		09/23/21 14:23	74-87-3	
Cyclohexane	<0.35	ug/m3	2.8	0.35	1.58		09/23/21 14:23	110-82-7	
Dibromochloromethane	<0.81	ug/m3	2.7	0.81	1.58		09/23/21 14:23	124-48-1	
1,2-Dibromoethane (EDB)	<0.47	ug/m3	1.2	0.47	1.58		09/23/21 14:23	106-93-4	
1,2-Dichlorobenzene	<0.64	ug/m3	4.8	0.64	1.58		09/23/21 14:23	95-50-1	
1,3-Dichlorobenzene	<0.80	ug/m3	4.8	0.80	1.58		09/23/21 14:23	541-73-1	
1,4-Dichlorobenzene	<1.4	ug/m3	4.8	1.4	1.58		09/23/21 14:23	106-46-7	
Dichlorodifluoromethane	2.2	ug/m3	1.6	0.30	1.58		09/23/21 14:23	75-71-8	
1,1-Dichloroethane	<0.26	ug/m3	1.3	0.26	1.58		09/23/21 14:23	75-34-3	
1,2-Dichloroethane	<0.31	ug/m3	1.3	0.31	1.58		09/23/21 14:23	107-06-2	
1,1-Dichloroethene	<0.22	ug/m3	1.3	0.22	1.58		09/23/21 14:23	75-35-4	
cis-1,2-Dichloroethene	1.7	ug/m3	1.3	0.31	1.58		09/23/21 14:23	156-59-2	
trans-1,2-Dichloroethene	<0.27	ug/m3	1.3	0.27	1.58		09/23/21 14:23	156-60-5	
1,2-Dichloropropane	<0.43	ug/m3	1.5	0.43	1.58		09/23/21 14:23	78-87-5	
cis-1,3-Dichloropropene	<0.40	ug/m3	3.6	0.40	1.58		09/23/21 14:23	10061-01-5	
trans-1,3-Dichloropropene	<0.86	ug/m3	3.6	0.86	1.58		09/23/21 14:23	10061-02-6	
Dichlorotetrafluoroethane	<0.32	ug/m3	2.2	0.32	1.58		09/23/21 14:23	76-14-2	
Ethanol	20.1	ug/m3	3.0	0.94	1.58		09/23/21 14:23	64-17-5	
Ethyl acetate	1.6	ug/m3	1.2	0.21	1.58		09/23/21 14:23	141-78-6	
Ethylbenzene	<0.49	ug/m3	1.4	0.49	1.58		09/23/21 14:23	100-41-4	
4-Ethyltoluene	<0.75	ug/m3	4.0	0.75	1.58		09/23/21 14:23	622-96-8	
n-Heptane	<0.29	ug/m3	1.3	0.29	1.58		09/23/21 14:23	142-82-5	
Hexachloro-1,3-butadiene	<1.9	ug/m3	8.6	1.9	1.58		09/23/21 14:23	87-68-3	
n-Hexane	0.50J	ug/m3	1.1	0.30	1.58		09/23/21 14:23	110-54-3	
2-Hexanone	4.0J	ug/m3	6.6	0.70	1.58		09/23/21 14:23	591-78-6	
Methylene Chloride	<0.94	ug/m3	5.6	0.94	1.58		09/23/21 14:23	75-09-2	
4-Methyl-2-pentanone (MIBK)	1.1J	ug/m3	6.6	0.51	1.58		09/23/21 14:23	108-10-1	
Methyl-tert-butyl ether	<0.20	ug/m3	5.8	0.20	1.58		09/23/21 14:23	1634-04-4	
Naphthalene	<3.4	ug/m3	4.2	3.4	1.58		09/23/21 14:23	91-20-3	
2-Propanol	7.4	ug/m3	4.0	0.80	1.58		09/23/21 14:23	67-63-0	
Propylene	<0.21	ug/m3	1.4	0.21	1.58		09/23/21 14:23	115-07-1	
Styrene	<0.61	ug/m3	3.4	0.61	1.58		09/23/21 14:23	100-42-5	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.

ANALYTICAL RESULTS

Project: CONBRACO
Pace Project No.: 92561801

Sample: IA-3 Lab ID: 92561801011 Collected: 09/14/21 16:48 Received: 09/16/21 12:00 Matrix: Air

Parameters	Results	Units	Report			Prepared	Analyzed	CAS No.	Qual
			Limit	MDL	DF				
TO15 MSV AIR									
Analytical Method: TO-15									
Pace Analytical Services - Minneapolis									
1,1,2,2-Tetrachloroethane	<0.59	ug/m3	2.2	0.59	1.58		09/23/21 14:23	79-34-5	
Tetrachloroethene	2.2	ug/m3	1.1	0.46	1.58		09/23/21 14:23	127-18-4	
Tetrahydrofuran	<0.28	ug/m3	0.95	0.28	1.58		09/23/21 14:23	109-99-9	
Toluene	1.4	ug/m3	1.2	0.39	1.58		09/23/21 14:23	108-88-3	
1,2,4-Trichlorobenzene	<7.7	ug/m3	11.9	7.7	1.58		09/23/21 14:23	120-82-1	
1,1,1-Trichloroethane	<0.29	ug/m3	1.8	0.29	1.58		09/23/21 14:23	71-55-6	
1,1,2-Trichloroethane	<0.31	ug/m3	0.88	0.31	1.58		09/23/21 14:23	79-00-5	
Trichloroethene	148	ug/m3	0.86	0.31	1.58		09/23/21 14:23	79-01-6	
Trichlorofluoromethane	13.9	ug/m3	1.8	0.37	1.58		09/23/21 14:23	75-69-4	
1,1,2-Trichlorotrifluoroethane	1.8J	ug/m3	2.5	0.46	1.58		09/23/21 14:23	76-13-1	
1,2,4-Trimethylbenzene	<0.56	ug/m3	1.6	0.56	1.58		09/23/21 14:23	95-63-6	
1,3,5-Trimethylbenzene	<0.46	ug/m3	1.6	0.46	1.58		09/23/21 14:23	108-67-8	
Vinyl acetate	<0.33	ug/m3	1.1	0.33	1.58		09/23/21 14:23	108-05-4	
Vinyl chloride	<0.14	ug/m3	0.41	0.14	1.58		09/23/21 14:23	75-01-4	
m&p-Xylene	<1.0	ug/m3	2.8	1.0	1.58		09/23/21 14:23	179601-23-1	
o-Xylene	<0.43	ug/m3	1.4	0.43	1.58		09/23/21 14:23	95-47-6	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.

ANALYTICAL RESULTS

Project: CONBRACO

Pace Project No.: 92561801

Sample: IA-10 Lab ID: 92561801012 Collected: 09/14/21 17:25 Received: 09/16/21 12:00 Matrix: Air

Parameters	Results	Units	Report			Prepared	Analyzed	CAS No.	Qual
			Limit	MDL	DF				
TO15 MSV AIR									
Analytical Method: TO-15									
Pace Analytical Services - Minneapolis									
Acetone	371	ug/m3	8.7	2.6	1.44		09/23/21 22:38	67-64-1	
Benzene	0.35J	ug/m3	0.47	0.16	1.44		09/23/21 22:38	71-43-2	
Benzyl chloride	<1.3	ug/m3	3.8	1.3	1.44		09/23/21 22:38	100-44-7	
Bromodichloromethane	<0.34	ug/m3	2.0	0.34	1.44		09/23/21 22:38	75-27-4	
Bromoform	<2.3	ug/m3	7.6	2.3	1.44		09/23/21 22:38	75-25-2	
Bromomethane	<0.22	ug/m3	1.1	0.22	1.44		09/23/21 22:38	74-83-9	
1,3-Butadiene	<0.17	ug/m3	0.65	0.17	1.44		09/23/21 22:38	106-99-0	
2-Butanone (MEK)	2.6J	ug/m3	4.3	0.67	1.44		09/23/21 22:38	78-93-3	
Carbon disulfide	1.6	ug/m3	0.91	0.19	1.44		09/23/21 22:38	75-15-0	
Carbon tetrachloride	0.48J	ug/m3	1.8	0.40	1.44		09/23/21 22:38	56-23-5	
Chlorobenzene	<0.22	ug/m3	1.3	0.22	1.44		09/23/21 22:38	108-90-7	
Chloroethane	<0.32	ug/m3	0.77	0.32	1.44		09/23/21 22:38	75-00-3	
Chloroform	<0.26	ug/m3	0.71	0.26	1.44		09/23/21 22:38	67-66-3	
Chloromethane	0.96	ug/m3	0.60	0.12	1.44		09/23/21 22:38	74-87-3	
Cyclohexane	0.37J	ug/m3	2.5	0.32	1.44		09/23/21 22:38	110-82-7	
Dibromochloromethane	<0.74	ug/m3	2.5	0.74	1.44		09/23/21 22:38	124-48-1	
1,2-Dibromoethane (EDB)	<0.43	ug/m3	1.1	0.43	1.44		09/23/21 22:38	106-93-4	
1,2-Dichlorobenzene	<0.58	ug/m3	4.4	0.58	1.44		09/23/21 22:38	95-50-1	
1,3-Dichlorobenzene	<0.73	ug/m3	4.4	0.73	1.44		09/23/21 22:38	541-73-1	
1,4-Dichlorobenzene	<1.3	ug/m3	4.4	1.3	1.44		09/23/21 22:38	106-46-7	
Dichlorodifluoromethane	2.1	ug/m3	1.5	0.27	1.44		09/23/21 22:38	75-71-8	
1,1-Dichloroethane	<0.24	ug/m3	1.2	0.24	1.44		09/23/21 22:38	75-34-3	
1,2-Dichloroethane	<0.28	ug/m3	1.2	0.28	1.44		09/23/21 22:38	107-06-2	
1,1-Dichloroethene	<0.20	ug/m3	1.2	0.20	1.44		09/23/21 22:38	75-35-4	
cis-1,2-Dichloroethene	<0.28	ug/m3	1.2	0.28	1.44		09/23/21 22:38	156-59-2	
trans-1,2-Dichloroethene	<0.24	ug/m3	1.2	0.24	1.44		09/23/21 22:38	156-60-5	
1,2-Dichloropropane	<0.39	ug/m3	1.4	0.39	1.44		09/23/21 22:38	78-87-5	
cis-1,3-Dichloropropene	<0.37	ug/m3	3.3	0.37	1.44		09/23/21 22:38	10061-01-5	
trans-1,3-Dichloropropene	<0.78	ug/m3	3.3	0.78	1.44		09/23/21 22:38	10061-02-6	
Dichlorotetrafluoroethane	<0.29	ug/m3	2.0	0.29	1.44		09/23/21 22:38	76-14-2	
Ethanol	133	ug/m3	2.8	0.85	1.44		09/23/21 22:38	64-17-5	
Ethyl acetate	<0.19	ug/m3	1.1	0.19	1.44		09/23/21 22:38	141-78-6	
Ethylbenzene	<0.44	ug/m3	1.3	0.44	1.44		09/23/21 22:38	100-41-4	
4-Ethyltoluene	<0.68	ug/m3	3.6	0.68	1.44		09/23/21 22:38	622-96-8	
n-Heptane	0.78J	ug/m3	1.2	0.26	1.44		09/23/21 22:38	142-82-5	
Hexachloro-1,3-butadiene	<1.8	ug/m3	7.8	1.8	1.44		09/23/21 22:38	87-68-3	
n-Hexane	0.75J	ug/m3	1.0	0.28	1.44		09/23/21 22:38	110-54-3	
2-Hexanone	<0.64	ug/m3	6.0	0.64	1.44		09/23/21 22:38	591-78-6	
Methylene Chloride	<0.85	ug/m3	5.1	0.85	1.44		09/23/21 22:38	75-09-2	
4-Methyl-2-pentanone (MIBK)	<0.46	ug/m3	6.0	0.46	1.44		09/23/21 22:38	108-10-1	
Methyl-tert-butyl ether	<0.18	ug/m3	5.3	0.18	1.44		09/23/21 22:38	1634-04-4	
Naphthalene	<3.1	ug/m3	3.8	3.1	1.44		09/23/21 22:38	91-20-3	
2-Propanol	36.8	ug/m3	3.6	0.73	1.44		09/23/21 22:38	67-63-0	
Propylene	<0.19	ug/m3	1.3	0.19	1.44		09/23/21 22:38	115-07-1	
Styrene	<0.55	ug/m3	3.1	0.55	1.44		09/23/21 22:38	100-42-5	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.

ANALYTICAL RESULTS

Project: CONBRACO

Pace Project No.: 92561801

Sample: IA-10 **Lab ID: 92561801012** Collected: 09/14/21 17:25 Received: 09/16/21 12:00 Matrix: Air

Parameters	Results	Units	Report			Prepared	Analyzed	CAS No.	Qual
			Limit	MDL	DF				
TO15 MSV AIR									
Analytical Method: TO-15									
Pace Analytical Services - Minneapolis									
1,1,2,2-Tetrachloroethane	<0.54	ug/m3	2.0	0.54	1.44		09/23/21 22:38	79-34-5	
Tetrachloroethene	<0.42	ug/m3	0.99	0.42	1.44		09/23/21 22:38	127-18-4	
Tetrahydrofuran	<0.26	ug/m3	0.86	0.26	1.44		09/23/21 22:38	109-99-9	
Toluene	1.7	ug/m3	1.1	0.35	1.44		09/23/21 22:38	108-88-3	
1,2,4-Trichlorobenzene	<7.0	ug/m3	10.9	7.0	1.44		09/23/21 22:38	120-82-1	
1,1,1-Trichloroethane	<0.27	ug/m3	1.6	0.27	1.44		09/23/21 22:38	71-55-6	
1,1,2-Trichloroethane	<0.28	ug/m3	0.80	0.28	1.44		09/23/21 22:38	79-00-5	
Trichloroethene	9.2	ug/m3	0.79	0.28	1.44		09/23/21 22:38	79-01-6	
Trichlorofluoromethane	3.8	ug/m3	1.6	0.34	1.44		09/23/21 22:38	75-69-4	
1,1,2-Trichlorotrifluoroethane	0.72J	ug/m3	2.2	0.42	1.44		09/23/21 22:38	76-13-1	
1,2,4-Trimethylbenzene	<0.51	ug/m3	1.4	0.51	1.44		09/23/21 22:38	95-63-6	
1,3,5-Trimethylbenzene	<0.42	ug/m3	1.4	0.42	1.44		09/23/21 22:38	108-67-8	
Vinyl acetate	<0.30	ug/m3	1.0	0.30	1.44		09/23/21 22:38	108-05-4	
Vinyl chloride	<0.12	ug/m3	0.37	0.12	1.44		09/23/21 22:38	75-01-4	
m&p-Xylene	0.97J	ug/m3	2.5	0.92	1.44		09/23/21 22:38	179601-23-1	
o-Xylene	0.41J	ug/m3	1.3	0.39	1.44		09/23/21 22:38	95-47-6	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.

ANALYTICAL RESULTS

Project: CONBRACO

Pace Project No.: 92561801

Sample: IA-11 Lab ID: 92561801013 Collected: 09/14/21 18:45 Received: 09/16/21 12:00 Matrix: Air

Parameters	Results	Units	Report			Prepared	Analyzed	CAS No.	Qual
			Limit	MDL	DF				
TO15 MSV AIR									
Analytical Method: TO-15									
Pace Analytical Services - Minneapolis									
Acetone	409	ug/m3	9.7	2.9	1.61		09/23/21 20:16	67-64-1	
Benzene	0.36J	ug/m3	0.52	0.18	1.61		09/23/21 20:16	71-43-2	
Benzyl chloride	<1.4	ug/m3	4.2	1.4	1.61		09/23/21 20:16	100-44-7	
Bromodichloromethane	<0.38	ug/m3	2.2	0.38	1.61		09/23/21 20:16	75-27-4	
Bromoform	<2.6	ug/m3	8.5	2.6	1.61		09/23/21 20:16	75-25-2	
Bromomethane	<0.24	ug/m3	1.3	0.24	1.61		09/23/21 20:16	74-83-9	
1,3-Butadiene	<0.19	ug/m3	0.72	0.19	1.61		09/23/21 20:16	106-99-0	
2-Butanone (MEK)	1.5J	ug/m3	4.8	0.75	1.61		09/23/21 20:16	78-93-3	
Carbon disulfide	<0.21	ug/m3	1.0	0.21	1.61		09/23/21 20:16	75-15-0	
Carbon tetrachloride	<0.45	ug/m3	2.1	0.45	1.61		09/23/21 20:16	56-23-5	
Chlorobenzene	<0.25	ug/m3	1.5	0.25	1.61		09/23/21 20:16	108-90-7	
Chloroethane	<0.36	ug/m3	0.86	0.36	1.61		09/23/21 20:16	75-00-3	
Chloroform	0.30J	ug/m3	0.80	0.29	1.61		09/23/21 20:16	67-66-3	
Chloromethane	0.89	ug/m3	0.68	0.14	1.61		09/23/21 20:16	74-87-3	
Cyclohexane	0.41J	ug/m3	2.8	0.36	1.61		09/23/21 20:16	110-82-7	
Dibromochloromethane	<0.83	ug/m3	2.8	0.83	1.61		09/23/21 20:16	124-48-1	
1,2-Dibromoethane (EDB)	<0.48	ug/m3	1.3	0.48	1.61		09/23/21 20:16	106-93-4	
1,2-Dichlorobenzene	<0.65	ug/m3	4.9	0.65	1.61		09/23/21 20:16	95-50-1	
1,3-Dichlorobenzene	<0.82	ug/m3	4.9	0.82	1.61		09/23/21 20:16	541-73-1	
1,4-Dichlorobenzene	<1.4	ug/m3	4.9	1.4	1.61		09/23/21 20:16	106-46-7	
Dichlorodifluoromethane	2.4	ug/m3	1.6	0.30	1.61		09/23/21 20:16	75-71-8	
1,1-Dichloroethane	<0.27	ug/m3	1.3	0.27	1.61		09/23/21 20:16	75-34-3	
1,2-Dichloroethane	<0.31	ug/m3	1.3	0.31	1.61		09/23/21 20:16	107-06-2	
1,1-Dichloroethene	<0.22	ug/m3	1.3	0.22	1.61		09/23/21 20:16	75-35-4	
cis-1,2-Dichloroethene	0.32J	ug/m3	1.3	0.31	1.61		09/23/21 20:16	156-59-2	
trans-1,2-Dichloroethene	<0.27	ug/m3	1.3	0.27	1.61		09/23/21 20:16	156-60-5	
1,2-Dichloropropane	<0.43	ug/m3	1.5	0.43	1.61		09/23/21 20:16	78-87-5	
cis-1,3-Dichloropropene	<0.41	ug/m3	3.7	0.41	1.61		09/23/21 20:16	10061-01-5	
trans-1,3-Dichloropropene	<0.88	ug/m3	3.7	0.88	1.61		09/23/21 20:16	10061-02-6	
Dichlorotetrafluoroethane	<0.33	ug/m3	2.3	0.33	1.61		09/23/21 20:16	76-14-2	
Ethanol	163	ug/m3	3.1	0.95	1.61		09/23/21 20:16	64-17-5	
Ethyl acetate	0.71J	ug/m3	1.2	0.21	1.61		09/23/21 20:16	141-78-6	
Ethylbenzene	<0.50	ug/m3	1.4	0.50	1.61		09/23/21 20:16	100-41-4	
4-Ethyltoluene	<0.76	ug/m3	4.0	0.76	1.61		09/23/21 20:16	622-96-8	
n-Heptane	<0.29	ug/m3	1.3	0.29	1.61		09/23/21 20:16	142-82-5	
Hexachloro-1,3-butadiene	<2.0	ug/m3	8.7	2.0	1.61		09/23/21 20:16	87-68-3	
n-Hexane	0.67J	ug/m3	1.2	0.31	1.61		09/23/21 20:16	110-54-3	
2-Hexanone	<0.71	ug/m3	6.7	0.71	1.61		09/23/21 20:16	591-78-6	
Methylene Chloride	1.6J	ug/m3	5.7	0.95	1.61		09/23/21 20:16	75-09-2	
4-Methyl-2-pentanone (MIBK)	<0.52	ug/m3	6.7	0.52	1.61		09/23/21 20:16	108-10-1	
Methyl-tert-butyl ether	<0.20	ug/m3	5.9	0.20	1.61		09/23/21 20:16	1634-04-4	
Naphthalene	<3.5	ug/m3	4.3	3.5	1.61		09/23/21 20:16	91-20-3	
2-Propanol	439	ug/m3	4.0	0.82	1.61		09/23/21 20:16	67-63-0	
Propylene	<0.21	ug/m3	1.4	0.21	1.61		09/23/21 20:16	115-07-1	
Styrene	<0.62	ug/m3	3.5	0.62	1.61		09/23/21 20:16	100-42-5	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.

ANALYTICAL RESULTS

Project: CONBRACO
Pace Project No.: 92561801

Sample: IA-11 Lab ID: 92561801013 Collected: 09/14/21 18:45 Received: 09/16/21 12:00 Matrix: Air

Parameters	Results	Units	Report			Prepared	Analyzed	CAS No.	Qual
			Limit	MDL	DF				
TO15 MSV AIR									
Analytical Method: TO-15									
Pace Analytical Services - Minneapolis									
1,1,2,2-Tetrachloroethane	<0.60	ug/m3	2.3	0.60	1.61		09/23/21 20:16	79-34-5	
Tetrachloroethene	1.6	ug/m3	1.1	0.47	1.61		09/23/21 20:16	127-18-4	
Tetrahydrofuran	<0.29	ug/m3	0.97	0.29	1.61		09/23/21 20:16	109-99-9	
Toluene	1.7	ug/m3	1.2	0.39	1.61		09/23/21 20:16	108-88-3	
1,2,4-Trichlorobenzene	<7.9	ug/m3	12.1	7.9	1.61		09/23/21 20:16	120-82-1	
1,1,1-Trichloroethane	<0.30	ug/m3	1.8	0.30	1.61		09/23/21 20:16	71-55-6	
1,1,2-Trichloroethane	<0.32	ug/m3	0.89	0.32	1.61		09/23/21 20:16	79-00-5	
Trichloroethene	154	ug/m3	0.88	0.32	1.61		09/23/21 20:16	79-01-6	
Trichlorofluoromethane	16.3	ug/m3	1.8	0.38	1.61		09/23/21 20:16	75-69-4	
1,1,2-Trichlorotrifluoroethane	1.2J	ug/m3	2.5	0.47	1.61		09/23/21 20:16	76-13-1	
1,2,4-Trimethylbenzene	<0.57	ug/m3	1.6	0.57	1.61		09/23/21 20:16	95-63-6	
1,3,5-Trimethylbenzene	<0.47	ug/m3	1.6	0.47	1.61		09/23/21 20:16	108-67-8	
Vinyl acetate	<0.33	ug/m3	1.2	0.33	1.61		09/23/21 20:16	108-05-4	
Vinyl chloride	<0.14	ug/m3	0.42	0.14	1.61		09/23/21 20:16	75-01-4	
m&p-Xylene	1.2J	ug/m3	2.8	1.0	1.61		09/23/21 20:16	179601-23-1	
o-Xylene	0.47J	ug/m3	1.4	0.44	1.61		09/23/21 20:16	95-47-6	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.

ANALYTICAL RESULTS

Project: CONBRACO

Pace Project No.: 92561801

Sample: IA-4 Lab ID: 92561801014 Collected: 09/14/21 16:16 Received: 09/16/21 12:00 Matrix: Air

Parameters	Results	Units	Report			Prepared	Analyzed	CAS No.	Qual
			Limit	MDL	DF				
TO15 MSV AIR									
Analytical Method: TO-15									
Pace Analytical Services - Minneapolis									
Acetone	25.1	ug/m3	9.7	2.9	1.61		09/23/21 23:07	67-64-1	
Benzene	0.28J	ug/m3	0.52	0.18	1.61		09/23/21 23:07	71-43-2	
Benzyl chloride	<1.4	ug/m3	4.2	1.4	1.61		09/23/21 23:07	100-44-7	
Bromodichloromethane	<0.38	ug/m3	2.2	0.38	1.61		09/23/21 23:07	75-27-4	
Bromoform	<2.6	ug/m3	8.5	2.6	1.61		09/23/21 23:07	75-25-2	
Bromomethane	<0.24	ug/m3	1.3	0.24	1.61		09/23/21 23:07	74-83-9	
1,3-Butadiene	<0.19	ug/m3	0.72	0.19	1.61		09/23/21 23:07	106-99-0	
2-Butanone (MEK)	6.6	ug/m3	4.8	0.75	1.61		09/23/21 23:07	78-93-3	
Carbon disulfide	<0.21	ug/m3	1.0	0.21	1.61		09/23/21 23:07	75-15-0	
Carbon tetrachloride	<0.45	ug/m3	2.1	0.45	1.61		09/23/21 23:07	56-23-5	
Chlorobenzene	<0.25	ug/m3	1.5	0.25	1.61		09/23/21 23:07	108-90-7	
Chloroethane	<0.36	ug/m3	0.86	0.36	1.61		09/23/21 23:07	75-00-3	
Chloroform	0.65J	ug/m3	0.80	0.29	1.61		09/23/21 23:07	67-66-3	
Chloromethane	4.5	ug/m3	0.68	0.14	1.61		09/23/21 23:07	74-87-3	
Cyclohexane	<0.36	ug/m3	2.8	0.36	1.61		09/23/21 23:07	110-82-7	
Dibromochloromethane	<0.83	ug/m3	2.8	0.83	1.61		09/23/21 23:07	124-48-1	
1,2-Dibromoethane (EDB)	<0.48	ug/m3	1.3	0.48	1.61		09/23/21 23:07	106-93-4	
1,2-Dichlorobenzene	<0.65	ug/m3	4.9	0.65	1.61		09/23/21 23:07	95-50-1	
1,3-Dichlorobenzene	<0.82	ug/m3	4.9	0.82	1.61		09/23/21 23:07	541-73-1	
1,4-Dichlorobenzene	<1.4	ug/m3	4.9	1.4	1.61		09/23/21 23:07	106-46-7	
Dichlorodifluoromethane	2.0	ug/m3	1.6	0.30	1.61		09/23/21 23:07	75-71-8	
1,1-Dichloroethane	<0.27	ug/m3	1.3	0.27	1.61		09/23/21 23:07	75-34-3	
1,2-Dichloroethane	<0.31	ug/m3	1.3	0.31	1.61		09/23/21 23:07	107-06-2	
1,1-Dichloroethene	<0.22	ug/m3	1.3	0.22	1.61		09/23/21 23:07	75-35-4	
cis-1,2-Dichloroethene	3.2	ug/m3	1.3	0.31	1.61		09/23/21 23:07	156-59-2	
trans-1,2-Dichloroethene	<0.27	ug/m3	1.3	0.27	1.61		09/23/21 23:07	156-60-5	
1,2-Dichloropropane	<0.43	ug/m3	1.5	0.43	1.61		09/23/21 23:07	78-87-5	
cis-1,3-Dichloropropene	<0.41	ug/m3	3.7	0.41	1.61		09/23/21 23:07	10061-01-5	
trans-1,3-Dichloropropene	<0.88	ug/m3	3.7	0.88	1.61		09/23/21 23:07	10061-02-6	
Dichlorotetrafluoroethane	<0.33	ug/m3	2.3	0.33	1.61		09/23/21 23:07	76-14-2	
Ethanol	152	ug/m3	3.1	0.95	1.61		09/23/21 23:07	64-17-5	
Ethyl acetate	3.9	ug/m3	1.2	0.21	1.61		09/23/21 23:07	141-78-6	
Ethylbenzene	<0.50	ug/m3	1.4	0.50	1.61		09/23/21 23:07	100-41-4	
4-Ethyltoluene	<0.76	ug/m3	4.0	0.76	1.61		09/23/21 23:07	622-96-8	
n-Heptane	<0.29	ug/m3	1.3	0.29	1.61		09/23/21 23:07	142-82-5	
Hexachloro-1,3-butadiene	<2.0	ug/m3	8.7	2.0	1.61		09/23/21 23:07	87-68-3	
n-Hexane	0.63J	ug/m3	1.2	0.31	1.61		09/23/21 23:07	110-54-3	
2-Hexanone	<0.71	ug/m3	6.7	0.71	1.61		09/23/21 23:07	591-78-6	
Methylene Chloride	<0.95	ug/m3	5.7	0.95	1.61		09/23/21 23:07	75-09-2	
4-Methyl-2-pentanone (MIBK)	<0.52	ug/m3	6.7	0.52	1.61		09/23/21 23:07	108-10-1	
Methyl-tert-butyl ether	<0.20	ug/m3	5.9	0.20	1.61		09/23/21 23:07	1634-04-4	
Naphthalene	<3.5	ug/m3	4.3	3.5	1.61		09/23/21 23:07	91-20-3	
2-Propanol	3.5J	ug/m3	4.0	0.82	1.61		09/23/21 23:07	67-63-0	
Propylene	<0.21	ug/m3	1.4	0.21	1.61		09/23/21 23:07	115-07-1	
Styrene	<0.62	ug/m3	3.5	0.62	1.61		09/23/21 23:07	100-42-5	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.

ANALYTICAL RESULTS

Project: CONBRACO

Pace Project No.: 92561801

Sample: IA-4 Lab ID: 92561801014 Collected: 09/14/21 16:16 Received: 09/16/21 12:00 Matrix: Air

Parameters	Results	Units	Report			Prepared	Analyzed	CAS No.	Qual
			Limit	MDL	DF				
TO15 MSV AIR									
Analytical Method: TO-15									
Pace Analytical Services - Minneapolis									
1,1,2,2-Tetrachloroethane	<0.60	ug/m3	2.3	0.60	1.61		09/23/21 23:07	79-34-5	
Tetrachloroethene	1.3	ug/m3	1.1	0.47	1.61		09/23/21 23:07	127-18-4	
Tetrahydrofuran	<0.29	ug/m3	0.97	0.29	1.61		09/23/21 23:07	109-99-9	
Toluene	1.9	ug/m3	1.2	0.39	1.61		09/23/21 23:07	108-88-3	
1,2,4-Trichlorobenzene	<7.9	ug/m3	12.1	7.9	1.61		09/23/21 23:07	120-82-1	
1,1,1-Trichloroethane	<0.30	ug/m3	1.8	0.30	1.61		09/23/21 23:07	71-55-6	
1,1,2-Trichloroethane	<0.32	ug/m3	0.89	0.32	1.61		09/23/21 23:07	79-00-5	
Trichloroethene	107	ug/m3	0.88	0.32	1.61		09/23/21 23:07	79-01-6	
Trichlorofluoromethane	8.1	ug/m3	1.8	0.38	1.61		09/23/21 23:07	75-69-4	
1,1,2-Trichlorotrifluoroethane	0.80J	ug/m3	2.5	0.47	1.61		09/23/21 23:07	76-13-1	
1,2,4-Trimethylbenzene	<0.57	ug/m3	1.6	0.57	1.61		09/23/21 23:07	95-63-6	
1,3,5-Trimethylbenzene	<0.47	ug/m3	1.6	0.47	1.61		09/23/21 23:07	108-67-8	
Vinyl acetate	<0.33	ug/m3	1.2	0.33	1.61		09/23/21 23:07	108-05-4	
Vinyl chloride	<0.14	ug/m3	0.42	0.14	1.61		09/23/21 23:07	75-01-4	
m&p-Xylene	<1.0	ug/m3	2.8	1.0	1.61		09/23/21 23:07	179601-23-1	
o-Xylene	<0.44	ug/m3	1.4	0.44	1.61		09/23/21 23:07	95-47-6	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.

ANALYTICAL RESULTS

Project: CONBRACO
Pace Project No.: 92561801

Sample: IA-9 **Lab ID: 92561801015** Collected: 09/14/21 18:04 Received: 09/16/21 12:00 Matrix: Air

Parameters	Results	Units	Report			Prepared	Analyzed	CAS No.	Qual
			Limit	MDL	DF				
TO15 MSV AIR									
Analytical Method: TO-15									
Pace Analytical Services - Minneapolis									
Acetone	285	ug/m3	18.7	5.6	3.1		09/23/21 21:41	67-64-1	
Benzene	<0.35	ug/m3	1.0	0.35	3.1		09/23/21 21:41	71-43-2	
Benzyl chloride	<2.8	ug/m3	8.2	2.8	3.1		09/23/21 21:41	100-44-7	
Bromodichloromethane	<0.73	ug/m3	4.2	0.73	3.1		09/23/21 21:41	75-27-4	
Bromoform	<5.0	ug/m3	16.3	5.0	3.1		09/23/21 21:41	75-25-2	
Bromomethane	<0.46	ug/m3	2.4	0.46	3.1		09/23/21 21:41	74-83-9	
1,3-Butadiene	<0.37	ug/m3	1.4	0.37	3.1		09/23/21 21:41	106-99-0	
2-Butanone (MEK)	3.1J	ug/m3	9.3	1.4	3.1		09/23/21 21:41	78-93-3	
Carbon disulfide	<0.40	ug/m3	2.0	0.40	3.1		09/23/21 21:41	75-15-0	
Carbon tetrachloride	<0.87	ug/m3	4.0	0.87	3.1		09/23/21 21:41	56-23-5	
Chlorobenzene	<0.48	ug/m3	2.9	0.48	3.1		09/23/21 21:41	108-90-7	
Chloroethane	<0.69	ug/m3	1.7	0.69	3.1		09/23/21 21:41	75-00-3	
Chloroform	<0.57	ug/m3	1.5	0.57	3.1		09/23/21 21:41	67-66-3	
Chloromethane	0.77J	ug/m3	1.3	0.26	3.1		09/23/21 21:41	74-87-3	
Cyclohexane	1.9J	ug/m3	5.4	0.69	3.1		09/23/21 21:41	110-82-7	
Dibromochloromethane	<1.6	ug/m3	5.4	1.6	3.1		09/23/21 21:41	124-48-1	
1,2-Dibromoethane (EDB)	<0.93	ug/m3	2.4	0.93	3.1		09/23/21 21:41	106-93-4	
1,2-Dichlorobenzene	<1.3	ug/m3	9.5	1.3	3.1		09/23/21 21:41	95-50-1	
1,3-Dichlorobenzene	<1.6	ug/m3	9.5	1.6	3.1		09/23/21 21:41	541-73-1	
1,4-Dichlorobenzene	<2.7	ug/m3	9.5	2.7	3.1		09/23/21 21:41	106-46-7	
Dichlorodifluoromethane	2.0J	ug/m3	3.1	0.58	3.1		09/23/21 21:41	75-71-8	
1,1-Dichloroethane	<0.51	ug/m3	2.6	0.51	3.1		09/23/21 21:41	75-34-3	
1,2-Dichloroethane	<0.60	ug/m3	2.6	0.60	3.1		09/23/21 21:41	107-06-2	
1,1-Dichloroethene	<0.43	ug/m3	2.5	0.43	3.1		09/23/21 21:41	75-35-4	
cis-1,2-Dichloroethene	<0.60	ug/m3	2.5	0.60	3.1		09/23/21 21:41	156-59-2	
trans-1,2-Dichloroethene	<0.52	ug/m3	2.5	0.52	3.1		09/23/21 21:41	156-60-5	
1,2-Dichloropropane	<0.83	ug/m3	2.9	0.83	3.1		09/23/21 21:41	78-87-5	
cis-1,3-Dichloropropene	<0.79	ug/m3	7.2	0.79	3.1		09/23/21 21:41	10061-01-5	
trans-1,3-Dichloropropene	<1.7	ug/m3	7.2	1.7	3.1		09/23/21 21:41	10061-02-6	
Dichlorotetrafluoroethane	<0.63	ug/m3	4.4	0.63	3.1		09/23/21 21:41	76-14-2	
Ethanol	195	ug/m3	6.0	1.8	3.1		09/23/21 21:41	64-17-5	
Ethyl acetate	29.8	ug/m3	2.3	0.41	3.1		09/23/21 21:41	141-78-6	
Ethylbenzene	<0.96	ug/m3	2.7	0.96	3.1		09/23/21 21:41	100-41-4	
4-Ethyltoluene	<1.5	ug/m3	7.8	1.5	3.1		09/23/21 21:41	622-96-8	
n-Heptane	<0.56	ug/m3	2.6	0.56	3.1		09/23/21 21:41	142-82-5	
Hexachloro-1,3-butadiene	<3.8	ug/m3	16.8	3.8	3.1		09/23/21 21:41	87-68-3	
n-Hexane	7.6	ug/m3	2.2	0.59	3.1		09/23/21 21:41	110-54-3	
2-Hexanone	<1.4	ug/m3	12.9	1.4	3.1		09/23/21 21:41	591-78-6	
Methylene Chloride	3.9J	ug/m3	10.9	1.8	3.1		09/23/21 21:41	75-09-2	
4-Methyl-2-pentanone (MIBK)	<1.0	ug/m3	12.9	1.0	3.1		09/23/21 21:41	108-10-1	
Methyl-tert-butyl ether	<0.39	ug/m3	11.3	0.39	3.1		09/23/21 21:41	1634-04-4	
Naphthalene	<6.7	ug/m3	8.2	6.7	3.1		09/23/21 21:41	91-20-3	
2-Propanol	532	ug/m3	7.8	1.6	3.1		09/23/21 21:41	67-63-0	
Propylene	<0.40	ug/m3	2.7	0.40	3.1		09/23/21 21:41	115-07-1	
Styrene	<1.2	ug/m3	6.7	1.2	3.1		09/23/21 21:41	100-42-5	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.

ANALYTICAL RESULTS

Project: CONBRACO
Pace Project No.: 92561801

Sample: IA-9 Lab ID: 92561801015 Collected: 09/14/21 18:04 Received: 09/16/21 12:00 Matrix: Air

Parameters	Results	Units	Report			Prepared	Analyzed	CAS No.	Qual
			Limit	MDL	DF				
TO15 MSV AIR									
Analytical Method: TO-15									
Pace Analytical Services - Minneapolis									
1,1,2,2-Tetrachloroethane	<1.2	ug/m3	4.3	1.2	3.1		09/23/21 21:41	79-34-5	
Tetrachloroethene	0.95J	ug/m3	2.1	0.91	3.1		09/23/21 21:41	127-18-4	
Tetrahydrofuran	<0.56	ug/m3	1.9	0.56	3.1		09/23/21 21:41	109-99-9	
Toluene	11.0	ug/m3	2.4	0.76	3.1		09/23/21 21:41	108-88-3	
1,2,4-Trichlorobenzene	<15.1	ug/m3	23.4	15.1	3.1		09/23/21 21:41	120-82-1	
1,1,1-Trichloroethane	<0.58	ug/m3	3.4	0.58	3.1		09/23/21 21:41	71-55-6	
1,1,2-Trichloroethane	<0.61	ug/m3	1.7	0.61	3.1		09/23/21 21:41	79-00-5	
Trichloroethene	73.7	ug/m3	1.7	0.61	3.1		09/23/21 21:41	79-01-6	
Trichlorofluoromethane	13.8	ug/m3	3.5	0.72	3.1		09/23/21 21:41	75-69-4	
1,1,2-Trichlorotrifluoroethane	<0.90	ug/m3	4.8	0.90	3.1		09/23/21 21:41	76-13-1	
1,2,4-Trimethylbenzene	2.2J	ug/m3	3.1	1.1	3.1		09/23/21 21:41	95-63-6	
1,3,5-Trimethylbenzene	<0.90	ug/m3	3.1	0.90	3.1		09/23/21 21:41	108-67-8	
Vinyl acetate	<0.64	ug/m3	2.2	0.64	3.1		09/23/21 21:41	108-05-4	
Vinyl chloride	<0.27	ug/m3	0.81	0.27	3.1		09/23/21 21:41	75-01-4	
m&p-Xylene	2.2J	ug/m3	5.5	2.0	3.1		09/23/21 21:41	179601-23-1	
o-Xylene	0.94J	ug/m3	2.7	0.84	3.1		09/23/21 21:41	95-47-6	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.

ANALYTICAL RESULTS

Project: CONBRACO

Pace Project No.: 92561801

Sample: **BACKGROUND** Lab ID: **92561801016** Collected: 09/14/21 18:08 Received: 09/16/21 12:00 Matrix: Air

Parameters	Results	Units	Report			Prepared	Analyzed	CAS No.	Qual
			Limit	MDL	DF				
TO15 MSV AIR									
Analytical Method: TO-15									
Pace Analytical Services - Minneapolis									
Acetone	8.5J	ug/m3	9.7	2.9	1.61		09/23/21 19:45	67-64-1	
Benzene	0.53	ug/m3	0.52	0.18	1.61		09/23/21 19:45	71-43-2	
Benzyl chloride	<1.4	ug/m3	4.2	1.4	1.61		09/23/21 19:45	100-44-7	
Bromodichloromethane	<0.38	ug/m3	2.2	0.38	1.61		09/23/21 19:45	75-27-4	
Bromoform	<2.6	ug/m3	8.5	2.6	1.61		09/23/21 19:45	75-25-2	
Bromomethane	<0.24	ug/m3	1.3	0.24	1.61		09/23/21 19:45	74-83-9	
1,3-Butadiene	<0.19	ug/m3	0.72	0.19	1.61		09/23/21 19:45	106-99-0	
2-Butanone (MEK)	<0.75	ug/m3	4.8	0.75	1.61		09/23/21 19:45	78-93-3	
Carbon disulfide	<0.21	ug/m3	1.0	0.21	1.61		09/23/21 19:45	75-15-0	
Carbon tetrachloride	<0.45	ug/m3	2.1	0.45	1.61		09/23/21 19:45	56-23-5	
Chlorobenzene	<0.25	ug/m3	1.5	0.25	1.61		09/23/21 19:45	108-90-7	
Chloroethane	<0.36	ug/m3	0.86	0.36	1.61		09/23/21 19:45	75-00-3	
Chloroform	<0.29	ug/m3	0.80	0.29	1.61		09/23/21 19:45	67-66-3	
Chloromethane	0.73	ug/m3	0.68	0.14	1.61		09/23/21 19:45	74-87-3	
Cyclohexane	<0.36	ug/m3	2.8	0.36	1.61		09/23/21 19:45	110-82-7	
Dibromochloromethane	<0.83	ug/m3	2.8	0.83	1.61		09/23/21 19:45	124-48-1	
1,2-Dibromoethane (EDB)	<0.48	ug/m3	1.3	0.48	1.61		09/23/21 19:45	106-93-4	
1,2-Dichlorobenzene	<0.65	ug/m3	4.9	0.65	1.61		09/23/21 19:45	95-50-1	
1,3-Dichlorobenzene	<0.82	ug/m3	4.9	0.82	1.61		09/23/21 19:45	541-73-1	
1,4-Dichlorobenzene	<1.4	ug/m3	4.9	1.4	1.61		09/23/21 19:45	106-46-7	
Dichlorodifluoromethane	2.1	ug/m3	1.6	0.30	1.61		09/23/21 19:45	75-71-8	
1,1-Dichloroethane	<0.27	ug/m3	1.3	0.27	1.61		09/23/21 19:45	75-34-3	
1,2-Dichloroethane	<0.31	ug/m3	1.3	0.31	1.61		09/23/21 19:45	107-06-2	
1,1-Dichloroethene	<0.22	ug/m3	1.3	0.22	1.61		09/23/21 19:45	75-35-4	
cis-1,2-Dichloroethene	<0.31	ug/m3	1.3	0.31	1.61		09/23/21 19:45	156-59-2	
trans-1,2-Dichloroethene	<0.27	ug/m3	1.3	0.27	1.61		09/23/21 19:45	156-60-5	
1,2-Dichloropropane	<0.43	ug/m3	1.5	0.43	1.61		09/23/21 19:45	78-87-5	
cis-1,3-Dichloropropene	<0.41	ug/m3	3.7	0.41	1.61		09/23/21 19:45	10061-01-5	
trans-1,3-Dichloropropene	<0.88	ug/m3	3.7	0.88	1.61		09/23/21 19:45	10061-02-6	
Dichlorotetrafluoroethane	<0.33	ug/m3	2.3	0.33	1.61		09/23/21 19:45	76-14-2	
Ethanol	4.2	ug/m3	3.1	0.95	1.61		09/23/21 19:45	64-17-5	
Ethyl acetate	<0.21	ug/m3	1.2	0.21	1.61		09/23/21 19:45	141-78-6	
Ethylbenzene	<0.50	ug/m3	1.4	0.50	1.61		09/23/21 19:45	100-41-4	
4-Ethyltoluene	<0.76	ug/m3	4.0	0.76	1.61		09/23/21 19:45	622-96-8	
n-Heptane	<0.29	ug/m3	1.3	0.29	1.61		09/23/21 19:45	142-82-5	
Hexachloro-1,3-butadiene	<2.0	ug/m3	8.7	2.0	1.61		09/23/21 19:45	87-68-3	
n-Hexane	<0.31	ug/m3	1.2	0.31	1.61		09/23/21 19:45	110-54-3	
2-Hexanone	<0.71	ug/m3	6.7	0.71	1.61		09/23/21 19:45	591-78-6	
Methylene Chloride	<0.95	ug/m3	5.7	0.95	1.61		09/23/21 19:45	75-09-2	
4-Methyl-2-pentanone (MIBK)	<0.52	ug/m3	6.7	0.52	1.61		09/23/21 19:45	108-10-1	
Methyl-tert-butyl ether	<0.20	ug/m3	5.9	0.20	1.61		09/23/21 19:45	1634-04-4	
Naphthalene	<3.5	ug/m3	4.3	3.5	1.61		09/23/21 19:45	91-20-3	
2-Propanol	1.1J	ug/m3	4.0	0.82	1.61		09/23/21 19:45	67-63-0	
Propylene	<0.21	ug/m3	1.4	0.21	1.61		09/23/21 19:45	115-07-1	
Styrene	<0.62	ug/m3	3.5	0.62	1.61		09/23/21 19:45	100-42-5	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.

ANALYTICAL RESULTS

Project: CONBRACO
Pace Project No.: 92561801

Sample: BACKGROUND **Lab ID: 92561801016** Collected: 09/14/21 18:08 Received: 09/16/21 12:00 Matrix: Air

Parameters	Results	Units	Report			Prepared	Analyzed	CAS No.	Qual
			Limit	MDL	DF				
TO15 MSV AIR									
Analytical Method: TO-15									
Pace Analytical Services - Minneapolis									
1,1,2,2-Tetrachloroethane	<0.60	ug/m3	2.3	0.60	1.61		09/23/21 19:45	79-34-5	
Tetrachloroethene	<0.47	ug/m3	1.1	0.47	1.61		09/23/21 19:45	127-18-4	
Tetrahydrofuran	<0.29	ug/m3	0.97	0.29	1.61		09/23/21 19:45	109-99-9	
Toluene	1.4	ug/m3	1.2	0.39	1.61		09/23/21 19:45	108-88-3	
1,2,4-Trichlorobenzene	<7.9	ug/m3	12.1	7.9	1.61		09/23/21 19:45	120-82-1	
1,1,1-Trichloroethane	<0.30	ug/m3	1.8	0.30	1.61		09/23/21 19:45	71-55-6	
1,1,2-Trichloroethane	<0.32	ug/m3	0.89	0.32	1.61		09/23/21 19:45	79-00-5	
Trichloroethene	0.86J	ug/m3	0.88	0.32	1.61		09/23/21 19:45	79-01-6	
Trichlorofluoromethane	1.2J	ug/m3	1.8	0.38	1.61		09/23/21 19:45	75-69-4	
1,1,2-Trichlorotrifluoroethane	<0.47	ug/m3	2.5	0.47	1.61		09/23/21 19:45	76-13-1	
1,2,4-Trimethylbenzene	<0.57	ug/m3	1.6	0.57	1.61		09/23/21 19:45	95-63-6	
1,3,5-Trimethylbenzene	<0.47	ug/m3	1.6	0.47	1.61		09/23/21 19:45	108-67-8	
Vinyl acetate	<0.33	ug/m3	1.2	0.33	1.61		09/23/21 19:45	108-05-4	
Vinyl chloride	<0.14	ug/m3	0.42	0.14	1.61		09/23/21 19:45	75-01-4	
m&p-Xylene	<1.0	ug/m3	2.8	1.0	1.61		09/23/21 19:45	179601-23-1	
o-Xylene	<0.44	ug/m3	1.4	0.44	1.61		09/23/21 19:45	95-47-6	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.

ANALYTICAL RESULTS

Project: CONBRACO

Pace Project No.: 92561801

Sample: IA-14 Cert# 1604 Lab ID: 92561801017 Collected: 09/14/21 17:32 Received: 09/16/21 13:15 Matrix: Air

Parameters	Results	Units	Report			Prepared	Analyzed	CAS No.	Qual
			Limit	MDL	DF				
Individual Can Certification									
Analytical Method: TO-15									
Pace Analytical Services - Minneapolis									
Acetone	<0.90	ug/m3	3.0	0.90	0.5		09/07/21 11:37	67-64-1	
Benzene	<0.057	ug/m3	0.16	0.057	0.5		09/07/21 11:37	71-43-2	
Benzyl chloride	<0.44	ug/m3	1.3	0.44	0.5		09/07/21 11:37	100-44-7	
Bromodichloromethane	<0.12	ug/m3	0.68	0.12	0.5		09/07/21 11:37	75-27-4	
Bromoform	<0.81	ug/m3	2.6	0.81	0.5		09/07/21 11:37	75-25-2	
Bromomethane	<0.075	ug/m3	0.39	0.075	0.5		09/07/21 11:37	74-83-9	
1,3-Butadiene	<0.060	ug/m3	0.22	0.060	0.5		09/07/21 11:37	106-99-0	
2-Butanone (MEK)	<0.23	ug/m3	1.5	0.23	0.5		09/07/21 11:37	78-93-3	
Carbon disulfide	<0.064	ug/m3	0.32	0.064	0.5		09/07/21 11:37	75-15-0	
Carbon tetrachloride	<0.14	ug/m3	0.64	0.14	0.5		09/07/21 11:37	56-23-5	
Chlorobenzene	<0.078	ug/m3	0.47	0.078	0.5		09/07/21 11:37	108-90-7	
Chloroethane	<0.11	ug/m3	0.27	0.11	0.5		09/07/21 11:37	75-00-3	
Chloroform	<0.092	ug/m3	0.25	0.092	0.5		09/07/21 11:37	67-66-3	
Chloromethane	<0.043	ug/m3	0.21	0.043	0.5		09/07/21 11:37	74-87-3	
Cyclohexane	<0.11	ug/m3	0.88	0.11	0.5		09/07/21 11:37	110-82-7	
Dibromochloromethane	<0.26	ug/m3	0.86	0.26	0.5		09/07/21 11:37	124-48-1	
1,2-Dibromoethane (EDB)	<0.15	ug/m3	0.39	0.15	0.5		09/07/21 11:37	106-93-4	
1,2-Dichlorobenzene	<0.20	ug/m3	1.5	0.20	0.5		09/07/21 11:37	95-50-1	
1,3-Dichlorobenzene	<0.25	ug/m3	1.5	0.25	0.5		09/07/21 11:37	541-73-1	
1,4-Dichlorobenzene	<0.44	ug/m3	1.5	0.44	0.5		09/07/21 11:37	106-46-7	
Dichlorodifluoromethane	<0.094	ug/m3	0.50	0.094	0.5		09/07/21 11:37	75-71-8	
1,1-Dichloroethane	<0.082	ug/m3	0.41	0.082	0.5		09/07/21 11:37	75-34-3	
1,2-Dichloroethane	<0.097	ug/m3	0.41	0.097	0.5		09/07/21 11:37	107-06-2	
1,1-Dichloroethene	<0.069	ug/m3	0.40	0.069	0.5		09/07/21 11:37	75-35-4	
cis-1,2-Dichloroethene	<0.098	ug/m3	0.40	0.098	0.5		09/07/21 11:37	156-59-2	
trans-1,2-Dichloroethene	<0.084	ug/m3	0.40	0.084	0.5		09/07/21 11:37	156-60-5	
1,2-Dichloropropane	<0.13	ug/m3	0.47	0.13	0.5		09/07/21 11:37	78-87-5	
cis-1,3-Dichloropropene	<0.13	ug/m3	1.2	0.13	0.5		09/07/21 11:37	10061-01-5	
trans-1,3-Dichloropropene	<0.27	ug/m3	1.2	0.27	0.5		09/07/21 11:37	10061-02-6	
Dichlorotetrafluoroethane	<0.10	ug/m3	0.71	0.10	0.5		09/07/21 11:37	76-14-2	
Ethanol	<0.30	ug/m3	0.96	0.30	0.5		09/07/21 11:37	64-17-5	
Ethyl acetate	<0.066	ug/m3	0.37	0.066	0.5		09/07/21 11:37	141-78-6	
Ethylbenzene	<0.15	ug/m3	0.44	0.15	0.5		09/07/21 11:37	100-41-4	
4-Ethyltoluene	<0.24	ug/m3	1.2	0.24	0.5		09/07/21 11:37	622-96-8	
n-Heptane	0.30J	ug/m3	0.42	0.090	0.5		09/07/21 11:37	142-82-5	
Hexachloro-1,3-butadiene	<0.62	ug/m3	2.7	0.62	0.5		09/07/21 11:37	87-68-3	
n-Hexane	<0.096	ug/m3	0.36	0.096	0.5		09/07/21 11:37	110-54-3	
2-Hexanone	<0.22	ug/m3	2.1	0.22	0.5		09/07/21 11:37	591-78-6	
Methylene Chloride	<0.30	ug/m3	1.8	0.30	0.5		09/07/21 11:37	75-09-2	
4-Methyl-2-pentanone (MIBK)	<0.16	ug/m3	2.1	0.16	0.5		09/07/21 11:37	108-10-1	
Methyl-tert-butyl ether	<0.063	ug/m3	1.8	0.063	0.5		09/07/21 11:37	1634-04-4	
Naphthalene	<1.1	ug/m3	1.3	1.1	0.5		09/07/21 11:37	91-20-3	
2-Propanol	<0.25	ug/m3	1.2	0.25	0.5		09/07/21 11:37	67-63-0	
Propylene	<0.065	ug/m3	0.44	0.065	0.5		09/07/21 11:37	115-07-1	
Styrene	<0.19	ug/m3	0.43	0.19	0.5		09/07/21 11:37	100-42-5	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.

ANALYTICAL RESULTS

Project: CONBRACO
Pace Project No.: 92561801

Sample: IA-14 Cert# 1604 **Lab ID: 92561801017** Collected: 09/14/21 17:32 Received: 09/16/21 13:15 Matrix: Air

Parameters	Results	Units	Report			Prepared	Analyzed	CAS No.	Qual
			Limit	MDL	DF				
Individual Can Certification									
Analytical Method: TO-15									
Pace Analytical Services - Minneapolis									
1,1,2,2-Tetrachloroethane	<0.19	ug/m3	0.70	0.19	0.5		09/07/21 11:37	79-34-5	
Tetrachloroethene	<0.15	ug/m3	0.34	0.15	0.5		09/07/21 11:37	127-18-4	
Tetrahydrofuran	<0.090	ug/m3	0.30	0.090	0.5		09/07/21 11:37	109-99-9	
Toluene	<0.12	ug/m3	0.38	0.12	0.5		09/07/21 11:37	108-88-3	
1,2,4-Trichlorobenzene	<2.4	ug/m3	3.8	2.4	0.5		09/07/21 11:37	120-82-1	
1,1,1-Trichloroethane	<0.093	ug/m3	0.56	0.093	0.5		09/07/21 11:37	71-55-6	
1,1,2-Trichloroethane	<0.098	ug/m3	0.28	0.098	0.5		09/07/21 11:37	79-00-5	
Trichloroethene	<0.098	ug/m3	0.27	0.098	0.5		09/07/21 11:37	79-01-6	
Trichlorofluoromethane	<0.12	ug/m3	0.57	0.12	0.5		09/07/21 11:37	75-69-4	
1,1,2-Trichlorotrifluoroethane	<0.14	ug/m3	0.78	0.14	0.5		09/07/21 11:37	76-13-1	
1,2,4-Trimethylbenzene	<0.18	ug/m3	0.50	0.18	0.5		09/07/21 11:37	95-63-6	
1,3,5-Trimethylbenzene	<0.14	ug/m3	0.50	0.14	0.5		09/07/21 11:37	108-67-8	
Vinyl acetate	<0.10	ug/m3	0.36	0.10	0.5		09/07/21 11:37	108-05-4	
Vinyl chloride	<0.043	ug/m3	0.13	0.043	0.5		09/07/21 11:37	75-01-4	
m&p-Xylene	<0.32	ug/m3	0.88	0.32	0.5		09/07/21 11:37	179601-23-1	
o-Xylene	<0.14	ug/m3	0.44	0.14	0.5		09/07/21 11:37	95-47-6	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.

ANALYTICAL RESULTS

Project: CONBRACO
Pace Project No.: 92561801

Sample: IA-6 Cert# 2319 Lab ID: 92561801018 Collected: 09/14/21 16:29 Received: 09/16/21 13:15 Matrix: Air

Parameters	Results	Units	Report			Prepared	Analyzed	CAS No.	Qual
			Limit	MDL	DF				
Individual Can Certification									
Analytical Method: TO-15									
Pace Analytical Services - Minneapolis									
Acetone	<0.90	ug/m3	3.0	0.90	0.5		09/08/21 06:15	67-64-1	
Benzene	<0.057	ug/m3	0.16	0.057	0.5		09/08/21 06:15	71-43-2	
Benzyl chloride	<0.44	ug/m3	1.3	0.44	0.5		09/08/21 06:15	100-44-7	
Bromodichloromethane	<0.12	ug/m3	0.68	0.12	0.5		09/08/21 06:15	75-27-4	
Bromoform	<0.81	ug/m3	2.6	0.81	0.5		09/08/21 06:15	75-25-2	
Bromomethane	<0.075	ug/m3	0.39	0.075	0.5		09/08/21 06:15	74-83-9	
1,3-Butadiene	<0.060	ug/m3	0.22	0.060	0.5		09/08/21 06:15	106-99-0	
2-Butanone (MEK)	<0.23	ug/m3	1.5	0.23	0.5		09/08/21 06:15	78-93-3	
Carbon disulfide	<0.064	ug/m3	0.32	0.064	0.5		09/08/21 06:15	75-15-0	
Carbon tetrachloride	<0.14	ug/m3	0.64	0.14	0.5		09/08/21 06:15	56-23-5	
Chlorobenzene	<0.078	ug/m3	0.47	0.078	0.5		09/08/21 06:15	108-90-7	
Chloroethane	<0.11	ug/m3	0.27	0.11	0.5		09/08/21 06:15	75-00-3	
Chloroform	<0.092	ug/m3	0.25	0.092	0.5		09/08/21 06:15	67-66-3	
Chloromethane	<0.043	ug/m3	0.21	0.043	0.5		09/08/21 06:15	74-87-3	
Cyclohexane	<0.11	ug/m3	0.88	0.11	0.5		09/08/21 06:15	110-82-7	
Dibromochloromethane	<0.26	ug/m3	0.86	0.26	0.5		09/08/21 06:15	124-48-1	
1,2-Dibromoethane (EDB)	<0.15	ug/m3	0.39	0.15	0.5		09/08/21 06:15	106-93-4	
1,2-Dichlorobenzene	<0.20	ug/m3	1.5	0.20	0.5		09/08/21 06:15	95-50-1	
1,3-Dichlorobenzene	<0.25	ug/m3	1.5	0.25	0.5		09/08/21 06:15	541-73-1	
1,4-Dichlorobenzene	<0.44	ug/m3	1.5	0.44	0.5		09/08/21 06:15	106-46-7	
Dichlorodifluoromethane	<0.094	ug/m3	0.50	0.094	0.5		09/08/21 06:15	75-71-8	
1,1-Dichloroethane	<0.082	ug/m3	0.41	0.082	0.5		09/08/21 06:15	75-34-3	
1,2-Dichloroethane	<0.097	ug/m3	0.41	0.097	0.5		09/08/21 06:15	107-06-2	
1,1-Dichloroethene	<0.069	ug/m3	0.40	0.069	0.5		09/08/21 06:15	75-35-4	
cis-1,2-Dichloroethene	<0.098	ug/m3	0.40	0.098	0.5		09/08/21 06:15	156-59-2	
trans-1,2-Dichloroethene	<0.084	ug/m3	0.40	0.084	0.5		09/08/21 06:15	156-60-5	
1,2-Dichloropropane	<0.13	ug/m3	0.47	0.13	0.5		09/08/21 06:15	78-87-5	
cis-1,3-Dichloropropene	<0.13	ug/m3	1.2	0.13	0.5		09/08/21 06:15	10061-01-5	
trans-1,3-Dichloropropene	<0.27	ug/m3	1.2	0.27	0.5		09/08/21 06:15	10061-02-6	
Dichlorotetrafluoroethane	<0.10	ug/m3	0.71	0.10	0.5		09/08/21 06:15	76-14-2	
Ethanol	<0.30	ug/m3	0.96	0.30	0.5		09/08/21 06:15	64-17-5	
Ethyl acetate	<0.066	ug/m3	0.37	0.066	0.5		09/08/21 06:15	141-78-6	
Ethylbenzene	<0.15	ug/m3	0.44	0.15	0.5		09/08/21 06:15	100-41-4	
4-Ethyltoluene	<0.24	ug/m3	1.2	0.24	0.5		09/08/21 06:15	622-96-8	
n-Heptane	<0.090	ug/m3	0.42	0.090	0.5		09/08/21 06:15	142-82-5	
Hexachloro-1,3-butadiene	<0.62	ug/m3	2.7	0.62	0.5		09/08/21 06:15	87-68-3	
n-Hexane	<0.096	ug/m3	0.36	0.096	0.5		09/08/21 06:15	110-54-3	
2-Hexanone	<0.22	ug/m3	2.1	0.22	0.5		09/08/21 06:15	591-78-6	
Methylene Chloride	<0.30	ug/m3	1.8	0.30	0.5		09/08/21 06:15	75-09-2	
4-Methyl-2-pentanone (MIBK)	<0.16	ug/m3	2.1	0.16	0.5		09/08/21 06:15	108-10-1	
Methyl-tert-butyl ether	<0.063	ug/m3	1.8	0.063	0.5		09/08/21 06:15	1634-04-4	
Naphthalene	<1.1	ug/m3	1.3	1.1	0.5		09/08/21 06:15	91-20-3	
2-Propanol	<0.25	ug/m3	1.2	0.25	0.5		09/08/21 06:15	67-63-0	
Propylene	<0.065	ug/m3	0.44	0.065	0.5		09/08/21 06:15	115-07-1	
Styrene	<0.19	ug/m3	0.43	0.19	0.5		09/08/21 06:15	100-42-5	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.

ANALYTICAL RESULTS

Project: CONBRACO
Pace Project No.: 92561801

Sample: IA-6 Cert# 2319 **Lab ID: 92561801018** Collected: 09/14/21 16:29 Received: 09/16/21 13:15 Matrix: Air

Parameters	Results	Units	Report			Prepared	Analyzed	CAS No.	Qual
			Limit	MDL	DF				
Individual Can Certification									
Analytical Method: TO-15									
Pace Analytical Services - Minneapolis									
1,1,2,2-Tetrachloroethane	<0.19	ug/m3	0.70	0.19	0.5		09/08/21 06:15	79-34-5	
Tetrachloroethene	<0.15	ug/m3	0.34	0.15	0.5		09/08/21 06:15	127-18-4	
Tetrahydrofuran	<0.090	ug/m3	0.30	0.090	0.5		09/08/21 06:15	109-99-9	
Toluene	<0.12	ug/m3	0.38	0.12	0.5		09/08/21 06:15	108-88-3	
1,2,4-Trichlorobenzene	<2.4	ug/m3	3.8	2.4	0.5		09/08/21 06:15	120-82-1	
1,1,1-Trichloroethane	<0.093	ug/m3	0.56	0.093	0.5		09/08/21 06:15	71-55-6	
1,1,2-Trichloroethane	<0.098	ug/m3	0.28	0.098	0.5		09/08/21 06:15	79-00-5	
Trichloroethene	<0.098	ug/m3	0.27	0.098	0.5		09/08/21 06:15	79-01-6	
Trichlorofluoromethane	<0.12	ug/m3	0.57	0.12	0.5		09/08/21 06:15	75-69-4	
1,1,2-Trichlorotrifluoroethane	<0.14	ug/m3	0.78	0.14	0.5		09/08/21 06:15	76-13-1	
1,2,4-Trimethylbenzene	<0.18	ug/m3	0.50	0.18	0.5		09/08/21 06:15	95-63-6	
1,3,5-Trimethylbenzene	<0.14	ug/m3	0.50	0.14	0.5		09/08/21 06:15	108-67-8	
Vinyl acetate	<0.10	ug/m3	0.36	0.10	0.5		09/08/21 06:15	108-05-4	
Vinyl chloride	<0.043	ug/m3	0.13	0.043	0.5		09/08/21 06:15	75-01-4	
m&p-Xylene	<0.32	ug/m3	0.88	0.32	0.5		09/08/21 06:15	179601-23-1	
o-Xylene	<0.14	ug/m3	0.44	0.14	0.5		09/08/21 06:15	95-47-6	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.

ANALYTICAL RESULTS

Project: CONBRACO

Pace Project No.: 92561801

Sample: IA-1 Cert #0799 Lab ID: 92561801019 Collected: 09/14/21 18:15 Received: 09/16/21 13:15 Matrix: Air

Parameters	Results	Units	Report			Prepared	Analyzed	CAS No.	Qual
			Limit	MDL	DF				
Individual Can Certification									
Analytical Method: TO-15									
Pace Analytical Services - Minneapolis									
Acetone	<0.90	ug/m3	3.0	0.90	0.5		09/07/21 12:18	67-64-1	
Benzene	<0.057	ug/m3	0.16	0.057	0.5		09/07/21 12:18	71-43-2	
Benzyl chloride	<0.44	ug/m3	1.3	0.44	0.5		09/07/21 12:18	100-44-7	
Bromodichloromethane	<0.12	ug/m3	0.68	0.12	0.5		09/07/21 12:18	75-27-4	
Bromoform	<0.81	ug/m3	2.6	0.81	0.5		09/07/21 12:18	75-25-2	
Bromomethane	<0.075	ug/m3	0.39	0.075	0.5		09/07/21 12:18	74-83-9	
1,3-Butadiene	<0.060	ug/m3	0.22	0.060	0.5		09/07/21 12:18	106-99-0	
2-Butanone (MEK)	<0.23	ug/m3	1.5	0.23	0.5		09/07/21 12:18	78-93-3	
Carbon disulfide	<0.064	ug/m3	0.32	0.064	0.5		09/07/21 12:18	75-15-0	
Carbon tetrachloride	<0.14	ug/m3	0.64	0.14	0.5		09/07/21 12:18	56-23-5	
Chlorobenzene	<0.078	ug/m3	0.47	0.078	0.5		09/07/21 12:18	108-90-7	
Chloroethane	<0.11	ug/m3	0.27	0.11	0.5		09/07/21 12:18	75-00-3	
Chloroform	<0.092	ug/m3	0.25	0.092	0.5		09/07/21 12:18	67-66-3	
Chloromethane	<0.043	ug/m3	0.21	0.043	0.5		09/07/21 12:18	74-87-3	
Cyclohexane	<0.11	ug/m3	0.88	0.11	0.5		09/07/21 12:18	110-82-7	
Dibromochloromethane	<0.26	ug/m3	0.86	0.26	0.5		09/07/21 12:18	124-48-1	
1,2-Dibromoethane (EDB)	<0.15	ug/m3	0.39	0.15	0.5		09/07/21 12:18	106-93-4	
1,2-Dichlorobenzene	<0.20	ug/m3	1.5	0.20	0.5		09/07/21 12:18	95-50-1	
1,3-Dichlorobenzene	<0.25	ug/m3	1.5	0.25	0.5		09/07/21 12:18	541-73-1	
1,4-Dichlorobenzene	<0.44	ug/m3	1.5	0.44	0.5		09/07/21 12:18	106-46-7	
Dichlorodifluoromethane	<0.094	ug/m3	0.50	0.094	0.5		09/07/21 12:18	75-71-8	
1,1-Dichloroethane	<0.082	ug/m3	0.41	0.082	0.5		09/07/21 12:18	75-34-3	
1,2-Dichloroethane	<0.097	ug/m3	0.41	0.097	0.5		09/07/21 12:18	107-06-2	
1,1-Dichloroethene	<0.069	ug/m3	0.40	0.069	0.5		09/07/21 12:18	75-35-4	
cis-1,2-Dichloroethene	<0.098	ug/m3	0.40	0.098	0.5		09/07/21 12:18	156-59-2	
trans-1,2-Dichloroethene	<0.084	ug/m3	0.40	0.084	0.5		09/07/21 12:18	156-60-5	
1,2-Dichloropropane	<0.13	ug/m3	0.47	0.13	0.5		09/07/21 12:18	78-87-5	
cis-1,3-Dichloropropene	<0.13	ug/m3	1.2	0.13	0.5		09/07/21 12:18	10061-01-5	
trans-1,3-Dichloropropene	<0.27	ug/m3	1.2	0.27	0.5		09/07/21 12:18	10061-02-6	
Dichlorotetrafluoroethane	<0.10	ug/m3	0.71	0.10	0.5		09/07/21 12:18	76-14-2	
Ethanol	<0.30	ug/m3	0.96	0.30	0.5		09/07/21 12:18	64-17-5	
Ethyl acetate	<0.066	ug/m3	0.37	0.066	0.5		09/07/21 12:18	141-78-6	
Ethylbenzene	<0.15	ug/m3	0.44	0.15	0.5		09/07/21 12:18	100-41-4	
4-Ethyltoluene	<0.24	ug/m3	1.2	0.24	0.5		09/07/21 12:18	622-96-8	
n-Heptane	<0.090	ug/m3	0.42	0.090	0.5		09/07/21 12:18	142-82-5	
Hexachloro-1,3-butadiene	<0.62	ug/m3	2.7	0.62	0.5		09/07/21 12:18	87-68-3	
n-Hexane	<0.096	ug/m3	0.36	0.096	0.5		09/07/21 12:18	110-54-3	
2-Hexanone	<0.22	ug/m3	2.1	0.22	0.5		09/07/21 12:18	591-78-6	
Methylene Chloride	<0.30	ug/m3	1.8	0.30	0.5		09/07/21 12:18	75-09-2	
4-Methyl-2-pentanone (MIBK)	<0.16	ug/m3	2.1	0.16	0.5		09/07/21 12:18	108-10-1	
Methyl-tert-butyl ether	<0.063	ug/m3	1.8	0.063	0.5		09/07/21 12:18	1634-04-4	
Naphthalene	<1.1	ug/m3	1.3	1.1	0.5		09/07/21 12:18	91-20-3	
2-Propanol	<0.25	ug/m3	1.2	0.25	0.5		09/07/21 12:18	67-63-0	
Propylene	<0.065	ug/m3	0.44	0.065	0.5		09/07/21 12:18	115-07-1	
Styrene	<0.19	ug/m3	0.43	0.19	0.5		09/07/21 12:18	100-42-5	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.

ANALYTICAL RESULTS

Project: CONBRACO
Pace Project No.: 92561801

Sample: IA-1 Cert #0799 **Lab ID: 92561801019** Collected: 09/14/21 18:15 Received: 09/16/21 13:15 Matrix: Air

Parameters	Results	Units	Report			Prepared	Analyzed	CAS No.	Qual
			Limit	MDL	DF				
Individual Can Certification									
Analytical Method: TO-15									
Pace Analytical Services - Minneapolis									
1,1,2,2-Tetrachloroethane	<0.19	ug/m3	0.70	0.19	0.5		09/07/21 12:18	79-34-5	
Tetrachloroethene	<0.15	ug/m3	0.34	0.15	0.5		09/07/21 12:18	127-18-4	
Tetrahydrofuran	<0.090	ug/m3	0.30	0.090	0.5		09/07/21 12:18	109-99-9	
Toluene	<0.12	ug/m3	0.38	0.12	0.5		09/07/21 12:18	108-88-3	
1,2,4-Trichlorobenzene	<2.4	ug/m3	3.8	2.4	0.5		09/07/21 12:18	120-82-1	
1,1,1-Trichloroethane	<0.093	ug/m3	0.56	0.093	0.5		09/07/21 12:18	71-55-6	
1,1,2-Trichloroethane	<0.098	ug/m3	0.28	0.098	0.5		09/07/21 12:18	79-00-5	
Trichloroethene	<0.098	ug/m3	0.27	0.098	0.5		09/07/21 12:18	79-01-6	
Trichlorofluoromethane	<0.12	ug/m3	0.57	0.12	0.5		09/07/21 12:18	75-69-4	
1,1,2-Trichlorotrifluoroethane	<0.14	ug/m3	0.78	0.14	0.5		09/07/21 12:18	76-13-1	
1,2,4-Trimethylbenzene	<0.18	ug/m3	0.50	0.18	0.5		09/07/21 12:18	95-63-6	
1,3,5-Trimethylbenzene	<0.14	ug/m3	0.50	0.14	0.5		09/07/21 12:18	108-67-8	
Vinyl acetate	<0.10	ug/m3	0.36	0.10	0.5		09/07/21 12:18	108-05-4	
Vinyl chloride	<0.043	ug/m3	0.13	0.043	0.5		09/07/21 12:18	75-01-4	
m&p-Xylene	<0.32	ug/m3	0.88	0.32	0.5		09/07/21 12:18	179601-23-1	
o-Xylene	<0.14	ug/m3	0.44	0.14	0.5		09/07/21 12:18	95-47-6	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.

ANALYTICAL RESULTS

Project: CONBRACO
Pace Project No.: 92561801

Sample: IA-12 Cert# 0145 **Lab ID: 92561801020** Collected: 09/14/21 17:35 Received: 09/16/21 13:15 Matrix: Air

Parameters	Results	Units	Report			Prepared	Analyzed	CAS No.	Qual
			Limit	MDL	DF				
Individual Can Certification									
Analytical Method: TO-15									
Pace Analytical Services - Minneapolis									
Acetone	<0.90	ug/m3	3.0	0.90	0.5		09/07/21 12:44	67-64-1	
Benzene	<0.057	ug/m3	0.16	0.057	0.5		09/07/21 12:44	71-43-2	
Benzyl chloride	<0.44	ug/m3	1.3	0.44	0.5		09/07/21 12:44	100-44-7	
Bromodichloromethane	<0.12	ug/m3	0.68	0.12	0.5		09/07/21 12:44	75-27-4	
Bromoform	<0.81	ug/m3	2.6	0.81	0.5		09/07/21 12:44	75-25-2	
Bromomethane	<0.075	ug/m3	0.39	0.075	0.5		09/07/21 12:44	74-83-9	
1,3-Butadiene	<0.060	ug/m3	0.22	0.060	0.5		09/07/21 12:44	106-99-0	
2-Butanone (MEK)	<0.23	ug/m3	1.5	0.23	0.5		09/07/21 12:44	78-93-3	
Carbon disulfide	<0.064	ug/m3	0.32	0.064	0.5		09/07/21 12:44	75-15-0	
Carbon tetrachloride	<0.14	ug/m3	0.64	0.14	0.5		09/07/21 12:44	56-23-5	
Chlorobenzene	<0.078	ug/m3	0.47	0.078	0.5		09/07/21 12:44	108-90-7	
Chloroethane	<0.11	ug/m3	0.27	0.11	0.5		09/07/21 12:44	75-00-3	
Chloroform	<0.092	ug/m3	0.25	0.092	0.5		09/07/21 12:44	67-66-3	
Chloromethane	<0.043	ug/m3	0.21	0.043	0.5		09/07/21 12:44	74-87-3	
Cyclohexane	<0.11	ug/m3	0.88	0.11	0.5		09/07/21 12:44	110-82-7	
Dibromochloromethane	<0.26	ug/m3	0.86	0.26	0.5		09/07/21 12:44	124-48-1	
1,2-Dibromoethane (EDB)	<0.15	ug/m3	0.39	0.15	0.5		09/07/21 12:44	106-93-4	
1,2-Dichlorobenzene	<0.20	ug/m3	1.5	0.20	0.5		09/07/21 12:44	95-50-1	
1,3-Dichlorobenzene	<0.25	ug/m3	1.5	0.25	0.5		09/07/21 12:44	541-73-1	
1,4-Dichlorobenzene	<0.44	ug/m3	1.5	0.44	0.5		09/07/21 12:44	106-46-7	
Dichlorodifluoromethane	<0.094	ug/m3	0.50	0.094	0.5		09/07/21 12:44	75-71-8	
1,1-Dichloroethane	<0.082	ug/m3	0.41	0.082	0.5		09/07/21 12:44	75-34-3	
1,2-Dichloroethane	<0.097	ug/m3	0.41	0.097	0.5		09/07/21 12:44	107-06-2	
1,1-Dichloroethene	<0.069	ug/m3	0.40	0.069	0.5		09/07/21 12:44	75-35-4	
cis-1,2-Dichloroethene	<0.098	ug/m3	0.40	0.098	0.5		09/07/21 12:44	156-59-2	
trans-1,2-Dichloroethene	<0.084	ug/m3	0.40	0.084	0.5		09/07/21 12:44	156-60-5	
1,2-Dichloropropane	<0.13	ug/m3	0.47	0.13	0.5		09/07/21 12:44	78-87-5	
cis-1,3-Dichloropropene	<0.13	ug/m3	1.2	0.13	0.5		09/07/21 12:44	10061-01-5	
trans-1,3-Dichloropropene	<0.27	ug/m3	1.2	0.27	0.5		09/07/21 12:44	10061-02-6	
Dichlorotetrafluoroethane	<0.10	ug/m3	0.71	0.10	0.5		09/07/21 12:44	76-14-2	
Ethanol	<0.30	ug/m3	0.96	0.30	0.5		09/07/21 12:44	64-17-5	
Ethyl acetate	<0.066	ug/m3	0.37	0.066	0.5		09/07/21 12:44	141-78-6	
Ethylbenzene	<0.15	ug/m3	0.44	0.15	0.5		09/07/21 12:44	100-41-4	
4-Ethyltoluene	<0.24	ug/m3	1.2	0.24	0.5		09/07/21 12:44	622-96-8	
n-Heptane	<0.090	ug/m3	0.42	0.090	0.5		09/07/21 12:44	142-82-5	
Hexachloro-1,3-butadiene	<0.62	ug/m3	2.7	0.62	0.5		09/07/21 12:44	87-68-3	
n-Hexane	<0.096	ug/m3	0.36	0.096	0.5		09/07/21 12:44	110-54-3	
2-Hexanone	<0.22	ug/m3	2.1	0.22	0.5		09/07/21 12:44	591-78-6	
Methylene Chloride	<0.30	ug/m3	1.8	0.30	0.5		09/07/21 12:44	75-09-2	
4-Methyl-2-pentanone (MIBK)	<0.16	ug/m3	2.1	0.16	0.5		09/07/21 12:44	108-10-1	
Methyl-tert-butyl ether	<0.063	ug/m3	1.8	0.063	0.5		09/07/21 12:44	1634-04-4	
Naphthalene	<1.1	ug/m3	1.3	1.1	0.5		09/07/21 12:44	91-20-3	
2-Propanol	<0.25	ug/m3	1.2	0.25	0.5		09/07/21 12:44	67-63-0	
Propylene	<0.065	ug/m3	0.44	0.065	0.5		09/07/21 12:44	115-07-1	
Styrene	<0.19	ug/m3	0.43	0.19	0.5		09/07/21 12:44	100-42-5	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.

ANALYTICAL RESULTS

Project: CONBRACO
Pace Project No.: 92561801

Sample: IA-12 Cert# 0145 **Lab ID: 92561801020** Collected: 09/14/21 17:35 Received: 09/16/21 13:15 Matrix: Air

Parameters	Results	Units	Report			Prepared	Analyzed	CAS No.	Qual
			Limit	MDL	DF				
Individual Can Certification									
Analytical Method: TO-15									
Pace Analytical Services - Minneapolis									
1,1,2,2-Tetrachloroethane	<0.19	ug/m3	0.70	0.19	0.5		09/07/21 12:44	79-34-5	
Tetrachloroethene	<0.15	ug/m3	0.34	0.15	0.5		09/07/21 12:44	127-18-4	
Tetrahydrofuran	<0.090	ug/m3	0.30	0.090	0.5		09/07/21 12:44	109-99-9	
Toluene	<0.12	ug/m3	0.38	0.12	0.5		09/07/21 12:44	108-88-3	
1,2,4-Trichlorobenzene	<2.4	ug/m3	3.8	2.4	0.5		09/07/21 12:44	120-82-1	
1,1,1-Trichloroethane	<0.093	ug/m3	0.56	0.093	0.5		09/07/21 12:44	71-55-6	
1,1,2-Trichloroethane	<0.098	ug/m3	0.28	0.098	0.5		09/07/21 12:44	79-00-5	
Trichloroethene	<0.098	ug/m3	0.27	0.098	0.5		09/07/21 12:44	79-01-6	
Trichlorofluoromethane	<0.12	ug/m3	0.57	0.12	0.5		09/07/21 12:44	75-69-4	
1,1,2-Trichlorotrifluoroethane	<0.14	ug/m3	0.78	0.14	0.5		09/07/21 12:44	76-13-1	
1,2,4-Trimethylbenzene	<0.18	ug/m3	0.50	0.18	0.5		09/07/21 12:44	95-63-6	
1,3,5-Trimethylbenzene	<0.14	ug/m3	0.50	0.14	0.5		09/07/21 12:44	108-67-8	
Vinyl acetate	<0.10	ug/m3	0.36	0.10	0.5		09/07/21 12:44	108-05-4	
Vinyl chloride	<0.043	ug/m3	0.13	0.043	0.5		09/07/21 12:44	75-01-4	
m&p-Xylene	<0.32	ug/m3	0.88	0.32	0.5		09/07/21 12:44	179601-23-1	
o-Xylene	<0.14	ug/m3	0.44	0.14	0.5		09/07/21 12:44	95-47-6	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.

ANALYTICAL RESULTS

Project: CONBRACO

Pace Project No.: 92561801

Sample: IA-2 Cert# 0796 Lab ID: 92561801021 Collected: 09/14/21 18:16 Received: 09/16/21 13:15 Matrix: Air

Parameters	Results	Units	Report			Prepared	Analyzed	CAS No.	Qual
			Limit	MDL	DF				
Individual Can Certification									
Analytical Method: TO-15									
Pace Analytical Services - Minneapolis									
Acetone	<0.90	ug/m3	3.0	0.90	0.5		09/07/21 11:19	67-64-1	
Benzene	<0.057	ug/m3	0.16	0.057	0.5		09/07/21 11:19	71-43-2	
Benzyl chloride	<0.44	ug/m3	1.3	0.44	0.5		09/07/21 11:19	100-44-7	
Bromodichloromethane	<0.12	ug/m3	0.68	0.12	0.5		09/07/21 11:19	75-27-4	
Bromoform	<0.81	ug/m3	2.6	0.81	0.5		09/07/21 11:19	75-25-2	
Bromomethane	<0.075	ug/m3	0.39	0.075	0.5		09/07/21 11:19	74-83-9	
1,3-Butadiene	<0.060	ug/m3	0.22	0.060	0.5		09/07/21 11:19	106-99-0	
2-Butanone (MEK)	<0.23	ug/m3	1.5	0.23	0.5		09/07/21 11:19	78-93-3	
Carbon disulfide	<0.064	ug/m3	0.32	0.064	0.5		09/07/21 11:19	75-15-0	
Carbon tetrachloride	<0.14	ug/m3	0.64	0.14	0.5		09/07/21 11:19	56-23-5	
Chlorobenzene	<0.078	ug/m3	0.47	0.078	0.5		09/07/21 11:19	108-90-7	
Chloroethane	<0.11	ug/m3	0.27	0.11	0.5		09/07/21 11:19	75-00-3	
Chloroform	<0.092	ug/m3	0.25	0.092	0.5		09/07/21 11:19	67-66-3	
Chloromethane	<0.043	ug/m3	0.21	0.043	0.5		09/07/21 11:19	74-87-3	
Cyclohexane	<0.11	ug/m3	0.88	0.11	0.5		09/07/21 11:19	110-82-7	
Dibromochloromethane	<0.26	ug/m3	0.86	0.26	0.5		09/07/21 11:19	124-48-1	
1,2-Dibromoethane (EDB)	<0.15	ug/m3	0.39	0.15	0.5		09/07/21 11:19	106-93-4	
1,2-Dichlorobenzene	<0.20	ug/m3	1.5	0.20	0.5		09/07/21 11:19	95-50-1	
1,3-Dichlorobenzene	<0.25	ug/m3	1.5	0.25	0.5		09/07/21 11:19	541-73-1	
1,4-Dichlorobenzene	<0.44	ug/m3	1.5	0.44	0.5		09/07/21 11:19	106-46-7	
Dichlorodifluoromethane	<0.094	ug/m3	0.50	0.094	0.5		09/07/21 11:19	75-71-8	
1,1-Dichloroethane	<0.082	ug/m3	0.41	0.082	0.5		09/07/21 11:19	75-34-3	
1,2-Dichloroethane	<0.097	ug/m3	0.41	0.097	0.5		09/07/21 11:19	107-06-2	
1,1-Dichloroethene	<0.069	ug/m3	0.40	0.069	0.5		09/07/21 11:19	75-35-4	
cis-1,2-Dichloroethene	<0.098	ug/m3	0.40	0.098	0.5		09/07/21 11:19	156-59-2	
trans-1,2-Dichloroethene	<0.084	ug/m3	0.40	0.084	0.5		09/07/21 11:19	156-60-5	
1,2-Dichloropropane	<0.13	ug/m3	0.47	0.13	0.5		09/07/21 11:19	78-87-5	
cis-1,3-Dichloropropene	<0.13	ug/m3	1.2	0.13	0.5		09/07/21 11:19	10061-01-5	
trans-1,3-Dichloropropene	<0.27	ug/m3	1.2	0.27	0.5		09/07/21 11:19	10061-02-6	
Dichlorotetrafluoroethane	<0.10	ug/m3	0.71	0.10	0.5		09/07/21 11:19	76-14-2	
Ethanol	<0.30	ug/m3	0.96	0.30	0.5		09/07/21 11:19	64-17-5	
Ethyl acetate	<0.066	ug/m3	0.37	0.066	0.5		09/07/21 11:19	141-78-6	
Ethylbenzene	<0.15	ug/m3	0.44	0.15	0.5		09/07/21 11:19	100-41-4	
4-Ethyltoluene	<0.24	ug/m3	1.2	0.24	0.5		09/07/21 11:19	622-96-8	
n-Heptane	<0.090	ug/m3	0.42	0.090	0.5		09/07/21 11:19	142-82-5	
Hexachloro-1,3-butadiene	<0.62	ug/m3	2.7	0.62	0.5		09/07/21 11:19	87-68-3	
n-Hexane	<0.096	ug/m3	0.36	0.096	0.5		09/07/21 11:19	110-54-3	
2-Hexanone	<0.22	ug/m3	2.1	0.22	0.5		09/07/21 11:19	591-78-6	
Methylene Chloride	<0.30	ug/m3	1.8	0.30	0.5		09/07/21 11:19	75-09-2	
4-Methyl-2-pentanone (MIBK)	<0.16	ug/m3	2.1	0.16	0.5		09/07/21 11:19	108-10-1	
Methyl-tert-butyl ether	<0.063	ug/m3	1.8	0.063	0.5		09/07/21 11:19	1634-04-4	
Naphthalene	<1.1	ug/m3	1.3	1.1	0.5		09/07/21 11:19	91-20-3	
2-Propanol	<0.25	ug/m3	1.2	0.25	0.5		09/07/21 11:19	67-63-0	
Propylene	<0.065	ug/m3	0.44	0.065	0.5		09/07/21 11:19	115-07-1	
Styrene	<0.19	ug/m3	0.43	0.19	0.5		09/07/21 11:19	100-42-5	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.

ANALYTICAL RESULTS

Project: CONBRACO
Pace Project No.: 92561801

Sample: IA-2 Cert# 0796 **Lab ID: 92561801021** Collected: 09/14/21 18:16 Received: 09/16/21 13:15 Matrix: Air

Parameters	Results	Units	Report			Prepared	Analyzed	CAS No.	Qual
			Limit	MDL	DF				
Individual Can Certification									
Analytical Method: TO-15									
Pace Analytical Services - Minneapolis									
1,1,2,2-Tetrachloroethane	<0.19	ug/m3	0.70	0.19	0.5		09/07/21 11:19	79-34-5	
Tetrachloroethene	<0.15	ug/m3	0.34	0.15	0.5		09/07/21 11:19	127-18-4	
Tetrahydrofuran	<0.090	ug/m3	0.30	0.090	0.5		09/07/21 11:19	109-99-9	
Toluene	<0.12	ug/m3	0.38	0.12	0.5		09/07/21 11:19	108-88-3	
1,2,4-Trichlorobenzene	<2.4	ug/m3	3.8	2.4	0.5		09/07/21 11:19	120-82-1	
1,1,1-Trichloroethane	<0.093	ug/m3	0.56	0.093	0.5		09/07/21 11:19	71-55-6	
1,1,2-Trichloroethane	<0.098	ug/m3	0.28	0.098	0.5		09/07/21 11:19	79-00-5	
Trichloroethene	<0.098	ug/m3	0.27	0.098	0.5		09/07/21 11:19	79-01-6	
Trichlorofluoromethane	<0.12	ug/m3	0.57	0.12	0.5		09/07/21 11:19	75-69-4	
1,1,2-Trichlorotrifluoroethane	<0.14	ug/m3	0.78	0.14	0.5		09/07/21 11:19	76-13-1	
1,2,4-Trimethylbenzene	<0.18	ug/m3	0.50	0.18	0.5		09/07/21 11:19	95-63-6	
1,3,5-Trimethylbenzene	<0.14	ug/m3	0.50	0.14	0.5		09/07/21 11:19	108-67-8	
Vinyl acetate	<0.10	ug/m3	0.36	0.10	0.5		09/07/21 11:19	108-05-4	
Vinyl chloride	<0.043	ug/m3	0.13	0.043	0.5		09/07/21 11:19	75-01-4	
m&p-Xylene	<0.32	ug/m3	0.88	0.32	0.5		09/07/21 11:19	179601-23-1	
o-Xylene	<0.14	ug/m3	0.44	0.14	0.5		09/07/21 11:19	95-47-6	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.

ANALYTICAL RESULTS

Project: CONBRACO
Pace Project No.: 92561801

Sample: IA-8 Cert# 2109 **Lab ID: 92561801022** Collected: 09/14/21 16:52 Received: 09/16/21 13:15 Matrix: Air

Parameters	Results	Units	Report			Prepared	Analyzed	CAS No.	Qual
			Limit	MDL	DF				
Individual Can Certification									
Analytical Method: TO-15									
Pace Analytical Services - Minneapolis									
Acetone	<1.8	ug/m3	6.0	1.8	1		09/08/21 09:54	67-64-1	
Benzene	<0.11	ug/m3	0.32	0.11	1		09/08/21 09:54	71-43-2	
Benzyl chloride	<0.89	ug/m3	2.6	0.89	1		09/08/21 09:54	100-44-7	
Bromodichloromethane	<0.24	ug/m3	1.4	0.24	1		09/08/21 09:54	75-27-4	
Bromoform	<1.6	ug/m3	5.2	1.6	1		09/08/21 09:54	75-25-2	
Bromomethane	<0.15	ug/m3	0.79	0.15	1		09/08/21 09:54	74-83-9	
1,3-Butadiene	<0.12	ug/m3	0.45	0.12	1		09/08/21 09:54	106-99-0	
2-Butanone (MEK)	<0.46	ug/m3	3.0	0.46	1		09/08/21 09:54	78-93-3	
Carbon disulfide	<0.13	ug/m3	0.63	0.13	1		09/08/21 09:54	75-15-0	
Carbon tetrachloride	<0.28	ug/m3	1.3	0.28	1		09/08/21 09:54	56-23-5	
Chlorobenzene	<0.16	ug/m3	0.94	0.16	1		09/08/21 09:54	108-90-7	
Chloroethane	<0.22	ug/m3	0.54	0.22	1		09/08/21 09:54	75-00-3	
Chloroform	<0.18	ug/m3	0.50	0.18	1		09/08/21 09:54	67-66-3	
Chloromethane	<0.085	ug/m3	0.42	0.085	1		09/08/21 09:54	74-87-3	
Cyclohexane	<0.22	ug/m3	1.8	0.22	1		09/08/21 09:54	110-82-7	
Dibromochloromethane	<0.52	ug/m3	1.7	0.52	1		09/08/21 09:54	124-48-1	
1,2-Dibromoethane (EDB)	<0.30	ug/m3	0.78	0.30	1		09/08/21 09:54	106-93-4	
1,2-Dichlorobenzene	<0.40	ug/m3	3.1	0.40	1		09/08/21 09:54	95-50-1	
1,3-Dichlorobenzene	<0.51	ug/m3	3.1	0.51	1		09/08/21 09:54	541-73-1	
1,4-Dichlorobenzene	<0.88	ug/m3	3.1	0.88	1		09/08/21 09:54	106-46-7	
Dichlorodifluoromethane	<0.19	ug/m3	1.0	0.19	1		09/08/21 09:54	75-71-8	
1,1-Dichloroethane	<0.16	ug/m3	0.82	0.16	1		09/08/21 09:54	75-34-3	
1,2-Dichloroethane	<0.19	ug/m3	0.82	0.19	1		09/08/21 09:54	107-06-2	
1,1-Dichloroethene	<0.14	ug/m3	0.81	0.14	1		09/08/21 09:54	75-35-4	
cis-1,2-Dichloroethene	<0.20	ug/m3	0.81	0.20	1		09/08/21 09:54	156-59-2	
trans-1,2-Dichloroethene	<0.17	ug/m3	0.81	0.17	1		09/08/21 09:54	156-60-5	
1,2-Dichloropropane	<0.27	ug/m3	0.94	0.27	1		09/08/21 09:54	78-87-5	
cis-1,3-Dichloropropene	<0.26	ug/m3	2.3	0.26	1		09/08/21 09:54	10061-01-5	
trans-1,3-Dichloropropene	<0.54	ug/m3	2.3	0.54	1		09/08/21 09:54	10061-02-6	
Dichlorotetrafluoroethane	<0.20	ug/m3	1.4	0.20	1		09/08/21 09:54	76-14-2	
Ethanol	<0.59	ug/m3	1.9	0.59	1		09/08/21 09:54	64-17-5	
Ethyl acetate	<0.13	ug/m3	0.73	0.13	1		09/08/21 09:54	141-78-6	
Ethylbenzene	<0.31	ug/m3	0.88	0.31	1		09/08/21 09:54	100-41-4	
4-Ethyltoluene	<0.47	ug/m3	2.5	0.47	1		09/08/21 09:54	622-96-8	
n-Heptane	<0.18	ug/m3	0.83	0.18	1		09/08/21 09:54	142-82-5	
Hexachloro-1,3-butadiene	<1.2	ug/m3	5.4	1.2	1		09/08/21 09:54	87-68-3	
n-Hexane	<0.19	ug/m3	0.72	0.19	1		09/08/21 09:54	110-54-3	
2-Hexanone	<0.44	ug/m3	4.2	0.44	1		09/08/21 09:54	591-78-6	
Methylene Chloride	<0.59	ug/m3	3.5	0.59	1		09/08/21 09:54	75-09-2	
4-Methyl-2-pentanone (MIBK)	<0.32	ug/m3	4.2	0.32	1		09/08/21 09:54	108-10-1	
Methyl-tert-butyl ether	<0.13	ug/m3	3.7	0.13	1		09/08/21 09:54	1634-04-4	
Naphthalene	<2.2	ug/m3	2.7	2.2	1		09/08/21 09:54	91-20-3	
2-Propanol	<0.51	ug/m3	2.5	0.51	1		09/08/21 09:54	67-63-0	
Propylene	<0.13	ug/m3	0.88	0.13	1		09/08/21 09:54	115-07-1	
Styrene	<0.38	ug/m3	0.87	0.38	1		09/08/21 09:54	100-42-5	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.

ANALYTICAL RESULTS

Project: CONBRACO
Pace Project No.: 92561801

Sample: IA-8 Cert# 2109 **Lab ID: 92561801022** Collected: 09/14/21 16:52 Received: 09/16/21 13:15 Matrix: Air

Parameters	Results	Units	Report			Prepared	Analyzed	CAS No.	Qual
			Limit	MDL	DF				
Individual Can Certification									
Analytical Method: TO-15									
Pace Analytical Services - Minneapolis									
1,1,2,2-Tetrachloroethane	<0.37	ug/m3	1.4	0.37	1		09/08/21 09:54	79-34-5	
Tetrachloroethene	<0.29	ug/m3	0.69	0.29	1		09/08/21 09:54	127-18-4	
Tetrahydrofuran	<0.18	ug/m3	0.60	0.18	1		09/08/21 09:54	109-99-9	
Toluene	<0.24	ug/m3	0.77	0.24	1		09/08/21 09:54	108-88-3	
1,2,4-Trichlorobenzene	<4.9	ug/m3	7.5	4.9	1		09/08/21 09:54	120-82-1	
1,1,1-Trichloroethane	<0.19	ug/m3	1.1	0.19	1		09/08/21 09:54	71-55-6	
1,1,2-Trichloroethane	<0.20	ug/m3	0.56	0.20	1		09/08/21 09:54	79-00-5	
Trichloroethene	<0.20	ug/m3	0.55	0.20	1		09/08/21 09:54	79-01-6	
Trichlorofluoromethane	<0.23	ug/m3	1.1	0.23	1		09/08/21 09:54	75-69-4	
1,1,2-Trichlorotrifluoroethane	<0.29	ug/m3	1.6	0.29	1		09/08/21 09:54	76-13-1	
1,2,4-Trimethylbenzene	<0.35	ug/m3	1.0	0.35	1		09/08/21 09:54	95-63-6	
1,3,5-Trimethylbenzene	<0.29	ug/m3	1.0	0.29	1		09/08/21 09:54	108-67-8	
Vinyl acetate	<0.21	ug/m3	0.72	0.21	1		09/08/21 09:54	108-05-4	
Vinyl chloride	<0.087	ug/m3	0.26	0.087	1		09/08/21 09:54	75-01-4	
m&p-Xylene	<0.64	ug/m3	1.8	0.64	1		09/08/21 09:54	179601-23-1	
o-Xylene	<0.27	ug/m3	0.88	0.27	1		09/08/21 09:54	95-47-6	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.

ANALYTICAL RESULTS

Project: CONBRACO

Pace Project No.: 92561801

Sample: IA-5 Cert# 2811 Lab ID: 92561801023 Collected: 09/14/21 16:15 Received: 09/16/21 13:15 Matrix: Air

Parameters	Results	Units	Report			Prepared	Analyzed	CAS No.	Qual
			Limit	MDL	DF				
Individual Can Certification									
Analytical Method: TO-15									
Pace Analytical Services - Minneapolis									
Acetone	<0.90	ug/m3	3.0	0.90	0.5		09/07/21 14:22	67-64-1	
Benzene	<0.057	ug/m3	0.16	0.057	0.5		09/07/21 14:22	71-43-2	
Benzyl chloride	0.68J	ug/m3	1.3	0.44	0.5		09/07/21 14:22	100-44-7	
Bromodichloromethane	<0.12	ug/m3	0.68	0.12	0.5		09/07/21 14:22	75-27-4	
Bromoform	<0.81	ug/m3	2.6	0.81	0.5		09/07/21 14:22	75-25-2	
Bromomethane	<0.075	ug/m3	0.39	0.075	0.5		09/07/21 14:22	74-83-9	
1,3-Butadiene	<0.060	ug/m3	0.22	0.060	0.5		09/07/21 14:22	106-99-0	
2-Butanone (MEK)	<0.23	ug/m3	1.5	0.23	0.5		09/07/21 14:22	78-93-3	
Carbon disulfide	<0.064	ug/m3	0.32	0.064	0.5		09/07/21 14:22	75-15-0	
Carbon tetrachloride	<0.14	ug/m3	0.64	0.14	0.5		09/07/21 14:22	56-23-5	
Chlorobenzene	<0.078	ug/m3	0.47	0.078	0.5		09/07/21 14:22	108-90-7	
Chloroethane	<0.11	ug/m3	0.27	0.11	0.5		09/07/21 14:22	75-00-3	
Chloroform	<0.092	ug/m3	0.25	0.092	0.5		09/07/21 14:22	67-66-3	
Chloromethane	<0.043	ug/m3	0.21	0.043	0.5		09/07/21 14:22	74-87-3	
Cyclohexane	<0.11	ug/m3	0.88	0.11	0.5		09/07/21 14:22	110-82-7	
Dibromochloromethane	<0.26	ug/m3	0.86	0.26	0.5		09/07/21 14:22	124-48-1	
1,2-Dibromoethane (EDB)	<0.15	ug/m3	0.39	0.15	0.5		09/07/21 14:22	106-93-4	
1,2-Dichlorobenzene	0.52J	ug/m3	1.5	0.20	0.5		09/07/21 14:22	95-50-1	
1,3-Dichlorobenzene	0.47J	ug/m3	1.5	0.25	0.5		09/07/21 14:22	541-73-1	
1,4-Dichlorobenzene	0.54J	ug/m3	1.5	0.44	0.5		09/07/21 14:22	106-46-7	
Dichlorodifluoromethane	<0.094	ug/m3	0.50	0.094	0.5		09/07/21 14:22	75-71-8	
1,1-Dichloroethane	<0.082	ug/m3	0.41	0.082	0.5		09/07/21 14:22	75-34-3	
1,2-Dichloroethane	<0.097	ug/m3	0.41	0.097	0.5		09/07/21 14:22	107-06-2	
1,1-Dichloroethene	<0.069	ug/m3	0.40	0.069	0.5		09/07/21 14:22	75-35-4	
cis-1,2-Dichloroethene	<0.098	ug/m3	0.40	0.098	0.5		09/07/21 14:22	156-59-2	
trans-1,2-Dichloroethene	<0.084	ug/m3	0.40	0.084	0.5		09/07/21 14:22	156-60-5	
1,2-Dichloropropane	<0.13	ug/m3	0.47	0.13	0.5		09/07/21 14:22	78-87-5	
cis-1,3-Dichloropropene	<0.13	ug/m3	1.2	0.13	0.5		09/07/21 14:22	10061-01-5	
trans-1,3-Dichloropropene	<0.27	ug/m3	1.2	0.27	0.5		09/07/21 14:22	10061-02-6	
Dichlorotetrafluoroethane	<0.10	ug/m3	0.71	0.10	0.5		09/07/21 14:22	76-14-2	
Ethanol	<0.30	ug/m3	0.96	0.30	0.5		09/07/21 14:22	64-17-5	
Ethyl acetate	<0.066	ug/m3	0.37	0.066	0.5		09/07/21 14:22	141-78-6	
Ethylbenzene	<0.15	ug/m3	0.44	0.15	0.5		09/07/21 14:22	100-41-4	
4-Ethyltoluene	0.42J	ug/m3	1.2	0.24	0.5		09/07/21 14:22	622-96-8	
n-Heptane	<0.090	ug/m3	0.42	0.090	0.5		09/07/21 14:22	142-82-5	
Hexachloro-1,3-butadiene	1.1J	ug/m3	2.7	0.62	0.5		09/07/21 14:22	87-68-3	
n-Hexane	<0.096	ug/m3	0.36	0.096	0.5		09/07/21 14:22	110-54-3	
2-Hexanone	<0.22	ug/m3	2.1	0.22	0.5		09/07/21 14:22	591-78-6	
Methylene Chloride	<0.30	ug/m3	1.8	0.30	0.5		09/07/21 14:22	75-09-2	
4-Methyl-2-pentanone (MIBK)	<0.16	ug/m3	2.1	0.16	0.5		09/07/21 14:22	108-10-1	
Methyl-tert-butyl ether	<0.063	ug/m3	1.8	0.063	0.5		09/07/21 14:22	1634-04-4	
Naphthalene	<1.1	ug/m3	1.3	1.1	0.5		09/07/21 14:22	91-20-3	
2-Propanol	<0.25	ug/m3	1.2	0.25	0.5		09/07/21 14:22	67-63-0	
Propylene	<0.065	ug/m3	0.44	0.065	0.5		09/07/21 14:22	115-07-1	
Styrene	<0.19	ug/m3	0.43	0.19	0.5		09/07/21 14:22	100-42-5	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.

ANALYTICAL RESULTS

Project: CONBRACO
Pace Project No.: 92561801

Sample: IA-5 Cert# 2811 **Lab ID: 92561801023** Collected: 09/14/21 16:15 Received: 09/16/21 13:15 Matrix: Air

Parameters	Results	Units	Report			Prepared	Analyzed	CAS No.	Qual
			Limit	MDL	DF				
Individual Can Certification									
Analytical Method: TO-15									
Pace Analytical Services - Minneapolis									
1,1,2,2-Tetrachloroethane	<0.19	ug/m3	0.70	0.19	0.5		09/07/21 14:22	79-34-5	
Tetrachloroethene	<0.15	ug/m3	0.34	0.15	0.5		09/07/21 14:22	127-18-4	
Tetrahydrofuran	<0.090	ug/m3	0.30	0.090	0.5		09/07/21 14:22	109-99-9	
Toluene	<0.12	ug/m3	0.38	0.12	0.5		09/07/21 14:22	108-88-3	
1,2,4-Trichlorobenzene	2.5J	ug/m3	3.8	2.4	0.5		09/07/21 14:22	120-82-1	
1,1,1-Trichloroethane	<0.093	ug/m3	0.56	0.093	0.5		09/07/21 14:22	71-55-6	
1,1,2-Trichloroethane	<0.098	ug/m3	0.28	0.098	0.5		09/07/21 14:22	79-00-5	
Trichloroethene	<0.098	ug/m3	0.27	0.098	0.5		09/07/21 14:22	79-01-6	
Trichlorofluoromethane	<0.12	ug/m3	0.57	0.12	0.5		09/07/21 14:22	75-69-4	
1,1,2-Trichlorotrifluoroethane	<0.14	ug/m3	0.78	0.14	0.5		09/07/21 14:22	76-13-1	
1,2,4-Trimethylbenzene	<0.18	ug/m3	0.50	0.18	0.5		09/07/21 14:22	95-63-6	
1,3,5-Trimethylbenzene	<0.14	ug/m3	0.50	0.14	0.5		09/07/21 14:22	108-67-8	
Vinyl acetate	<0.10	ug/m3	0.36	0.10	0.5		09/07/21 14:22	108-05-4	
Vinyl chloride	<0.043	ug/m3	0.13	0.043	0.5		09/07/21 14:22	75-01-4	
m&p-Xylene	<0.32	ug/m3	0.88	0.32	0.5		09/07/21 14:22	179601-23-1	
o-Xylene	<0.14	ug/m3	0.44	0.14	0.5		09/07/21 14:22	95-47-6	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.

ANALYTICAL RESULTS

Project: CONBRACO

Pace Project No.: 92561801

Sample: DUP-1 Cert#1195 Lab ID: 92561801024 Collected: 09/14/21 17:09 Received: 09/16/21 13:15 Matrix: Air

Parameters	Results	Units	Report			Prepared	Analyzed	CAS No.	Qual
			Limit	MDL	DF				
Individual Can Certification									
Analytical Method: TO-15									
Pace Analytical Services - Minneapolis									
Acetone	<0.90	ug/m3	3.0	0.90	0.5		09/08/21 03:33	67-64-1	
Benzene	<0.057	ug/m3	0.16	0.057	0.5		09/08/21 03:33	71-43-2	
Benzyl chloride	<0.44	ug/m3	1.3	0.44	0.5		09/08/21 03:33	100-44-7	
Bromodichloromethane	<0.12	ug/m3	0.68	0.12	0.5		09/08/21 03:33	75-27-4	
Bromoform	<0.81	ug/m3	2.6	0.81	0.5		09/08/21 03:33	75-25-2	
Bromomethane	<0.075	ug/m3	0.39	0.075	0.5		09/08/21 03:33	74-83-9	
1,3-Butadiene	<0.060	ug/m3	0.22	0.060	0.5		09/08/21 03:33	106-99-0	
2-Butanone (MEK)	<0.23	ug/m3	1.5	0.23	0.5		09/08/21 03:33	78-93-3	
Carbon disulfide	<0.064	ug/m3	0.32	0.064	0.5		09/08/21 03:33	75-15-0	
Carbon tetrachloride	<0.14	ug/m3	0.64	0.14	0.5		09/08/21 03:33	56-23-5	
Chlorobenzene	<0.078	ug/m3	0.47	0.078	0.5		09/08/21 03:33	108-90-7	
Chloroethane	<0.11	ug/m3	0.27	0.11	0.5		09/08/21 03:33	75-00-3	
Chloroform	<0.092	ug/m3	0.25	0.092	0.5		09/08/21 03:33	67-66-3	
Chloromethane	<0.043	ug/m3	0.21	0.043	0.5		09/08/21 03:33	74-87-3	
Cyclohexane	<0.11	ug/m3	0.88	0.11	0.5		09/08/21 03:33	110-82-7	
Dibromochloromethane	<0.26	ug/m3	0.86	0.26	0.5		09/08/21 03:33	124-48-1	
1,2-Dibromoethane (EDB)	<0.15	ug/m3	0.39	0.15	0.5		09/08/21 03:33	106-93-4	
1,2-Dichlorobenzene	<0.20	ug/m3	1.5	0.20	0.5		09/08/21 03:33	95-50-1	
1,3-Dichlorobenzene	<0.25	ug/m3	1.5	0.25	0.5		09/08/21 03:33	541-73-1	
1,4-Dichlorobenzene	<0.44	ug/m3	1.5	0.44	0.5		09/08/21 03:33	106-46-7	
Dichlorodifluoromethane	<0.094	ug/m3	0.50	0.094	0.5		09/08/21 03:33	75-71-8	
1,1-Dichloroethane	<0.082	ug/m3	0.41	0.082	0.5		09/08/21 03:33	75-34-3	
1,2-Dichloroethane	<0.097	ug/m3	0.41	0.097	0.5		09/08/21 03:33	107-06-2	
1,1-Dichloroethene	<0.069	ug/m3	0.40	0.069	0.5		09/08/21 03:33	75-35-4	
cis-1,2-Dichloroethene	<0.098	ug/m3	0.40	0.098	0.5		09/08/21 03:33	156-59-2	
trans-1,2-Dichloroethene	<0.084	ug/m3	0.40	0.084	0.5		09/08/21 03:33	156-60-5	
1,2-Dichloropropane	<0.13	ug/m3	0.47	0.13	0.5		09/08/21 03:33	78-87-5	
cis-1,3-Dichloropropene	<0.13	ug/m3	1.2	0.13	0.5		09/08/21 03:33	10061-01-5	
trans-1,3-Dichloropropene	<0.27	ug/m3	1.2	0.27	0.5		09/08/21 03:33	10061-02-6	
Dichlorotetrafluoroethane	<0.10	ug/m3	0.71	0.10	0.5		09/08/21 03:33	76-14-2	
Ethanol	<0.30	ug/m3	0.96	0.30	0.5		09/08/21 03:33	64-17-5	
Ethyl acetate	<0.066	ug/m3	0.37	0.066	0.5		09/08/21 03:33	141-78-6	
Ethylbenzene	<0.15	ug/m3	0.44	0.15	0.5		09/08/21 03:33	100-41-4	
4-Ethyltoluene	<0.24	ug/m3	1.2	0.24	0.5		09/08/21 03:33	622-96-8	
n-Heptane	<0.090	ug/m3	0.42	0.090	0.5		09/08/21 03:33	142-82-5	
Hexachloro-1,3-butadiene	<0.62	ug/m3	2.7	0.62	0.5		09/08/21 03:33	87-68-3	
n-Hexane	<0.096	ug/m3	0.36	0.096	0.5		09/08/21 03:33	110-54-3	
2-Hexanone	<0.22	ug/m3	2.1	0.22	0.5		09/08/21 03:33	591-78-6	
Methylene Chloride	<0.30	ug/m3	1.8	0.30	0.5		09/08/21 03:33	75-09-2	
4-Methyl-2-pentanone (MIBK)	<0.16	ug/m3	2.1	0.16	0.5		09/08/21 03:33	108-10-1	
Methyl-tert-butyl ether	<0.063	ug/m3	1.8	0.063	0.5		09/08/21 03:33	1634-04-4	
Naphthalene	<1.1	ug/m3	1.3	1.1	0.5		09/08/21 03:33	91-20-3	
2-Propanol	<0.25	ug/m3	1.2	0.25	0.5		09/08/21 03:33	67-63-0	
Propylene	<0.065	ug/m3	0.44	0.065	0.5		09/08/21 03:33	115-07-1	
Styrene	<0.19	ug/m3	0.43	0.19	0.5		09/08/21 03:33	100-42-5	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.

ANALYTICAL RESULTS

Project: CONBRACO
Pace Project No.: 92561801

Sample: DUP-1 Cert#1195 **Lab ID: 92561801024** Collected: 09/14/21 17:09 Received: 09/16/21 13:15 Matrix: Air

Parameters	Results	Units	Report			Prepared	Analyzed	CAS No.	Qual
			Limit	MDL	DF				
Individual Can Certification									
Analytical Method: TO-15									
Pace Analytical Services - Minneapolis									
1,1,2,2-Tetrachloroethane	<0.19	ug/m3	0.70	0.19	0.5		09/08/21 03:33	79-34-5	
Tetrachloroethene	<0.15	ug/m3	0.34	0.15	0.5		09/08/21 03:33	127-18-4	
Tetrahydrofuran	<0.090	ug/m3	0.30	0.090	0.5		09/08/21 03:33	109-99-9	
Toluene	<0.12	ug/m3	0.38	0.12	0.5		09/08/21 03:33	108-88-3	
1,2,4-Trichlorobenzene	<2.4	ug/m3	3.8	2.4	0.5		09/08/21 03:33	120-82-1	
1,1,1-Trichloroethane	<0.093	ug/m3	0.56	0.093	0.5		09/08/21 03:33	71-55-6	
1,1,2-Trichloroethane	<0.098	ug/m3	0.28	0.098	0.5		09/08/21 03:33	79-00-5	
Trichloroethene	<0.098	ug/m3	0.27	0.098	0.5		09/08/21 03:33	79-01-6	
Trichlorofluoromethane	<0.12	ug/m3	0.57	0.12	0.5		09/08/21 03:33	75-69-4	
1,1,2-Trichlorotrifluoroethane	<0.14	ug/m3	0.78	0.14	0.5		09/08/21 03:33	76-13-1	
1,2,4-Trimethylbenzene	<0.18	ug/m3	0.50	0.18	0.5		09/08/21 03:33	95-63-6	
1,3,5-Trimethylbenzene	<0.14	ug/m3	0.50	0.14	0.5		09/08/21 03:33	108-67-8	
Vinyl acetate	<0.10	ug/m3	0.36	0.10	0.5		09/08/21 03:33	108-05-4	
Vinyl chloride	<0.043	ug/m3	0.13	0.043	0.5		09/08/21 03:33	75-01-4	
m&p-Xylene	<0.32	ug/m3	0.88	0.32	0.5		09/08/21 03:33	179601-23-1	
o-Xylene	<0.14	ug/m3	0.44	0.14	0.5		09/08/21 03:33	95-47-6	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.

ANALYTICAL RESULTS

Project: CONBRACO

Pace Project No.: 92561801

Sample: IA-13 Cert# 2803 **Lab ID: 92561801025** Collected: 09/14/21 17:09 Received: 09/16/21 13:15 Matrix: Air

Parameters	Results	Units	Report			Prepared	Analyzed	CAS No.	Qual
			Limit	MDL	DF				
Individual Can Certification									
Analytical Method: TO-15									
Pace Analytical Services - Minneapolis									
Acetone	<0.90	ug/m3	3.0	0.90	0.5		09/08/21 04:14	67-64-1	
Benzene	<0.057	ug/m3	0.16	0.057	0.5		09/08/21 04:14	71-43-2	
Benzyl chloride	<0.44	ug/m3	1.3	0.44	0.5		09/08/21 04:14	100-44-7	
Bromodichloromethane	<0.12	ug/m3	0.68	0.12	0.5		09/08/21 04:14	75-27-4	
Bromoform	<0.81	ug/m3	2.6	0.81	0.5		09/08/21 04:14	75-25-2	
Bromomethane	<0.075	ug/m3	0.39	0.075	0.5		09/08/21 04:14	74-83-9	
1,3-Butadiene	<0.060	ug/m3	0.22	0.060	0.5		09/08/21 04:14	106-99-0	
2-Butanone (MEK)	<0.23	ug/m3	1.5	0.23	0.5		09/08/21 04:14	78-93-3	
Carbon disulfide	<0.064	ug/m3	0.32	0.064	0.5		09/08/21 04:14	75-15-0	
Carbon tetrachloride	<0.14	ug/m3	0.64	0.14	0.5		09/08/21 04:14	56-23-5	
Chlorobenzene	<0.078	ug/m3	0.47	0.078	0.5		09/08/21 04:14	108-90-7	
Chloroethane	<0.11	ug/m3	0.27	0.11	0.5		09/08/21 04:14	75-00-3	
Chloroform	<0.092	ug/m3	0.25	0.092	0.5		09/08/21 04:14	67-66-3	
Chloromethane	<0.043	ug/m3	0.21	0.043	0.5		09/08/21 04:14	74-87-3	
Cyclohexane	<0.11	ug/m3	0.88	0.11	0.5		09/08/21 04:14	110-82-7	
Dibromochloromethane	<0.26	ug/m3	0.86	0.26	0.5		09/08/21 04:14	124-48-1	
1,2-Dibromoethane (EDB)	<0.15	ug/m3	0.39	0.15	0.5		09/08/21 04:14	106-93-4	
1,2-Dichlorobenzene	<0.20	ug/m3	1.5	0.20	0.5		09/08/21 04:14	95-50-1	
1,3-Dichlorobenzene	<0.25	ug/m3	1.5	0.25	0.5		09/08/21 04:14	541-73-1	
1,4-Dichlorobenzene	<0.44	ug/m3	1.5	0.44	0.5		09/08/21 04:14	106-46-7	
Dichlorodifluoromethane	<0.094	ug/m3	0.50	0.094	0.5		09/08/21 04:14	75-71-8	
1,1-Dichloroethane	<0.082	ug/m3	0.41	0.082	0.5		09/08/21 04:14	75-34-3	
1,2-Dichloroethane	<0.097	ug/m3	0.41	0.097	0.5		09/08/21 04:14	107-06-2	
1,1-Dichloroethene	<0.069	ug/m3	0.40	0.069	0.5		09/08/21 04:14	75-35-4	
cis-1,2-Dichloroethene	<0.098	ug/m3	0.40	0.098	0.5		09/08/21 04:14	156-59-2	
trans-1,2-Dichloroethene	<0.084	ug/m3	0.40	0.084	0.5		09/08/21 04:14	156-60-5	
1,2-Dichloropropane	<0.13	ug/m3	0.47	0.13	0.5		09/08/21 04:14	78-87-5	
cis-1,3-Dichloropropene	<0.13	ug/m3	1.2	0.13	0.5		09/08/21 04:14	10061-01-5	
trans-1,3-Dichloropropene	<0.27	ug/m3	1.2	0.27	0.5		09/08/21 04:14	10061-02-6	
Dichlorotetrafluoroethane	<0.10	ug/m3	0.71	0.10	0.5		09/08/21 04:14	76-14-2	
Ethanol	<0.30	ug/m3	0.96	0.30	0.5		09/08/21 04:14	64-17-5	
Ethyl acetate	<0.066	ug/m3	0.37	0.066	0.5		09/08/21 04:14	141-78-6	
Ethylbenzene	<0.15	ug/m3	0.44	0.15	0.5		09/08/21 04:14	100-41-4	
4-Ethyltoluene	<0.24	ug/m3	1.2	0.24	0.5		09/08/21 04:14	622-96-8	
n-Heptane	<0.090	ug/m3	0.42	0.090	0.5		09/08/21 04:14	142-82-5	
Hexachloro-1,3-butadiene	<0.62	ug/m3	2.7	0.62	0.5		09/08/21 04:14	87-68-3	
n-Hexane	<0.096	ug/m3	0.36	0.096	0.5		09/08/21 04:14	110-54-3	
2-Hexanone	<0.22	ug/m3	2.1	0.22	0.5		09/08/21 04:14	591-78-6	
Methylene Chloride	<0.30	ug/m3	1.8	0.30	0.5		09/08/21 04:14	75-09-2	
4-Methyl-2-pentanone (MIBK)	<0.16	ug/m3	2.1	0.16	0.5		09/08/21 04:14	108-10-1	
Methyl-tert-butyl ether	<0.063	ug/m3	1.8	0.063	0.5		09/08/21 04:14	1634-04-4	
Naphthalene	<1.1	ug/m3	1.3	1.1	0.5		09/08/21 04:14	91-20-3	
2-Propanol	<0.25	ug/m3	1.2	0.25	0.5		09/08/21 04:14	67-63-0	
Propylene	<0.065	ug/m3	0.44	0.065	0.5		09/08/21 04:14	115-07-1	
Styrene	<0.19	ug/m3	0.43	0.19	0.5		09/08/21 04:14	100-42-5	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.

ANALYTICAL RESULTS

Project: CONBRACO
Pace Project No.: 92561801

Sample: IA-13 Cert# 2803 **Lab ID: 92561801025** Collected: 09/14/21 17:09 Received: 09/16/21 13:15 Matrix: Air

Parameters	Results	Units	Report			Prepared	Analyzed	CAS No.	Qual
			Limit	MDL	DF				
Individual Can Certification									
Analytical Method: TO-15									
Pace Analytical Services - Minneapolis									
1,1,2,2-Tetrachloroethane	<0.19	ug/m3	0.70	0.19	0.5		09/08/21 04:14	79-34-5	
Tetrachloroethene	<0.15	ug/m3	0.34	0.15	0.5		09/08/21 04:14	127-18-4	
Tetrahydrofuran	<0.090	ug/m3	0.30	0.090	0.5		09/08/21 04:14	109-99-9	
Toluene	<0.12	ug/m3	0.38	0.12	0.5		09/08/21 04:14	108-88-3	
1,2,4-Trichlorobenzene	<2.4	ug/m3	3.8	2.4	0.5		09/08/21 04:14	120-82-1	
1,1,1-Trichloroethane	<0.093	ug/m3	0.56	0.093	0.5		09/08/21 04:14	71-55-6	
1,1,2-Trichloroethane	<0.098	ug/m3	0.28	0.098	0.5		09/08/21 04:14	79-00-5	
Trichloroethene	<0.098	ug/m3	0.27	0.098	0.5		09/08/21 04:14	79-01-6	
Trichlorofluoromethane	<0.12	ug/m3	0.57	0.12	0.5		09/08/21 04:14	75-69-4	
1,1,2-Trichlorotrifluoroethane	<0.14	ug/m3	0.78	0.14	0.5		09/08/21 04:14	76-13-1	
1,2,4-Trimethylbenzene	<0.18	ug/m3	0.50	0.18	0.5		09/08/21 04:14	95-63-6	
1,3,5-Trimethylbenzene	<0.14	ug/m3	0.50	0.14	0.5		09/08/21 04:14	108-67-8	
Vinyl acetate	<0.10	ug/m3	0.36	0.10	0.5		09/08/21 04:14	108-05-4	
Vinyl chloride	<0.043	ug/m3	0.13	0.043	0.5		09/08/21 04:14	75-01-4	
m&p-Xylene	<0.32	ug/m3	0.88	0.32	0.5		09/08/21 04:14	179601-23-1	
o-Xylene	<0.14	ug/m3	0.44	0.14	0.5		09/08/21 04:14	95-47-6	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.

ANALYTICAL RESULTS

Project: CONBRACO
Pace Project No.: 92561801

Sample: IA-7 Cert# 3629 Lab ID: 92561801026 Collected: 09/14/21 16:56 Received: 09/16/21 13:15 Matrix: Air

Parameters	Results	Units	Report			Prepared	Analyzed	CAS No.	Qual
			Limit	MDL	DF				
Individual Can Certification									
Analytical Method: TO-15									
Pace Analytical Services - Minneapolis									
Acetone	<0.90	ug/m3	3.0	0.90	0.5		09/08/21 05:35	67-64-1	
Benzene	<0.057	ug/m3	0.16	0.057	0.5		09/08/21 05:35	71-43-2	
Benzyl chloride	<0.44	ug/m3	1.3	0.44	0.5		09/08/21 05:35	100-44-7	
Bromodichloromethane	<0.12	ug/m3	0.68	0.12	0.5		09/08/21 05:35	75-27-4	
Bromoform	<0.81	ug/m3	2.6	0.81	0.5		09/08/21 05:35	75-25-2	
Bromomethane	<0.075	ug/m3	0.39	0.075	0.5		09/08/21 05:35	74-83-9	
1,3-Butadiene	<0.060	ug/m3	0.22	0.060	0.5		09/08/21 05:35	106-99-0	
2-Butanone (MEK)	<0.23	ug/m3	1.5	0.23	0.5		09/08/21 05:35	78-93-3	
Carbon disulfide	<0.064	ug/m3	0.32	0.064	0.5		09/08/21 05:35	75-15-0	
Carbon tetrachloride	<0.14	ug/m3	0.64	0.14	0.5		09/08/21 05:35	56-23-5	
Chlorobenzene	<0.078	ug/m3	0.47	0.078	0.5		09/08/21 05:35	108-90-7	
Chloroethane	<0.11	ug/m3	0.27	0.11	0.5		09/08/21 05:35	75-00-3	
Chloroform	<0.092	ug/m3	0.25	0.092	0.5		09/08/21 05:35	67-66-3	
Chloromethane	<0.043	ug/m3	0.21	0.043	0.5		09/08/21 05:35	74-87-3	
Cyclohexane	<0.11	ug/m3	0.88	0.11	0.5		09/08/21 05:35	110-82-7	
Dibromochloromethane	<0.26	ug/m3	0.86	0.26	0.5		09/08/21 05:35	124-48-1	
1,2-Dibromoethane (EDB)	<0.15	ug/m3	0.39	0.15	0.5		09/08/21 05:35	106-93-4	
1,2-Dichlorobenzene	<0.20	ug/m3	1.5	0.20	0.5		09/08/21 05:35	95-50-1	
1,3-Dichlorobenzene	<0.25	ug/m3	1.5	0.25	0.5		09/08/21 05:35	541-73-1	
1,4-Dichlorobenzene	<0.44	ug/m3	1.5	0.44	0.5		09/08/21 05:35	106-46-7	
Dichlorodifluoromethane	<0.094	ug/m3	0.50	0.094	0.5		09/08/21 05:35	75-71-8	
1,1-Dichloroethane	<0.082	ug/m3	0.41	0.082	0.5		09/08/21 05:35	75-34-3	
1,2-Dichloroethane	<0.097	ug/m3	0.41	0.097	0.5		09/08/21 05:35	107-06-2	
1,1-Dichloroethene	<0.069	ug/m3	0.40	0.069	0.5		09/08/21 05:35	75-35-4	
cis-1,2-Dichloroethene	<0.098	ug/m3	0.40	0.098	0.5		09/08/21 05:35	156-59-2	
trans-1,2-Dichloroethene	<0.084	ug/m3	0.40	0.084	0.5		09/08/21 05:35	156-60-5	
1,2-Dichloropropane	<0.13	ug/m3	0.47	0.13	0.5		09/08/21 05:35	78-87-5	
cis-1,3-Dichloropropene	<0.13	ug/m3	1.2	0.13	0.5		09/08/21 05:35	10061-01-5	
trans-1,3-Dichloropropene	<0.27	ug/m3	1.2	0.27	0.5		09/08/21 05:35	10061-02-6	
Dichlorotetrafluoroethane	<0.10	ug/m3	0.71	0.10	0.5		09/08/21 05:35	76-14-2	
Ethanol	<0.30	ug/m3	0.96	0.30	0.5		09/08/21 05:35	64-17-5	
Ethyl acetate	<0.066	ug/m3	0.37	0.066	0.5		09/08/21 05:35	141-78-6	
Ethylbenzene	<0.15	ug/m3	0.44	0.15	0.5		09/08/21 05:35	100-41-4	
4-Ethyltoluene	<0.24	ug/m3	1.2	0.24	0.5		09/08/21 05:35	622-96-8	
n-Heptane	<0.090	ug/m3	0.42	0.090	0.5		09/08/21 05:35	142-82-5	
Hexachloro-1,3-butadiene	<0.62	ug/m3	2.7	0.62	0.5		09/08/21 05:35	87-68-3	
n-Hexane	<0.096	ug/m3	0.36	0.096	0.5		09/08/21 05:35	110-54-3	
2-Hexanone	<0.22	ug/m3	2.1	0.22	0.5		09/08/21 05:35	591-78-6	
Methylene Chloride	<0.30	ug/m3	1.8	0.30	0.5		09/08/21 05:35	75-09-2	
4-Methyl-2-pentanone (MIBK)	<0.16	ug/m3	2.1	0.16	0.5		09/08/21 05:35	108-10-1	
Methyl-tert-butyl ether	<0.063	ug/m3	1.8	0.063	0.5		09/08/21 05:35	1634-04-4	
Naphthalene	<1.1	ug/m3	1.3	1.1	0.5		09/08/21 05:35	91-20-3	
2-Propanol	<0.25	ug/m3	1.2	0.25	0.5		09/08/21 05:35	67-63-0	
Propylene	<0.065	ug/m3	0.44	0.065	0.5		09/08/21 05:35	115-07-1	
Styrene	<0.19	ug/m3	0.43	0.19	0.5		09/08/21 05:35	100-42-5	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.

ANALYTICAL RESULTS

Project: CONBRACO
Pace Project No.: 92561801

Sample: IA-7 Cert# 3629 **Lab ID: 92561801026** Collected: 09/14/21 16:56 Received: 09/16/21 13:15 Matrix: Air

Parameters	Results	Units	Report			Prepared	Analyzed	CAS No.	Qual
			Limit	MDL	DF				
Individual Can Certification									
Analytical Method: TO-15									
Pace Analytical Services - Minneapolis									
1,1,2,2-Tetrachloroethane	<0.19	ug/m3	0.70	0.19	0.5		09/08/21 05:35	79-34-5	
Tetrachloroethene	<0.15	ug/m3	0.34	0.15	0.5		09/08/21 05:35	127-18-4	
Tetrahydrofuran	<0.090	ug/m3	0.30	0.090	0.5		09/08/21 05:35	109-99-9	
Toluene	<0.12	ug/m3	0.38	0.12	0.5		09/08/21 05:35	108-88-3	
1,2,4-Trichlorobenzene	<2.4	ug/m3	3.8	2.4	0.5		09/08/21 05:35	120-82-1	
1,1,1-Trichloroethane	<0.093	ug/m3	0.56	0.093	0.5		09/08/21 05:35	71-55-6	
1,1,2-Trichloroethane	<0.098	ug/m3	0.28	0.098	0.5		09/08/21 05:35	79-00-5	
Trichloroethene	<0.098	ug/m3	0.27	0.098	0.5		09/08/21 05:35	79-01-6	
Trichlorofluoromethane	<0.12	ug/m3	0.57	0.12	0.5		09/08/21 05:35	75-69-4	
1,1,2-Trichlorotrifluoroethane	<0.14	ug/m3	0.78	0.14	0.5		09/08/21 05:35	76-13-1	
1,2,4-Trimethylbenzene	<0.18	ug/m3	0.50	0.18	0.5		09/08/21 05:35	95-63-6	
1,3,5-Trimethylbenzene	<0.14	ug/m3	0.50	0.14	0.5		09/08/21 05:35	108-67-8	
Vinyl acetate	<0.10	ug/m3	0.36	0.10	0.5		09/08/21 05:35	108-05-4	
Vinyl chloride	<0.043	ug/m3	0.13	0.043	0.5		09/08/21 05:35	75-01-4	
m&p-Xylene	<0.32	ug/m3	0.88	0.32	0.5		09/08/21 05:35	179601-23-1	
o-Xylene	<0.14	ug/m3	0.44	0.14	0.5		09/08/21 05:35	95-47-6	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.

ANALYTICAL RESULTS

Project: CONBRACO

Pace Project No.: 92561801

Sample: IA-3 Cert # 0629 **Lab ID: 92561801027** Collected: 09/14/21 16:48 Received: 09/16/21 13:15 Matrix: Air

Parameters	Results	Units	Report			Prepared	Analyzed	CAS No.	Qual
			Limit	MDL	DF				
Individual Can Certification									
Analytical Method: TO-15									
Pace Analytical Services - Minneapolis									
Acetone	<9.0	ug/m3	30.2	9.0	5		09/03/21 10:53	67-64-1	
Benzene	<0.57	ug/m3	1.6	0.57	5		09/03/21 10:53	71-43-2	
Benzyl chloride	<4.4	ug/m3	13.2	4.4	5		09/03/21 10:53	100-44-7	
Bromodichloromethane	<1.2	ug/m3	6.8	1.2	5		09/03/21 10:53	75-27-4	
Bromoform	<8.1	ug/m3	26.2	8.1	5		09/03/21 10:53	75-25-2	
Bromomethane	<0.75	ug/m3	3.9	0.75	5		09/03/21 10:53	74-83-9	
1,3-Butadiene	<0.60	ug/m3	2.2	0.60	5		09/03/21 10:53	106-99-0	
2-Butanone (MEK)	<2.3	ug/m3	15.0	2.3	5		09/03/21 10:53	78-93-3	
Carbon disulfide	<0.64	ug/m3	3.2	0.64	5		09/03/21 10:53	75-15-0	
Carbon tetrachloride	<1.4	ug/m3	6.4	1.4	5		09/03/21 10:53	56-23-5	
Chlorobenzene	<0.78	ug/m3	4.7	0.78	5		09/03/21 10:53	108-90-7	
Chloroethane	<1.1	ug/m3	2.7	1.1	5		09/03/21 10:53	75-00-3	
Chloroform	<0.92	ug/m3	2.5	0.92	5		09/03/21 10:53	67-66-3	
Chloromethane	<0.43	ug/m3	2.1	0.43	5		09/03/21 10:53	74-87-3	
Cyclohexane	<1.1	ug/m3	8.8	1.1	5		09/03/21 10:53	110-82-7	
Dibromochloromethane	<2.6	ug/m3	8.6	2.6	5		09/03/21 10:53	124-48-1	
1,2-Dibromoethane (EDB)	<1.5	ug/m3	3.9	1.5	5		09/03/21 10:53	106-93-4	
1,2-Dichlorobenzene	<2.0	ug/m3	15.3	2.0	5		09/03/21 10:53	95-50-1	
1,3-Dichlorobenzene	<2.5	ug/m3	15.3	2.5	5		09/03/21 10:53	541-73-1	
1,4-Dichlorobenzene	<4.4	ug/m3	15.3	4.4	5		09/03/21 10:53	106-46-7	
Dichlorodifluoromethane	<0.94	ug/m3	5.0	0.94	5		09/03/21 10:53	75-71-8	
1,1-Dichloroethane	<0.82	ug/m3	4.1	0.82	5		09/03/21 10:53	75-34-3	
1,2-Dichloroethane	<0.97	ug/m3	4.1	0.97	5		09/03/21 10:53	107-06-2	
1,1-Dichloroethene	<0.69	ug/m3	4.0	0.69	5		09/03/21 10:53	75-35-4	
cis-1,2-Dichloroethene	<0.98	ug/m3	4.0	0.98	5		09/03/21 10:53	156-59-2	
trans-1,2-Dichloroethene	<0.84	ug/m3	4.0	0.84	5		09/03/21 10:53	156-60-5	
1,2-Dichloropropane	<1.3	ug/m3	4.7	1.3	5		09/03/21 10:53	78-87-5	
cis-1,3-Dichloropropene	<1.3	ug/m3	11.6	1.3	5		09/03/21 10:53	10061-01-5	
trans-1,3-Dichloropropene	<2.7	ug/m3	11.6	2.7	5		09/03/21 10:53	10061-02-6	
Dichlorotetrafluoroethane	<1.0	ug/m3	7.1	1.0	5		09/03/21 10:53	76-14-2	
Ethanol	<3.0	ug/m3	9.6	3.0	5		09/03/21 10:53	64-17-5	
Ethyl acetate	<0.66	ug/m3	3.7	0.66	5		09/03/21 10:53	141-78-6	
Ethylbenzene	<1.5	ug/m3	4.4	1.5	5		09/03/21 10:53	100-41-4	
4-Ethyltoluene	<2.4	ug/m3	12.5	2.4	5		09/03/21 10:53	622-96-8	
n-Heptane	<0.90	ug/m3	4.2	0.90	5		09/03/21 10:53	142-82-5	
Hexachloro-1,3-butadiene	<6.2	ug/m3	27.1	6.2	5		09/03/21 10:53	87-68-3	
n-Hexane	<0.96	ug/m3	3.6	0.96	5		09/03/21 10:53	110-54-3	
2-Hexanone	<2.2	ug/m3	20.8	2.2	5		09/03/21 10:53	591-78-6	
Methylene Chloride	<3.0	ug/m3	17.6	3.0	5		09/03/21 10:53	75-09-2	
4-Methyl-2-pentanone (MIBK)	<1.6	ug/m3	20.8	1.6	5		09/03/21 10:53	108-10-1	
Methyl-tert-butyl ether	<0.63	ug/m3	18.3	0.63	5		09/03/21 10:53	1634-04-4	
Naphthalene	<10.8	ug/m3	13.3	10.8	5		09/03/21 10:53	91-20-3	
2-Propanol	<2.5	ug/m3	12.5	2.5	5		09/03/21 10:53	67-63-0	
Propylene	<0.65	ug/m3	4.4	0.65	5		09/03/21 10:53	115-07-1	
Styrene	<1.9	ug/m3	4.3	1.9	5		09/03/21 10:53	100-42-5	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.

ANALYTICAL RESULTS

Project: CONBRACO
Pace Project No.: 92561801

Sample: IA-3 Cert # 0629 **Lab ID: 92561801027** Collected: 09/14/21 16:48 Received: 09/16/21 13:15 Matrix: Air

Parameters	Results	Units	Report			Prepared	Analyzed	CAS No.	Qual
			Limit	MDL	DF				
Individual Can Certification									
Analytical Method: TO-15									
Pace Analytical Services - Minneapolis									
1,1,2,2-Tetrachloroethane	<1.9	ug/m3	7.0	1.9	5		09/03/21 10:53	79-34-5	
Tetrachloroethene	<1.5	ug/m3	3.4	1.5	5		09/03/21 10:53	127-18-4	
Tetrahydrofuran	<0.90	ug/m3	3.0	0.90	5		09/03/21 10:53	109-99-9	
Toluene	<1.2	ug/m3	3.8	1.2	5		09/03/21 10:53	108-88-3	
1,2,4-Trichlorobenzene	<24.4	ug/m3	37.7	24.4	5		09/03/21 10:53	120-82-1	
1,1,1-Trichloroethane	<0.93	ug/m3	5.6	0.93	5		09/03/21 10:53	71-55-6	
1,1,2-Trichloroethane	<0.98	ug/m3	2.8	0.98	5		09/03/21 10:53	79-00-5	
Trichloroethene	<0.98	ug/m3	2.7	0.98	5		09/03/21 10:53	79-01-6	
Trichlorofluoromethane	<1.2	ug/m3	5.7	1.2	5		09/03/21 10:53	75-69-4	
1,1,2-Trichlorotrifluoroethane	<1.4	ug/m3	7.8	1.4	5		09/03/21 10:53	76-13-1	
1,2,4-Trimethylbenzene	2.2J	ug/m3	5.0	1.8	5		09/03/21 10:53	95-63-6	
1,3,5-Trimethylbenzene	<1.4	ug/m3	5.0	1.4	5		09/03/21 10:53	108-67-8	
Vinyl acetate	<1.0	ug/m3	3.6	1.0	5		09/03/21 10:53	108-05-4	
Vinyl chloride	<0.43	ug/m3	1.3	0.43	5		09/03/21 10:53	75-01-4	
m&p-Xylene	<3.2	ug/m3	8.8	3.2	5		09/03/21 10:53	179601-23-1	
o-Xylene	<1.4	ug/m3	4.4	1.4	5		09/03/21 10:53	95-47-6	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.

ANALYTICAL RESULTS

Project: CONBRACO

Pace Project No.: 92561801

Sample: IA-10 Cert# 3475 Lab ID: 92561801028 Collected: 09/14/21 17:35 Received: 09/16/21 13:15 Matrix: Air

Parameters	Results	Units	Report			Prepared	Analyzed	CAS No.	Qual
			Limit	MDL	DF				
Individual Can Certification									
Analytical Method: TO-15									
Pace Analytical Services - Minneapolis									
Acetone	<0.90	ug/m3	3.0	0.90	0.5		09/07/21 12:59	67-64-1	
Benzene	<0.057	ug/m3	0.16	0.057	0.5		09/07/21 12:59	71-43-2	
Benzyl chloride	<0.44	ug/m3	1.3	0.44	0.5		09/07/21 12:59	100-44-7	
Bromodichloromethane	<0.12	ug/m3	0.68	0.12	0.5		09/07/21 12:59	75-27-4	
Bromoform	<0.81	ug/m3	2.6	0.81	0.5		09/07/21 12:59	75-25-2	
Bromomethane	<0.075	ug/m3	0.39	0.075	0.5		09/07/21 12:59	74-83-9	
1,3-Butadiene	<0.060	ug/m3	0.22	0.060	0.5		09/07/21 12:59	106-99-0	
2-Butanone (MEK)	<0.23	ug/m3	1.5	0.23	0.5		09/07/21 12:59	78-93-3	
Carbon disulfide	<0.064	ug/m3	0.32	0.064	0.5		09/07/21 12:59	75-15-0	
Carbon tetrachloride	<0.14	ug/m3	0.64	0.14	0.5		09/07/21 12:59	56-23-5	
Chlorobenzene	<0.078	ug/m3	0.47	0.078	0.5		09/07/21 12:59	108-90-7	
Chloroethane	<0.11	ug/m3	0.27	0.11	0.5		09/07/21 12:59	75-00-3	
Chloroform	<0.092	ug/m3	0.25	0.092	0.5		09/07/21 12:59	67-66-3	
Chloromethane	<0.043	ug/m3	0.21	0.043	0.5		09/07/21 12:59	74-87-3	
Cyclohexane	<0.11	ug/m3	0.88	0.11	0.5		09/07/21 12:59	110-82-7	
Dibromochloromethane	<0.26	ug/m3	0.86	0.26	0.5		09/07/21 12:59	124-48-1	
1,2-Dibromoethane (EDB)	<0.15	ug/m3	0.39	0.15	0.5		09/07/21 12:59	106-93-4	
1,2-Dichlorobenzene	<0.20	ug/m3	1.5	0.20	0.5		09/07/21 12:59	95-50-1	
1,3-Dichlorobenzene	<0.25	ug/m3	1.5	0.25	0.5		09/07/21 12:59	541-73-1	
1,4-Dichlorobenzene	<0.44	ug/m3	1.5	0.44	0.5		09/07/21 12:59	106-46-7	
Dichlorodifluoromethane	<0.094	ug/m3	0.50	0.094	0.5		09/07/21 12:59	75-71-8	
1,1-Dichloroethane	<0.082	ug/m3	0.41	0.082	0.5		09/07/21 12:59	75-34-3	
1,2-Dichloroethane	<0.097	ug/m3	0.41	0.097	0.5		09/07/21 12:59	107-06-2	
1,1-Dichloroethene	<0.069	ug/m3	0.40	0.069	0.5		09/07/21 12:59	75-35-4	
cis-1,2-Dichloroethene	<0.098	ug/m3	0.40	0.098	0.5		09/07/21 12:59	156-59-2	
trans-1,2-Dichloroethene	<0.084	ug/m3	0.40	0.084	0.5		09/07/21 12:59	156-60-5	
1,2-Dichloropropane	<0.13	ug/m3	0.47	0.13	0.5		09/07/21 12:59	78-87-5	
cis-1,3-Dichloropropene	<0.13	ug/m3	1.2	0.13	0.5		09/07/21 12:59	10061-01-5	
trans-1,3-Dichloropropene	<0.27	ug/m3	1.2	0.27	0.5		09/07/21 12:59	10061-02-6	
Dichlorotetrafluoroethane	<0.10	ug/m3	0.71	0.10	0.5		09/07/21 12:59	76-14-2	
Ethanol	<0.30	ug/m3	0.96	0.30	0.5		09/07/21 12:59	64-17-5	
Ethyl acetate	<0.066	ug/m3	0.37	0.066	0.5		09/07/21 12:59	141-78-6	
Ethylbenzene	<0.15	ug/m3	0.44	0.15	0.5		09/07/21 12:59	100-41-4	
4-Ethyltoluene	<0.24	ug/m3	1.2	0.24	0.5		09/07/21 12:59	622-96-8	
n-Heptane	<0.090	ug/m3	0.42	0.090	0.5		09/07/21 12:59	142-82-5	
Hexachloro-1,3-butadiene	<0.62	ug/m3	2.7	0.62	0.5		09/07/21 12:59	87-68-3	
n-Hexane	<0.096	ug/m3	0.36	0.096	0.5		09/07/21 12:59	110-54-3	
2-Hexanone	<0.22	ug/m3	2.1	0.22	0.5		09/07/21 12:59	591-78-6	
Methylene Chloride	<0.30	ug/m3	1.8	0.30	0.5		09/07/21 12:59	75-09-2	
4-Methyl-2-pentanone (MIBK)	<0.16	ug/m3	2.1	0.16	0.5		09/07/21 12:59	108-10-1	
Methyl-tert-butyl ether	<0.063	ug/m3	1.8	0.063	0.5		09/07/21 12:59	1634-04-4	
Naphthalene	<1.1	ug/m3	1.3	1.1	0.5		09/07/21 12:59	91-20-3	
2-Propanol	<0.25	ug/m3	1.2	0.25	0.5		09/07/21 12:59	67-63-0	
Propylene	<0.065	ug/m3	0.44	0.065	0.5		09/07/21 12:59	115-07-1	
Styrene	<0.19	ug/m3	0.43	0.19	0.5		09/07/21 12:59	100-42-5	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.

ANALYTICAL RESULTS

Project: CONBRACO
Pace Project No.: 92561801

Sample: IA-10 Cert# 3475 **Lab ID: 92561801028** Collected: 09/14/21 17:35 Received: 09/16/21 13:15 Matrix: Air

Parameters	Results	Units	Report			Prepared	Analyzed	CAS No.	Qual
			Limit	MDL	DF				
Individual Can Certification									
Analytical Method: TO-15									
Pace Analytical Services - Minneapolis									
1,1,2,2-Tetrachloroethane	<0.19	ug/m3	0.70	0.19	0.5		09/07/21 12:59	79-34-5	
Tetrachloroethene	<0.15	ug/m3	0.34	0.15	0.5		09/07/21 12:59	127-18-4	
Tetrahydrofuran	<0.090	ug/m3	0.30	0.090	0.5		09/07/21 12:59	109-99-9	
Toluene	<0.12	ug/m3	0.38	0.12	0.5		09/07/21 12:59	108-88-3	
1,2,4-Trichlorobenzene	<2.4	ug/m3	3.8	2.4	0.5		09/07/21 12:59	120-82-1	
1,1,1-Trichloroethane	<0.093	ug/m3	0.56	0.093	0.5		09/07/21 12:59	71-55-6	
1,1,2-Trichloroethane	<0.098	ug/m3	0.28	0.098	0.5		09/07/21 12:59	79-00-5	
Trichloroethene	<0.098	ug/m3	0.27	0.098	0.5		09/07/21 12:59	79-01-6	
Trichlorofluoromethane	<0.12	ug/m3	0.57	0.12	0.5		09/07/21 12:59	75-69-4	
1,1,2-Trichlorotrifluoroethane	<0.14	ug/m3	0.78	0.14	0.5		09/07/21 12:59	76-13-1	
1,2,4-Trimethylbenzene	<0.18	ug/m3	0.50	0.18	0.5		09/07/21 12:59	95-63-6	
1,3,5-Trimethylbenzene	<0.14	ug/m3	0.50	0.14	0.5		09/07/21 12:59	108-67-8	
Vinyl acetate	<0.10	ug/m3	0.36	0.10	0.5		09/07/21 12:59	108-05-4	
Vinyl chloride	<0.043	ug/m3	0.13	0.043	0.5		09/07/21 12:59	75-01-4	
m&p-Xylene	<0.32	ug/m3	0.88	0.32	0.5		09/07/21 12:59	179601-23-1	
o-Xylene	<0.14	ug/m3	0.44	0.14	0.5		09/07/21 12:59	95-47-6	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.

ANALYTICAL RESULTS

Project: CONBRACO

Pace Project No.: 92561801

Sample: IA-11 Cert# 3408 Lab ID: 92561801029 Collected: 09/14/21 18:45 Received: 09/16/21 13:15 Matrix: Air

Parameters	Results	Units	Report			Prepared	Analyzed	CAS No.	Qual
			Limit	MDL	DF				
Individual Can Certification									
Analytical Method: TO-15									
Pace Analytical Services - Minneapolis									
Acetone	<0.90	ug/m3	3.0	0.90	0.5		09/08/21 09:27	67-64-1	
Benzene	<0.057	ug/m3	0.16	0.057	0.5		09/08/21 09:27	71-43-2	
Benzyl chloride	<0.44	ug/m3	1.3	0.44	0.5		09/08/21 09:27	100-44-7	
Bromodichloromethane	<0.12	ug/m3	0.68	0.12	0.5		09/08/21 09:27	75-27-4	
Bromoform	<0.81	ug/m3	2.6	0.81	0.5		09/08/21 09:27	75-25-2	
Bromomethane	<0.075	ug/m3	0.39	0.075	0.5		09/08/21 09:27	74-83-9	
1,3-Butadiene	<0.060	ug/m3	0.22	0.060	0.5		09/08/21 09:27	106-99-0	
2-Butanone (MEK)	<0.23	ug/m3	1.5	0.23	0.5		09/08/21 09:27	78-93-3	
Carbon disulfide	<0.064	ug/m3	0.32	0.064	0.5		09/08/21 09:27	75-15-0	
Carbon tetrachloride	<0.14	ug/m3	0.64	0.14	0.5		09/08/21 09:27	56-23-5	
Chlorobenzene	<0.078	ug/m3	0.47	0.078	0.5		09/08/21 09:27	108-90-7	
Chloroethane	<0.11	ug/m3	0.27	0.11	0.5		09/08/21 09:27	75-00-3	
Chloroform	<0.092	ug/m3	0.25	0.092	0.5		09/08/21 09:27	67-66-3	
Chloromethane	<0.043	ug/m3	0.21	0.043	0.5		09/08/21 09:27	74-87-3	
Cyclohexane	<0.11	ug/m3	0.88	0.11	0.5		09/08/21 09:27	110-82-7	
Dibromochloromethane	<0.26	ug/m3	0.86	0.26	0.5		09/08/21 09:27	124-48-1	
1,2-Dibromoethane (EDB)	<0.15	ug/m3	0.39	0.15	0.5		09/08/21 09:27	106-93-4	
1,2-Dichlorobenzene	<0.20	ug/m3	1.5	0.20	0.5		09/08/21 09:27	95-50-1	
1,3-Dichlorobenzene	<0.25	ug/m3	1.5	0.25	0.5		09/08/21 09:27	541-73-1	
1,4-Dichlorobenzene	<0.44	ug/m3	1.5	0.44	0.5		09/08/21 09:27	106-46-7	
Dichlorodifluoromethane	<0.094	ug/m3	0.50	0.094	0.5		09/08/21 09:27	75-71-8	
1,1-Dichloroethane	<0.082	ug/m3	0.41	0.082	0.5		09/08/21 09:27	75-34-3	
1,2-Dichloroethane	<0.097	ug/m3	0.41	0.097	0.5		09/08/21 09:27	107-06-2	
1,1-Dichloroethene	<0.069	ug/m3	0.40	0.069	0.5		09/08/21 09:27	75-35-4	
cis-1,2-Dichloroethene	<0.098	ug/m3	0.40	0.098	0.5		09/08/21 09:27	156-59-2	
trans-1,2-Dichloroethene	<0.084	ug/m3	0.40	0.084	0.5		09/08/21 09:27	156-60-5	
1,2-Dichloropropane	<0.13	ug/m3	0.47	0.13	0.5		09/08/21 09:27	78-87-5	
cis-1,3-Dichloropropene	<0.13	ug/m3	1.2	0.13	0.5		09/08/21 09:27	10061-01-5	
trans-1,3-Dichloropropene	<0.27	ug/m3	1.2	0.27	0.5		09/08/21 09:27	10061-02-6	
Dichlorotetrafluoroethane	<0.10	ug/m3	0.71	0.10	0.5		09/08/21 09:27	76-14-2	
Ethanol	<0.30	ug/m3	0.96	0.30	0.5		09/08/21 09:27	64-17-5	
Ethyl acetate	<0.066	ug/m3	0.37	0.066	0.5		09/08/21 09:27	141-78-6	
Ethylbenzene	<0.15	ug/m3	0.44	0.15	0.5		09/08/21 09:27	100-41-4	
4-Ethyltoluene	<0.24	ug/m3	1.2	0.24	0.5		09/08/21 09:27	622-96-8	
n-Heptane	<0.090	ug/m3	0.42	0.090	0.5		09/08/21 09:27	142-82-5	
Hexachloro-1,3-butadiene	<0.62	ug/m3	2.7	0.62	0.5		09/08/21 09:27	87-68-3	
n-Hexane	<0.096	ug/m3	0.36	0.096	0.5		09/08/21 09:27	110-54-3	
2-Hexanone	<0.22	ug/m3	2.1	0.22	0.5		09/08/21 09:27	591-78-6	
Methylene Chloride	<0.30	ug/m3	1.8	0.30	0.5		09/08/21 09:27	75-09-2	
4-Methyl-2-pentanone (MIBK)	<0.16	ug/m3	2.1	0.16	0.5		09/08/21 09:27	108-10-1	
Methyl-tert-butyl ether	<0.063	ug/m3	1.8	0.063	0.5		09/08/21 09:27	1634-04-4	
Naphthalene	<1.1	ug/m3	1.3	1.1	0.5		09/08/21 09:27	91-20-3	
2-Propanol	<0.25	ug/m3	1.2	0.25	0.5		09/08/21 09:27	67-63-0	
Propylene	<0.065	ug/m3	0.44	0.065	0.5		09/08/21 09:27	115-07-1	
Styrene	<0.19	ug/m3	0.43	0.19	0.5		09/08/21 09:27	100-42-5	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.

ANALYTICAL RESULTS

Project: CONBRACO
Pace Project No.: 92561801

Sample: IA-11 Cert# 3408 **Lab ID: 92561801029** Collected: 09/14/21 18:45 Received: 09/16/21 13:15 Matrix: Air

Parameters	Results	Units	Report			Prepared	Analyzed	CAS No.	Qual
			Limit	MDL	DF				
Individual Can Certification									
Analytical Method: TO-15									
Pace Analytical Services - Minneapolis									
1,1,2,2-Tetrachloroethane	<0.19	ug/m3	0.70	0.19	0.5		09/08/21 09:27	79-34-5	
Tetrachloroethene	<0.15	ug/m3	0.34	0.15	0.5		09/08/21 09:27	127-18-4	
Tetrahydrofuran	<0.090	ug/m3	0.30	0.090	0.5		09/08/21 09:27	109-99-9	
Toluene	<0.12	ug/m3	0.38	0.12	0.5		09/08/21 09:27	108-88-3	
1,2,4-Trichlorobenzene	<2.4	ug/m3	3.8	2.4	0.5		09/08/21 09:27	120-82-1	
1,1,1-Trichloroethane	<0.093	ug/m3	0.56	0.093	0.5		09/08/21 09:27	71-55-6	
1,1,2-Trichloroethane	<0.098	ug/m3	0.28	0.098	0.5		09/08/21 09:27	79-00-5	
Trichloroethene	<0.098	ug/m3	0.55	0.098	0.5		09/08/21 09:27	79-01-6	
Trichlorofluoromethane	<0.12	ug/m3	0.57	0.12	0.5		09/08/21 09:27	75-69-4	
1,1,2-Trichlorotrifluoroethane	<0.14	ug/m3	0.78	0.14	0.5		09/08/21 09:27	76-13-1	
1,2,4-Trimethylbenzene	<0.18	ug/m3	0.50	0.18	0.5		09/08/21 09:27	95-63-6	
1,3,5-Trimethylbenzene	<0.14	ug/m3	0.50	0.14	0.5		09/08/21 09:27	108-67-8	
Vinyl acetate	<0.10	ug/m3	0.36	0.10	0.5		09/08/21 09:27	108-05-4	
Vinyl chloride	<0.043	ug/m3	0.13	0.043	0.5		09/08/21 09:27	75-01-4	
m&p-Xylene	<0.32	ug/m3	0.88	0.32	0.5		09/08/21 09:27	179601-23-1	
o-Xylene	<0.14	ug/m3	0.44	0.14	0.5		09/08/21 09:27	95-47-6	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.

ANALYTICAL RESULTS

Project: CONBRACO

Pace Project No.: 92561801

Sample: IA-4 Cert# 2049 **Lab ID: 92561801030** Collected: 09/14/21 16:16 Received: 09/16/21 13:15 Matrix: Air

Parameters	Results	Units	Report			Prepared	Analyzed	CAS No.	Qual
			Limit	MDL	DF				
Individual Can Certification									
Analytical Method: TO-15									
Pace Analytical Services - Minneapolis									
Acetone	<0.90	ug/m3	3.0	0.90	0.5		09/09/21 04:27	67-64-1	
Benzene	<0.057	ug/m3	0.16	0.057	0.5		09/09/21 04:27	71-43-2	
Benzyl chloride	<0.44	ug/m3	1.3	0.44	0.5		09/09/21 04:27	100-44-7	
Bromodichloromethane	<0.12	ug/m3	0.68	0.12	0.5		09/09/21 04:27	75-27-4	
Bromoform	<0.81	ug/m3	2.6	0.81	0.5		09/09/21 04:27	75-25-2	
Bromomethane	<0.075	ug/m3	0.39	0.075	0.5		09/09/21 04:27	74-83-9	
1,3-Butadiene	<0.060	ug/m3	0.22	0.060	0.5		09/09/21 04:27	106-99-0	
2-Butanone (MEK)	<0.23	ug/m3	1.5	0.23	0.5		09/09/21 04:27	78-93-3	
Carbon disulfide	<0.064	ug/m3	0.32	0.064	0.5		09/09/21 04:27	75-15-0	
Carbon tetrachloride	<0.14	ug/m3	0.64	0.14	0.5		09/09/21 04:27	56-23-5	
Chlorobenzene	<0.078	ug/m3	0.47	0.078	0.5		09/09/21 04:27	108-90-7	
Chloroethane	<0.11	ug/m3	0.27	0.11	0.5		09/09/21 04:27	75-00-3	
Chloroform	<0.092	ug/m3	0.25	0.092	0.5		09/09/21 04:27	67-66-3	
Chloromethane	<0.043	ug/m3	0.21	0.043	0.5		09/09/21 04:27	74-87-3	
Cyclohexane	<0.11	ug/m3	0.88	0.11	0.5		09/09/21 04:27	110-82-7	
Dibromochloromethane	<0.26	ug/m3	0.86	0.26	0.5		09/09/21 04:27	124-48-1	
1,2-Dibromoethane (EDB)	<0.15	ug/m3	0.39	0.15	0.5		09/09/21 04:27	106-93-4	
1,2-Dichlorobenzene	<0.20	ug/m3	1.5	0.20	0.5		09/09/21 04:27	95-50-1	
1,3-Dichlorobenzene	<0.25	ug/m3	1.5	0.25	0.5		09/09/21 04:27	541-73-1	
1,4-Dichlorobenzene	<0.44	ug/m3	1.5	0.44	0.5		09/09/21 04:27	106-46-7	
Dichlorodifluoromethane	<0.094	ug/m3	0.50	0.094	0.5		09/09/21 04:27	75-71-8	
1,1-Dichloroethane	<0.082	ug/m3	0.41	0.082	0.5		09/09/21 04:27	75-34-3	
1,2-Dichloroethane	<0.097	ug/m3	0.41	0.097	0.5		09/09/21 04:27	107-06-2	
1,1-Dichloroethene	<0.069	ug/m3	0.40	0.069	0.5		09/09/21 04:27	75-35-4	
cis-1,2-Dichloroethene	<0.098	ug/m3	0.40	0.098	0.5		09/09/21 04:27	156-59-2	
trans-1,2-Dichloroethene	<0.084	ug/m3	0.40	0.084	0.5		09/09/21 04:27	156-60-5	
1,2-Dichloropropane	<0.13	ug/m3	0.47	0.13	0.5		09/09/21 04:27	78-87-5	
cis-1,3-Dichloropropene	<0.13	ug/m3	1.2	0.13	0.5		09/09/21 04:27	10061-01-5	
trans-1,3-Dichloropropene	<0.27	ug/m3	1.2	0.27	0.5		09/09/21 04:27	10061-02-6	
Dichlorotetrafluoroethane	<0.10	ug/m3	0.71	0.10	0.5		09/09/21 04:27	76-14-2	
Ethanol	<0.30	ug/m3	0.96	0.30	0.5		09/09/21 04:27	64-17-5	
Ethyl acetate	<0.066	ug/m3	0.37	0.066	0.5		09/09/21 04:27	141-78-6	
Ethylbenzene	<0.15	ug/m3	0.44	0.15	0.5		09/09/21 04:27	100-41-4	
4-Ethyltoluene	<0.24	ug/m3	1.2	0.24	0.5		09/09/21 04:27	622-96-8	
n-Heptane	<0.090	ug/m3	0.42	0.090	0.5		09/09/21 04:27	142-82-5	
Hexachloro-1,3-butadiene	<0.62	ug/m3	2.7	0.62	0.5		09/09/21 04:27	87-68-3	
n-Hexane	<0.096	ug/m3	0.36	0.096	0.5		09/09/21 04:27	110-54-3	
2-Hexanone	<0.22	ug/m3	2.1	0.22	0.5		09/09/21 04:27	591-78-6	
Methylene Chloride	<0.30	ug/m3	1.8	0.30	0.5		09/09/21 04:27	75-09-2	
4-Methyl-2-pentanone (MIBK)	<0.16	ug/m3	2.1	0.16	0.5		09/09/21 04:27	108-10-1	
Methyl-tert-butyl ether	<0.063	ug/m3	1.8	0.063	0.5		09/09/21 04:27	1634-04-4	
Naphthalene	<1.1	ug/m3	1.3	1.1	0.5		09/09/21 04:27	91-20-3	
2-Propanol	<0.25	ug/m3	1.2	0.25	0.5		09/09/21 04:27	67-63-0	
Propylene	<0.065	ug/m3	0.44	0.065	0.5		09/09/21 04:27	115-07-1	
Styrene	<0.19	ug/m3	0.43	0.19	0.5		09/09/21 04:27	100-42-5	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.

ANALYTICAL RESULTS

Project: CONBRACO
Pace Project No.: 92561801

Sample: IA-4 Cert# 2049 **Lab ID: 92561801030** Collected: 09/14/21 16:16 Received: 09/16/21 13:15 Matrix: Air

Parameters	Results	Units	Report			Prepared	Analyzed	CAS No.	Qual
			Limit	MDL	DF				
Individual Can Certification									
Analytical Method: TO-15									
Pace Analytical Services - Minneapolis									
1,1,2,2-Tetrachloroethane	<0.19	ug/m3	0.70	0.19	0.5		09/09/21 04:27	79-34-5	
Tetrachloroethene	<0.15	ug/m3	0.34	0.15	0.5		09/09/21 04:27	127-18-4	
Tetrahydrofuran	<0.090	ug/m3	0.30	0.090	0.5		09/09/21 04:27	109-99-9	
Toluene	<0.12	ug/m3	0.38	0.12	0.5		09/09/21 04:27	108-88-3	
1,2,4-Trichlorobenzene	<2.4	ug/m3	3.8	2.4	0.5		09/09/21 04:27	120-82-1	
1,1,1-Trichloroethane	<0.093	ug/m3	0.56	0.093	0.5		09/09/21 04:27	71-55-6	
1,1,2-Trichloroethane	<0.098	ug/m3	0.28	0.098	0.5		09/09/21 04:27	79-00-5	
Trichloroethene	<0.098	ug/m3	0.27	0.098	0.5		09/09/21 04:27	79-01-6	
Trichlorofluoromethane	<0.12	ug/m3	0.57	0.12	0.5		09/09/21 04:27	75-69-4	
1,1,2-Trichlorotrifluoroethane	<0.14	ug/m3	0.78	0.14	0.5		09/09/21 04:27	76-13-1	
1,2,4-Trimethylbenzene	<0.18	ug/m3	0.50	0.18	0.5		09/09/21 04:27	95-63-6	
1,3,5-Trimethylbenzene	<0.14	ug/m3	0.50	0.14	0.5		09/09/21 04:27	108-67-8	
Vinyl acetate	<0.10	ug/m3	0.36	0.10	0.5		09/09/21 04:27	108-05-4	
Vinyl chloride	<0.043	ug/m3	0.13	0.043	0.5		09/09/21 04:27	75-01-4	
m&p-Xylene	<0.32	ug/m3	0.88	0.32	0.5		09/09/21 04:27	179601-23-1	
o-Xylene	<0.14	ug/m3	0.44	0.14	0.5		09/09/21 04:27	95-47-6	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.

ANALYTICAL RESULTS

Project: CONBRACO

Pace Project No.: 92561801

Sample: IA-9 Cert# 0857 Lab ID: 92561801031 Collected: 09/14/21 18:04 Received: 09/16/21 13:15 Matrix: Air

Parameters	Results	Units	Report			Prepared	Analyzed	CAS No.	Qual
			Limit	MDL	DF				
Individual Can Certification									
Analytical Method: TO-15									
Pace Analytical Services - Minneapolis									
Acetone	<0.90	ug/m3	3.0	0.90	0.5		09/08/21 09:58	67-64-1	
Benzene	<0.057	ug/m3	0.16	0.057	0.5		09/08/21 09:58	71-43-2	
Benzyl chloride	<0.44	ug/m3	1.3	0.44	0.5		09/08/21 09:58	100-44-7	
Bromodichloromethane	<0.12	ug/m3	0.68	0.12	0.5		09/08/21 09:58	75-27-4	
Bromoform	<0.81	ug/m3	2.6	0.81	0.5		09/08/21 09:58	75-25-2	
Bromomethane	<0.075	ug/m3	0.39	0.075	0.5		09/08/21 09:58	74-83-9	
1,3-Butadiene	<0.060	ug/m3	0.22	0.060	0.5		09/08/21 09:58	106-99-0	
2-Butanone (MEK)	<0.23	ug/m3	1.5	0.23	0.5		09/08/21 09:58	78-93-3	
Carbon disulfide	<0.064	ug/m3	0.32	0.064	0.5		09/08/21 09:58	75-15-0	
Carbon tetrachloride	<0.14	ug/m3	0.64	0.14	0.5		09/08/21 09:58	56-23-5	
Chlorobenzene	<0.078	ug/m3	0.47	0.078	0.5		09/08/21 09:58	108-90-7	
Chloroethane	<0.11	ug/m3	0.27	0.11	0.5		09/08/21 09:58	75-00-3	
Chloroform	<0.092	ug/m3	0.25	0.092	0.5		09/08/21 09:58	67-66-3	
Chloromethane	<0.043	ug/m3	0.21	0.043	0.5		09/08/21 09:58	74-87-3	
Cyclohexane	<0.11	ug/m3	0.88	0.11	0.5		09/08/21 09:58	110-82-7	
Dibromochloromethane	<0.26	ug/m3	0.86	0.26	0.5		09/08/21 09:58	124-48-1	
1,2-Dibromoethane (EDB)	<0.15	ug/m3	0.39	0.15	0.5		09/08/21 09:58	106-93-4	
1,2-Dichlorobenzene	<0.20	ug/m3	1.5	0.20	0.5		09/08/21 09:58	95-50-1	
1,3-Dichlorobenzene	<0.25	ug/m3	1.5	0.25	0.5		09/08/21 09:58	541-73-1	
1,4-Dichlorobenzene	<0.44	ug/m3	1.5	0.44	0.5		09/08/21 09:58	106-46-7	
Dichlorodifluoromethane	<0.094	ug/m3	0.50	0.094	0.5		09/08/21 09:58	75-71-8	
1,1-Dichloroethane	<0.082	ug/m3	0.41	0.082	0.5		09/08/21 09:58	75-34-3	
1,2-Dichloroethane	<0.097	ug/m3	0.41	0.097	0.5		09/08/21 09:58	107-06-2	
1,1-Dichloroethene	<0.069	ug/m3	0.40	0.069	0.5		09/08/21 09:58	75-35-4	
cis-1,2-Dichloroethene	<0.098	ug/m3	0.40	0.098	0.5		09/08/21 09:58	156-59-2	
trans-1,2-Dichloroethene	<0.084	ug/m3	0.40	0.084	0.5		09/08/21 09:58	156-60-5	
1,2-Dichloropropane	<0.13	ug/m3	0.47	0.13	0.5		09/08/21 09:58	78-87-5	
cis-1,3-Dichloropropene	<0.13	ug/m3	1.2	0.13	0.5		09/08/21 09:58	10061-01-5	
trans-1,3-Dichloropropene	<0.27	ug/m3	1.2	0.27	0.5		09/08/21 09:58	10061-02-6	
Dichlorotetrafluoroethane	<0.10	ug/m3	0.71	0.10	0.5		09/08/21 09:58	76-14-2	
Ethanol	<0.30	ug/m3	0.96	0.30	0.5		09/08/21 09:58	64-17-5	
Ethyl acetate	<0.066	ug/m3	0.37	0.066	0.5		09/08/21 09:58	141-78-6	
Ethylbenzene	<0.15	ug/m3	0.44	0.15	0.5		09/08/21 09:58	100-41-4	
4-Ethyltoluene	<0.24	ug/m3	1.2	0.24	0.5		09/08/21 09:58	622-96-8	
n-Heptane	<0.090	ug/m3	0.42	0.090	0.5		09/08/21 09:58	142-82-5	
Hexachloro-1,3-butadiene	<0.62	ug/m3	2.7	0.62	0.5		09/08/21 09:58	87-68-3	
n-Hexane	<0.096	ug/m3	0.36	0.096	0.5		09/08/21 09:58	110-54-3	
2-Hexanone	<0.22	ug/m3	2.1	0.22	0.5		09/08/21 09:58	591-78-6	
Methylene Chloride	<0.30	ug/m3	1.8	0.30	0.5		09/08/21 09:58	75-09-2	
4-Methyl-2-pentanone (MIBK)	<0.16	ug/m3	2.1	0.16	0.5		09/08/21 09:58	108-10-1	
Methyl-tert-butyl ether	<0.063	ug/m3	1.8	0.063	0.5		09/08/21 09:58	1634-04-4	
Naphthalene	<1.1	ug/m3	1.3	1.1	0.5		09/08/21 09:58	91-20-3	
2-Propanol	<0.25	ug/m3	1.2	0.25	0.5		09/08/21 09:58	67-63-0	
Propylene	<0.065	ug/m3	0.44	0.065	0.5		09/08/21 09:58	115-07-1	
Styrene	<0.19	ug/m3	0.43	0.19	0.5		09/08/21 09:58	100-42-5	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.

ANALYTICAL RESULTS

Project: CONBRACO
Pace Project No.: 92561801

Sample: IA-9 Cert# 0857 **Lab ID: 92561801031** Collected: 09/14/21 18:04 Received: 09/16/21 13:15 Matrix: Air

Parameters	Results	Units	Report			Prepared	Analyzed	CAS No.	Qual
			Limit	MDL	DF				
Individual Can Certification									
Analytical Method: TO-15									
Pace Analytical Services - Minneapolis									
1,1,2,2-Tetrachloroethane	<0.19	ug/m3	0.70	0.19	0.5		09/08/21 09:58	79-34-5	
Tetrachloroethene	<0.15	ug/m3	0.34	0.15	0.5		09/08/21 09:58	127-18-4	
Tetrahydrofuran	<0.090	ug/m3	0.30	0.090	0.5		09/08/21 09:58	109-99-9	
Toluene	<0.12	ug/m3	0.38	0.12	0.5		09/08/21 09:58	108-88-3	
1,2,4-Trichlorobenzene	<2.4	ug/m3	3.8	2.4	0.5		09/08/21 09:58	120-82-1	
1,1,1-Trichloroethane	<0.093	ug/m3	0.56	0.093	0.5		09/08/21 09:58	71-55-6	
1,1,2-Trichloroethane	<0.098	ug/m3	0.28	0.098	0.5		09/08/21 09:58	79-00-5	
Trichloroethene	<0.098	ug/m3	0.55	0.098	0.5		09/08/21 09:58	79-01-6	
Trichlorofluoromethane	<0.12	ug/m3	0.57	0.12	0.5		09/08/21 09:58	75-69-4	
1,1,2-Trichlorotrifluoroethane	<0.14	ug/m3	0.78	0.14	0.5		09/08/21 09:58	76-13-1	
1,2,4-Trimethylbenzene	<0.18	ug/m3	0.50	0.18	0.5		09/08/21 09:58	95-63-6	
1,3,5-Trimethylbenzene	<0.14	ug/m3	0.50	0.14	0.5		09/08/21 09:58	108-67-8	
Vinyl acetate	<0.10	ug/m3	0.36	0.10	0.5		09/08/21 09:58	108-05-4	
Vinyl chloride	<0.043	ug/m3	0.13	0.043	0.5		09/08/21 09:58	75-01-4	
m&p-Xylene	<0.32	ug/m3	0.88	0.32	0.5		09/08/21 09:58	179601-23-1	
o-Xylene	<0.14	ug/m3	0.44	0.14	0.5		09/08/21 09:58	95-47-6	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.

ANALYTICAL RESULTS

Project: CONBRACO
Pace Project No.: 92561801

Sample: Background Cert# 3891 **Lab ID: 92561801032** Collected: 09/14/21 18:08 Received: 09/16/21 13:15 Matrix: Air

Parameters	Results	Units	Report			Prepared	Analyzed	CAS No.	Qual
			Limit	MDL	DF				
Individual Can Certification									
Analytical Method: TO-15									
Pace Analytical Services - Minneapolis									
Acetone	<0.90	ug/m3	3.0	0.90	0.5		09/07/21 10:49	67-64-1	
Benzene	<0.057	ug/m3	0.16	0.057	0.5		09/07/21 10:49	71-43-2	
Benzyl chloride	<0.44	ug/m3	1.3	0.44	0.5		09/07/21 10:49	100-44-7	
Bromodichloromethane	<0.12	ug/m3	0.68	0.12	0.5		09/07/21 10:49	75-27-4	
Bromoform	<0.81	ug/m3	2.6	0.81	0.5		09/07/21 10:49	75-25-2	
Bromomethane	<0.075	ug/m3	0.39	0.075	0.5		09/07/21 10:49	74-83-9	
1,3-Butadiene	<0.060	ug/m3	0.22	0.060	0.5		09/07/21 10:49	106-99-0	
2-Butanone (MEK)	<0.23	ug/m3	1.5	0.23	0.5		09/07/21 10:49	78-93-3	
Carbon disulfide	<0.064	ug/m3	0.32	0.064	0.5		09/07/21 10:49	75-15-0	
Carbon tetrachloride	<0.14	ug/m3	0.64	0.14	0.5		09/07/21 10:49	56-23-5	
Chlorobenzene	<0.078	ug/m3	0.47	0.078	0.5		09/07/21 10:49	108-90-7	
Chloroethane	<0.11	ug/m3	0.27	0.11	0.5		09/07/21 10:49	75-00-3	
Chloroform	<0.092	ug/m3	0.25	0.092	0.5		09/07/21 10:49	67-66-3	
Chloromethane	<0.043	ug/m3	0.21	0.043	0.5		09/07/21 10:49	74-87-3	
Cyclohexane	<0.11	ug/m3	0.88	0.11	0.5		09/07/21 10:49	110-82-7	
Dibromochloromethane	<0.26	ug/m3	0.86	0.26	0.5		09/07/21 10:49	124-48-1	
1,2-Dibromoethane (EDB)	<0.15	ug/m3	0.39	0.15	0.5		09/07/21 10:49	106-93-4	
1,2-Dichlorobenzene	0.39J	ug/m3	1.5	0.20	0.5		09/07/21 10:49	95-50-1	
1,3-Dichlorobenzene	<0.25	ug/m3	1.5	0.25	0.5		09/07/21 10:49	541-73-1	
1,4-Dichlorobenzene	<0.44	ug/m3	1.5	0.44	0.5		09/07/21 10:49	106-46-7	
Dichlorodifluoromethane	<0.094	ug/m3	0.50	0.094	0.5		09/07/21 10:49	75-71-8	
1,1-Dichloroethane	<0.082	ug/m3	0.41	0.082	0.5		09/07/21 10:49	75-34-3	
1,2-Dichloroethane	<0.097	ug/m3	0.41	0.097	0.5		09/07/21 10:49	107-06-2	
1,1-Dichloroethene	<0.069	ug/m3	0.40	0.069	0.5		09/07/21 10:49	75-35-4	
cis-1,2-Dichloroethene	<0.098	ug/m3	0.40	0.098	0.5		09/07/21 10:49	156-59-2	
trans-1,2-Dichloroethene	<0.084	ug/m3	0.40	0.084	0.5		09/07/21 10:49	156-60-5	
1,2-Dichloropropane	<0.13	ug/m3	0.47	0.13	0.5		09/07/21 10:49	78-87-5	
cis-1,3-Dichloropropene	<0.13	ug/m3	1.2	0.13	0.5		09/07/21 10:49	10061-01-5	
trans-1,3-Dichloropropene	<0.27	ug/m3	1.2	0.27	0.5		09/07/21 10:49	10061-02-6	
Dichlorotetrafluoroethane	<0.10	ug/m3	0.71	0.10	0.5		09/07/21 10:49	76-14-2	
Ethanol	<0.30	ug/m3	0.96	0.30	0.5		09/07/21 10:49	64-17-5	
Ethyl acetate	<0.066	ug/m3	0.37	0.066	0.5		09/07/21 10:49	141-78-6	
Ethylbenzene	<0.15	ug/m3	0.44	0.15	0.5		09/07/21 10:49	100-41-4	
4-Ethyltoluene	<0.24	ug/m3	1.2	0.24	0.5		09/07/21 10:49	622-96-8	
n-Heptane	<0.090	ug/m3	0.42	0.090	0.5		09/07/21 10:49	142-82-5	
Hexachloro-1,3-butadiene	<0.62	ug/m3	2.7	0.62	0.5		09/07/21 10:49	87-68-3	
n-Hexane	<0.096	ug/m3	0.36	0.096	0.5		09/07/21 10:49	110-54-3	
2-Hexanone	<0.22	ug/m3	2.1	0.22	0.5		09/07/21 10:49	591-78-6	
Methylene Chloride	<0.30	ug/m3	1.8	0.30	0.5		09/07/21 10:49	75-09-2	
4-Methyl-2-pentanone (MIBK)	<0.16	ug/m3	2.1	0.16	0.5		09/07/21 10:49	108-10-1	
Methyl-tert-butyl ether	<0.063	ug/m3	1.8	0.063	0.5		09/07/21 10:49	1634-04-4	
Naphthalene	<1.1	ug/m3	1.3	1.1	0.5		09/07/21 10:49	91-20-3	
2-Propanol	<0.25	ug/m3	1.2	0.25	0.5		09/07/21 10:49	67-63-0	
Propylene	<0.065	ug/m3	0.44	0.065	0.5		09/07/21 10:49	115-07-1	
Styrene	<0.19	ug/m3	0.43	0.19	0.5		09/07/21 10:49	100-42-5	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.

ANALYTICAL RESULTS

Project: CONBRACO
Pace Project No.: 92561801

Sample: Background Cert# 3891 **Lab ID: 92561801032** Collected: 09/14/21 18:08 Received: 09/16/21 13:15 Matrix: Air

Parameters	Results	Units	Report			Prepared	Analyzed	CAS No.	Qual
			Limit	MDL	DF				
Individual Can Certification									
Analytical Method: TO-15									
Pace Analytical Services - Minneapolis									
1,1,2,2-Tetrachloroethane	<0.19	ug/m3	0.70	0.19	0.5		09/07/21 10:49	79-34-5	
Tetrachloroethene	<0.15	ug/m3	0.34	0.15	0.5		09/07/21 10:49	127-18-4	
Tetrahydrofuran	<0.090	ug/m3	0.30	0.090	0.5		09/07/21 10:49	109-99-9	
Toluene	<0.12	ug/m3	0.38	0.12	0.5		09/07/21 10:49	108-88-3	
1,2,4-Trichlorobenzene	<2.4	ug/m3	3.8	2.4	0.5		09/07/21 10:49	120-82-1	
1,1,1-Trichloroethane	<0.093	ug/m3	0.56	0.093	0.5		09/07/21 10:49	71-55-6	
1,1,2-Trichloroethane	<0.098	ug/m3	0.28	0.098	0.5		09/07/21 10:49	79-00-5	
Trichloroethene	<0.098	ug/m3	0.27	0.098	0.5		09/07/21 10:49	79-01-6	
Trichlorofluoromethane	<0.12	ug/m3	0.57	0.12	0.5		09/07/21 10:49	75-69-4	
1,1,2-Trichlorotrifluoroethane	<0.14	ug/m3	0.78	0.14	0.5		09/07/21 10:49	76-13-1	
1,2,4-Trimethylbenzene	<0.18	ug/m3	0.50	0.18	0.5		09/07/21 10:49	95-63-6	
1,3,5-Trimethylbenzene	<0.14	ug/m3	0.50	0.14	0.5		09/07/21 10:49	108-67-8	
Vinyl acetate	<0.10	ug/m3	0.36	0.10	0.5		09/07/21 10:49	108-05-4	
Vinyl chloride	<0.043	ug/m3	0.13	0.043	0.5		09/07/21 10:49	75-01-4	
m&p-Xylene	<0.32	ug/m3	0.88	0.32	0.5		09/07/21 10:49	179601-23-1	
o-Xylene	<0.14	ug/m3	0.44	0.14	0.5		09/07/21 10:49	95-47-6	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.

QUALIFIERS

Project: CONBRACO

Pace Project No.: 92561801

DEFINITIONS

DF - Dilution Factor, if reported, represents the factor applied to the reported data due to dilution of the sample aliquot.

ND - Not Detected at or above adjusted reporting limit.

TNTC - Too Numerous To Count

J - Estimated concentration above the adjusted method detection limit and below the adjusted reporting limit.

MDL - Adjusted Method Detection Limit.

PQL - Practical Quantitation Limit.

RL - Reporting Limit - The lowest concentration value that meets project requirements for quantitative data with known precision and bias for a specific analyte in a specific matrix.

S - Surrogate

1,2-Diphenylhydrazine decomposes to and cannot be separated from Azobenzene using Method 8270. The result for each analyte is a combined concentration.

Consistent with EPA guidelines, unrounded data are displayed and have been used to calculate % recovery and RPD values.

LCS(D) - Laboratory Control Sample (Duplicate)

MS(D) - Matrix Spike (Duplicate)

DUP - Sample Duplicate

RPD - Relative Percent Difference

NC - Not Calculable.

SG - Silica Gel - Clean-Up

U - Indicates the compound was analyzed for, but not detected.

Acid preservation may not be appropriate for 2 Chloroethylvinyl ether.

A separate vial preserved to a pH of 4-5 is recommended in SW846 Chapter 4 for the analysis of Acrolein and Acrylonitrile by EPA Method 8260.

N-Nitrosodiphenylamine decomposes and cannot be separated from Diphenylamine using Method 8270. The result reported for each analyte is a combined concentration.

Reported results are not rounded until the final step prior to reporting. Therefore, calculated parameters that are typically reported as "Total" may vary slightly from the sum of the reported component parameters.

Pace Analytical is TNI accredited. Contact your Pace PM for the current list of accredited analytes.

TNI - The NELAC Institute.

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.

QUALITY CONTROL DATA CROSS REFERENCE TABLE

Project: CONBRACO

Pace Project No.: 92561801

Lab ID	Sample ID	QC Batch Method	QC Batch	Analytical Method	Analytical Batch
92561801001	IA-14	TO-15	771937		
92561801002	IA-6	TO-15	771937		
92561801003	IA-1	TO-15	771937		
92561801004	IA-12	TO-15	771937		
92561801005	IA-2	TO-15	771937		
92561801006	IA-8	TO-15	771937		
92561801007	IA-5	TO-15	771937		
92561801008	DUP-1	TO-15	771937		
92561801009	IA-13	TO-15	771937		
92561801010	IA-7	TO-15	771937		
92561801011	IA-3	TO-15	771937		
92561801012	IA-10	TO-15	771937		
92561801013	IA-11	TO-15	771937		
92561801014	IA-4	TO-15	771937		
92561801015	IA-9	TO-15	771937		
92561801016	BACKGROUND	TO-15	771937		
92561801017	IA-14 Cert# 1604	TO-15	771546		
92561801018	IA-6 Cert# 2319	TO-15	771546		
92561801019	IA-1 Cert #0799	TO-15	771546		
92561801020	IA-12 Cert# 0145	TO-15	771546		
92561801021	IA-2 Cert# 0796	TO-15	771546		
92561801022	IA-8 Cert# 2109	TO-15	771546		
92561801023	IA-5 Cert# 2811	TO-15	771546		
92561801024	DUP-1 Cert#1195	TO-15	771546		
92561801025	IA-13 Cert# 2803	TO-15	771546		
92561801026	IA-7 Cert# 3629	TO-15	771546		
92561801027	IA-3 Cert # 0629	TO-15	771546		
92561801028	IA-10 Cert# 3475	TO-15	771546		
92561801029	IA-11 Cert# 3408	TO-15	771546		
92561801030	IA-4 Cert# 2049	TO-15	771546		
92561801031	IA-9 Cert# 0857	TO-15	771546		
92561801032	Background Cert# 3891	TO-15	771546		

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.

WO#: 92561801



92561801

AIR: CHAIN-OF-CUSTODY / Analytical Request Document

The Chain-of-Custody is a LEGAL DOCUMENT. All relevant fields must be completed accurately.

Section A Required Client Information: Company: Shield Engineering Address: 4301 Target Creek Rd Charlotte, NC 28208 Email To: wbarfield@shieldengineering.com Phone: 204-394-6913 Fax: Requested Due Date/TAT: ASAP w/out extra charge		Section B Required Project Information: Report To: wbarfield@shieldengineering.com Copy To: Purchase Order No.: 1030214-09 Project Name: Co-brac Project Number: 1030214-09		Section C Invoice Information: Attention: Les Barfield Company Name: Address: Pace Quote Reference: Pace Project Manager/Sales Rep. Pace Profile #: 42604		51398 Page: 1 of 2										
Section D Required Client Information AIR SAMPLE ID Sample IDs MUST BE UNIQUE		Valid Media Codes MEDIA CODE Tedlar Bag TB 1 Liter Summa Can 1LC 6 Liter Summa Can 6LC Low Volume Puff LVP High Volume Puff HVP Other PM10		COLLECTED MEDIA CODE PID Reading (Client only)		ACCEPTED BY / AFFILIATION DATE TIME										
ITEM #	IA-14	IA-6	IA-1	IA-12	IA-2	IA-8	IA-5	DUP-1	IA-13	IA-7	IA-3	IA-10				
	9-14-21 9:45	9-14-21 9:00	9-14-21 10:50	9-14-21 10:00	9-14-21 10:48	9-14-21 9:34	9-14-21 9:13	9-14-21 9:57	9-14-21 9:57	9-14-21 9:36	9-14-21 9:08	9-14-21 10:07	9-15-21 9:15	9-16-2009 12:00		
	6LC	6LC	6LC	6LC	6LC	6LC	6LC	6LC	6LC	6LC	6LC	6LC	6LC	6LC	6LC	6LC
	9:32	16:09	18:15	17:35	18:16	16:52	16:15	17:04	17:04	16:56	16:48	17:25	12:47	12:00		
	-33	-29	-32	-30	-32	-27	-29	-28	-28	-29	-31	-30	1315	1200		
	0261	0335	0866	2091	1275	1104	1283	0744	1407	1681	2932	2075				
	0261	0262	0263	0264	0265	0266	0267	0268	0269	0270	0271	0272				
	PM10	TO-3 BTEX	TO-3M (Methane)	TO-14	TO-15 Full List VOCs	TO-15 Short List BTEX	TO-15 Short List Chlorinated	TO-15 Short List (Other)								
	Method:	Method:	Method:	Method:	Method:	Method:	Method:	Method:	Method:	Method:	Method:	Method:	Method:	Method:	Method:	Method:
	3C - Fixed Gas (%)	3C - Fixed Gas (%)	3C - Fixed Gas (%)	3C - Fixed Gas (%)	3C - Fixed Gas (%)	3C - Fixed Gas (%)	3C - Fixed Gas (%)	3C - Fixed Gas (%)	3C - Fixed Gas (%)	3C - Fixed Gas (%)	3C - Fixed Gas (%)	3C - Fixed Gas (%)	3C - Fixed Gas (%)	3C - Fixed Gas (%)	3C - Fixed Gas (%)	3C - Fixed Gas (%)
	Report Level II. <input checked="" type="checkbox"/> III. <input type="checkbox"/> IV. <input type="checkbox"/> Other	Report Level II. <input checked="" type="checkbox"/> III. <input type="checkbox"/> IV. <input type="checkbox"/> Other	Report Level II. <input checked="" type="checkbox"/> III. <input type="checkbox"/> IV. <input type="checkbox"/> Other	Report Level II. <input checked="" type="checkbox"/> III. <input type="checkbox"/> IV. <input type="checkbox"/> Other	Report Level II. <input checked="" type="checkbox"/> III. <input type="checkbox"/> IV. <input type="checkbox"/> Other	Report Level II. <input checked="" type="checkbox"/> III. <input type="checkbox"/> IV. <input type="checkbox"/> Other	Report Level II. <input checked="" type="checkbox"/> III. <input type="checkbox"/> IV. <input type="checkbox"/> Other	Report Level II. <input checked="" type="checkbox"/> III. <input type="checkbox"/> IV. <input type="checkbox"/> Other	Report Level II. <input checked="" type="checkbox"/> III. <input type="checkbox"/> IV. <input type="checkbox"/> Other	Report Level II. <input checked="" type="checkbox"/> III. <input type="checkbox"/> IV. <input type="checkbox"/> Other	Report Level II. <input checked="" type="checkbox"/> III. <input type="checkbox"/> IV. <input type="checkbox"/> Other	Report Level II. <input checked="" type="checkbox"/> III. <input type="checkbox"/> IV. <input type="checkbox"/> Other	Report Level II. <input checked="" type="checkbox"/> III. <input type="checkbox"/> IV. <input type="checkbox"/> Other	Report Level II. <input checked="" type="checkbox"/> III. <input type="checkbox"/> IV. <input type="checkbox"/> Other	Report Level II. <input checked="" type="checkbox"/> III. <input type="checkbox"/> IV. <input type="checkbox"/> Other	Report Level II. <input checked="" type="checkbox"/> III. <input type="checkbox"/> IV. <input type="checkbox"/> Other
	Location of Sampling by State NC	Location of Sampling by State NC	Location of Sampling by State NC	Location of Sampling by State NC	Location of Sampling by State NC	Location of Sampling by State NC	Location of Sampling by State NC	Location of Sampling by State NC	Location of Sampling by State NC	Location of Sampling by State NC	Location of Sampling by State NC	Location of Sampling by State NC	Location of Sampling by State NC	Location of Sampling by State NC	Location of Sampling by State NC	Location of Sampling by State NC
	Reporting Units ug/m ³ ppbv Other	Reporting Units ug/m ³ ppbv Other	Reporting Units ug/m ³ ppbv Other	Reporting Units ug/m ³ ppbv Other	Reporting Units ug/m ³ ppbv Other	Reporting Units ug/m ³ ppbv Other	Reporting Units ug/m ³ ppbv Other	Reporting Units ug/m ³ ppbv Other	Reporting Units ug/m ³ ppbv Other	Reporting Units ug/m ³ ppbv Other	Reporting Units ug/m ³ ppbv Other	Reporting Units ug/m ³ ppbv Other	Reporting Units ug/m ³ ppbv Other	Reporting Units ug/m ³ ppbv Other	Reporting Units ug/m ³ ppbv Other	Reporting Units ug/m ³ ppbv Other
	Program	Program	Program	Program	Program	Program	Program	Program	Program	Program	Program	Program	Program	Program	Program	Program
	<input type="checkbox"/> UST <input type="checkbox"/> Superfund <input type="checkbox"/> Emissions <input type="checkbox"/> Clean Air Act	<input type="checkbox"/> UST <input type="checkbox"/> Superfund <input type="checkbox"/> Emissions <input type="checkbox"/> Clean Air Act	<input type="checkbox"/> UST <input type="checkbox"/> Superfund <input type="checkbox"/> Emissions <input type="checkbox"/> Clean Air Act	<input type="checkbox"/> UST <input type="checkbox"/> Superfund <input type="checkbox"/> Emissions <input type="checkbox"/> Clean Air Act	<input type="checkbox"/> UST <input type="checkbox"/> Superfund <input type="checkbox"/> Emissions <input type="checkbox"/> Clean Air Act	<input type="checkbox"/> UST <input type="checkbox"/> Superfund <input type="checkbox"/> Emissions <input type="checkbox"/> Clean Air Act	<input type="checkbox"/> UST <input type="checkbox"/> Superfund <input type="checkbox"/> Emissions <input type="checkbox"/> Clean Air Act	<input type="checkbox"/> UST <input type="checkbox"/> Superfund <input type="checkbox"/> Emissions <input type="checkbox"/> Clean Air Act	<input type="checkbox"/> UST <input type="checkbox"/> Superfund <input type="checkbox"/> Emissions <input type="checkbox"/> Clean Air Act	<input type="checkbox"/> UST <input type="checkbox"/> Superfund <input type="checkbox"/> Emissions <input type="checkbox"/> Clean Air Act	<input type="checkbox"/> UST <input type="checkbox"/> Superfund <input type="checkbox"/> Emissions <input type="checkbox"/> Clean Air Act	<input type="checkbox"/> UST <input type="checkbox"/> Superfund <input type="checkbox"/> Emissions <input type="checkbox"/> Clean Air Act	<input type="checkbox"/> UST <input type="checkbox"/> Superfund <input type="checkbox"/> Emissions <input type="checkbox"/> Clean Air Act	<input type="checkbox"/> UST <input type="checkbox"/> Superfund <input type="checkbox"/> Emissions <input type="checkbox"/> Clean Air Act	<input type="checkbox"/> UST <input type="checkbox"/> Superfund <input type="checkbox"/> Emissions <input type="checkbox"/> Clean Air Act	<input type="checkbox"/> UST <input type="checkbox"/> Superfund <input type="checkbox"/> Emissions <input type="checkbox"/> Clean Air Act
	<input type="checkbox"/> Voluntary Clean Up <input type="checkbox"/> Dry Clean <input checked="" type="checkbox"/> RCRA <input type="checkbox"/> Other	<input type="checkbox"/> Voluntary Clean Up <input type="checkbox"/> Dry Clean <input checked="" type="checkbox"/> RCRA <input type="checkbox"/> Other	<input type="checkbox"/> Voluntary Clean Up <input type="checkbox"/> Dry Clean <input checked="" type="checkbox"/> RCRA <input type="checkbox"/> Other	<input type="checkbox"/> Voluntary Clean Up <input type="checkbox"/> Dry Clean <input checked="" type="checkbox"/> RCRA <input type="checkbox"/> Other	<input type="checkbox"/> Voluntary Clean Up <input type="checkbox"/> Dry Clean <input checked="" type="checkbox"/> RCRA <input type="checkbox"/> Other	<input type="checkbox"/> Voluntary Clean Up <input type="checkbox"/> Dry Clean <input checked="" type="checkbox"/> RCRA <input type="checkbox"/> Other	<input type="checkbox"/> Voluntary Clean Up <input type="checkbox"/> Dry Clean <input checked="" type="checkbox"/> RCRA <input type="checkbox"/> Other	<input type="checkbox"/> Voluntary Clean Up <input type="checkbox"/> Dry Clean <input checked="" type="checkbox"/> RCRA <input type="checkbox"/> Other	<input type="checkbox"/> Voluntary Clean Up <input type="checkbox"/> Dry Clean <input checked="" type="checkbox"/> RCRA <input type="checkbox"/> Other	<input type="checkbox"/> Voluntary Clean Up <input type="checkbox"/> Dry Clean <input checked="" type="checkbox"/> RCRA <input type="checkbox"/> Other	<input type="checkbox"/> Voluntary Clean Up <input type="checkbox"/> Dry Clean <input checked="" type="checkbox"/> RCRA <input type="checkbox"/> Other	<input type="checkbox"/> Voluntary Clean Up <input type="checkbox"/> Dry Clean <input checked="" type="checkbox"/> RCRA <input type="checkbox"/> Other	<input type="checkbox"/> Voluntary Clean Up <input type="checkbox"/> Dry Clean <input checked="" type="checkbox"/> RCRA <input type="checkbox"/> Other	<input type="checkbox"/> Voluntary Clean Up <input type="checkbox"/> Dry Clean <input checked="" type="checkbox"/> RCRA <input type="checkbox"/> Other	<input type="checkbox"/> Voluntary Clean Up <input type="checkbox"/> Dry Clean <input checked="" type="checkbox"/> RCRA <input type="checkbox"/> Other	<input type="checkbox"/> Voluntary Clean Up <input type="checkbox"/> Dry Clean <input checked="" type="checkbox"/> RCRA <input type="checkbox"/> Other
	Temp In C	Temp In C	Temp In C	Temp In C	Temp In C	Temp In C	Temp In C	Temp In C	Temp In C	Temp In C	Temp In C	Temp In C	Temp In C	Temp In C	Temp In C	Temp In C
	Received on Ice	Received on Ice	Received on Ice	Received on Ice	Received on Ice	Received on Ice	Received on Ice	Received on Ice	Received on Ice	Received on Ice	Received on Ice	Received on Ice	Received on Ice	Received on Ice	Received on Ice	Received on Ice
	Custody Sealed Cooler	Custody Sealed Cooler	Custody Sealed Cooler	Custody Sealed Cooler	Custody Sealed Cooler	Custody Sealed Cooler	Custody Sealed Cooler	Custody Sealed Cooler	Custody Sealed Cooler	Custody Sealed Cooler	Custody Sealed Cooler	Custody Sealed Cooler	Custody Sealed Cooler	Custody Sealed Cooler	Custody Sealed Cooler	Custody Sealed Cooler
	Samples Intact	Samples Intact	Samples Intact	Samples Intact	Samples Intact	Samples Intact	Samples Intact	Samples Intact	Samples Intact	Samples Intact	Samples Intact	Samples Intact	Samples Intact	Samples Intact	Samples Intact	Samples Intact

Comments:
 Detection levels need to meet applicable MDEQ PWM Non-Residential Indoor Air Screening Levels (attached).
 Level II Deliverables ORIGINAL
 Even container #2703 arrived with no vacuum on gauge, return for no charge

SAMPLER NAME AND SIGNATURE
 PRINT Name of SAMPLER: Jeff Nicholson
 SIGNATURE of SAMPLER: Jeff Nicholson
 DATE Signed (MM/DD/YY) 09-15-21

WO#: 92561801

PM: RNB Due Date: 09/27/21
CLIENT: 92-Shield

AIR: CHAIN-OF-CUSTODY / Analytical Request Document

The Chain-of-Custody is a LEGAL DOCUMENT. All relevant fields must be completed accurately.

Section A Required Client Information: Company: <u>Shield Engineering</u> Address: Email To: Phone: Fax: Requested Due Date/TAT: <u>ASAP w/out Extra Charge</u>		Section B Required Project Information: Report To: Copy To: Purchase Order No.: Project Name: Project Number: <u>42504</u>		Section C Invoice Information: Attention: <u>Wes Barfield</u> Company Name: Address: Pace Quote Reference: Pace Project Manager/Sales Rep. Pace Profile #: <u>42504</u>		Program: <input type="checkbox"/> UST Superfund <input type="checkbox"/> Emissions <input type="checkbox"/> Clean Air Act <input type="checkbox"/> Voluntary Clean Up <input type="checkbox"/> Dry Clean <input checked="" type="checkbox"/> RCRA <input type="checkbox"/> Other Location of Sampling by State: <u>NC</u> Reporting Units: <input checked="" type="checkbox"/> ug/m ³ <input type="checkbox"/> mg/m ³ <input type="checkbox"/> PPMV <input type="checkbox"/> PPMV Report Level: II <input checked="" type="checkbox"/> III <input type="checkbox"/> IV <input type="checkbox"/> Other		Page: <u>2</u> of <u>2</u>			
Section D Required Client Information AIR SAMPLE ID Sample IDs MUST BE UNIQUE <u>IA-11</u> <u>IA-4</u> <u>IA-9</u> <u>BackGround</u>		Valid Media Codes MEDIA CODE Tedlar Bag TB 1 Liter Summa Can 1LC 6 Liter Summa Can 6LC Low Volume Puff LVP High Volume Puff HVP Other PM10		COLLECTED PID Reading (Client only) MEDIA CODE <u>6LC</u> DATE TIME DATE TIME <u>9-14-21 10:13 9-14-21 18:45</u> <u>9-14-21 9:16 9-14-21 16:16</u> <u>9-14-21 10:03 9-14-21 18:04</u> <u>9-14-21 10:24 9-14-21 18:08</u>		Canister Pressure (Initial Field - In Hg) Canister Pressure (Final Field - In Hg) <u>-34 -3</u> <u>-29 -6</u> <u>-28 -18</u> <u>-30 -5</u>		Summa Can Number Flow Control Number <u>3475</u> <u>0872</u> <u>2049</u> <u>0712</u> <u>0857</u> <u>2069</u> <u>3891</u> <u>1429</u>		Method: <input type="checkbox"/> PM10 <input type="checkbox"/> 3C - Fixed Gas (%) <input type="checkbox"/> TO-3 BTEX <input type="checkbox"/> TO-3M (Methane) <input type="checkbox"/> TO-14 <input type="checkbox"/> TO-15 Full List VOCs <input type="checkbox"/> TO-15 Short List BTEX <input type="checkbox"/> TO-15 Short List Chlorinated <input type="checkbox"/> TO-15 Short List (Other) <u>013</u> <u>014</u> <u>015</u> <u>016</u>	
Comments: Detection levels need to meet applicable NCDER DWM Non-Residential Indoor Air Screening Levels - See Attached Level II Deliverables		RELINQUISHED BY / AFFILIATION <u>Jeff Nicholson/Shield</u> DATE <u>9-15-21</u>		ACCEPTED BY / AFFILIATION <u>J. M. Barber</u> DATE <u>9-16-21</u>		TIME <u>12:47</u> <u>13:15</u>		SAMPLE CONDITIONS Temp in C Received on Ice Custody Sealed Cooler Samples Intact			
SAMPALER NAME AND SIGNATURE PRINT Name of SAMPALER: <u>Jeff Nicholson</u> SIGNATURE of SAMPALER: <u>Jeff Nicholson</u> DATE Signed (MM/DD/YY): <u>09-15-21</u>		ORIGINAL									

North Carolina Department of Environmental Quality
 Non-Residential Vapor Intrusion Screening Levels
 Not To Be Used for Structures Frequented or Occupied By Children
 June 2021
 (based on May 2021 USEPA Regional Screening Tables)

Note: If a screening level is exceeded and/or more than five compounds with non-cancer effects are present risk should be evaluated using NCEQ Risk Calculator.

CAS #	Chemical Name:	Screening Level (GWSL) ug/L	Sub-slab and Exterior Soil Gas Screening Level (SGSL) ug/m ³	Indoor Air Screening Level (IASL) ug/m ³	Does Chemical Have a Non-Cancer Effect?
30560-19-1	Acaphate	THQ=0.2	THQ=0.2	THQ=0.2	
75-07-0	Acetaldehyde	2.0E+03	5.6E+02	5.6E+00	Yes
34256-82-1	Acetochlor				
67-64-1	Acetone	1.9E+07	2.7E+06	2.7E+04	Yes
75-86-5	Acetone Cyanohydrin	Not Volatile	Not Volatile	1.8E+00	Yes
75-05-8	Acetonitrile	3.7E+04	5.3E+03	5.3E+01	Yes
98-86-2	Acetophenone				
53-96-3	Acetylaminofluorene, 2-	Not Volatile	Not Volatile	9.4E-03	
107-02-8	Acrolein	3.5E+00	1.8E+00	1.8E-02	Yes
79-06-1	Acrylamide	Not Volatile	Not Volatile	1.2E-01	Yes
79-10-7	Acrylic Acid	5.8E+04	8.8E+01	8.8E-01	Yes
107-13-1	Acrylonitrile	3.2E+01	1.8E+01	1.8E-01	Yes
111-69-3	Adiponitrile	Not Volatile	Not Volatile	5.3E+00	Yes
15972-60-8	Alachlor				
116-06-3	Aldicarb				
1646-88-4	Aldicarb Sulfone				
1646-87-3	Aldicarb sulfoxide				
309-00-2	Aldrin	1.4E+00	2.5E-01	2.5E-03	Yes
107-18-6	Allyl Alcohol	4.3E+02	8.8E+00	8.8E-02	Yes
107-05-1	Allyl Chloride	1.9E+00	8.8E+01	8.8E-01	Yes
7429-90-5	Aluminum	Not Volatile	Not Volatile	4.4E+00	Yes
20859-73-8	Aluminum Phosphide				
834-12-8	Ametyn				
92-67-1	Aminobiphenyl, 4-	Not Volatile	Not Volatile	2.0E-03	
591-27-5	Aminophenol, m-				
95-55-6	Aminophenol, o-				
123-30-8	Aminophenol, p-				
33089-61-1	Amirtraz				
7664-41-7	Ammonia	6.7E+05	4.4E+04	4.4E+02	Yes
131-74-8	Ammonium Picrate				
7773-06-0	Ammonium Sulfamate				
75-85-4	Amyl Alcohol, tert-	4.7E+03	2.6E+02	2.6E+00	Yes
62-53-3	Aniline	Not Volatile	Not Volatile	8.8E-01	Yes
84-65-1	Anthraquinone, 9,10-				
7440-36-0	Antimony (metalloid)	Not Volatile	Not Volatile	2.6E-01	Yes
1314-60-9	Antimony Pentoxide				
1332-81-6	Antimony Tetroxide				
1309-64-4	Antimony Trioxide	Not Volatile	Not Volatile	1.8E-01	Yes
7440-38-2	Arsenic, Inorganic	Not Volatile	Not Volatile	2.9E-03	Yes
7784-42-1	Arsine	Not Volatile	Not Volatile	4.4E-02	Yes
1332-21-4	Asbestos (units in fibers)				
3337-71-1	Asulam				

North Carolina Department of Environmental Quality
 Non-Residential Vapor Intrusion Screening Levels
 Not To Be Used for Structures Frequented or Occupied by Children
 June 2021
 (based on May 2021 USEPA Regional Screening Tables)

Note: If a screening level is exceeded and/or more than five compounds with non-cancer effects are present risk should be evaluated using NCEQ Risk Calculator.

CAS #	Chemical Name:	Ground Water Screening Level (GWSL) ug/L	Soil Gas Screening Level (SGL) ug/m ³	Sub-slab and Exterior	Indoor Air Screening Level (IASL) ug/m ³	Does Chemical Have a Non-Cancer Effect?
2104-96-3	Bromophos					
106-94-5	Bromopropane, 1-	2.9E+02	8.8E+03	N	8.8E+01	Yes
1689-84-5	Bromoxynil					
1689-99-2	Bromoxynil Octanoate					
106-99-0	Butadiene, 1,3-	1.4E-01	4.1E+01	C	4.1E-01	Yes
94-82-6	Butanoic acid, 4-(2,4-dichlorophenoxy)-					
71-36-3	Butanol, N-					
78-92-2	Butyl alcohol, sec-	7.1E+07	2.6E+06	N	2.6E+04	Yes
2008-41-5	Butylate					
25013-16-5	Butylated hydroxyanisole	Not Volatile	Not Volatile		2.2E+02	C
128-37-0	Butylated hydroxytoluene					
104-51-8	Butylbenzene, n-					
135-98-8	Butylbenzene, sec-					
98-06-6	Butylbenzene, tert-					
75-60-5	Caodylic Acid					
7440-43-9	Cadmium (Diel)	Not Volatile	Not Volatile		6.8E-03	C
7440-43-9	Cadmium (Water)	Not Volatile	Not Volatile		6.8E-03	C
105-60-2	Caprolactam	Not Volatile	Not Volatile		1.9E+00	N
2425-06-1	Captafol	Not Volatile	Not Volatile		2.9E-01	C
133-06-2	Captan	Not Volatile	Not Volatile		1.9E+01	C
63-25-2	Carbaryl					
1563-66-2	Carbofuran					
75-15-0	Carbon Disulfide	1.0E+03	6.1E+04	N	6.1E+02	N
56-23-5	Carbon Tetrachloride	1.8E+00	2.0E+02	C	2.0E+00	C
463-58-1	Carbonyl Sulfide	3.5E+00	8.8E+03	N	8.8E+01	N
55285-14-8	Carbosulfan					
5234-68-4	Carboxin					
1306-38-3	Ceric oxide	Not Volatile	Not Volatile		7.9E-01	N
302-17-0	Chloral Hydrate					
133-90-4	Chloramphenicol					
E701235	Chloramines, Organic					
118-75-2	Chloranil					
5103-71-9	Chlordane (alpha)					
5103-74-2	Chlordane (gamma)					
12789-03-6	Chlordane (technical mixture)	6.2E+01	1.2E+01	C	1.2E-01	C
143-50-0	Chlordecone (Kepone)	Not Volatile	Not Volatile		2.7E-03	C
470-90-6	Chlorovinphos					
90982-32-4	Chlorimuron, Ethyl-					
7782-50-5	Chlorine	2.7E-01	1.3E+01	N	1.3E-01	N
10049-04-4	Chlorine Dioxide	1.1E-01	1.8E+01	N	1.8E-01	N
7758-19-2	Chlorite (Sodium Salt)					
75-68-3	Chloro-1,1-difluoroethane, 1-	1.8E+04	4.4E+06	N	4.4E+04	N

North Carolina Department of Environmental Quality
 Non-Residential Vapor Intrusion Screening Levels
 Not To Be Used for Structures Frequented or Occupied By Children
 June 2021
 (based on May 2021 USEPA Regional Screening Tables)

Note: If a screening level is exceeded and/or more than five compounds with non-cancer effects are present risk should be evaluated using NCDEQ Risk Calculator.

CAS #	Chemical Name:	Ground Water Screening Level (GWSL) ug/L	Sub-slab and Exterior Soil Gas Screening Level (SGSL) ug/m ³	Indoor Air Screening Level (IASL) ug/m ³	Does Chemical Have a Non-Cancer Effect?
59-50-7	Cresol, p-chloro-m-				
1319-77-3	Cresols	Not Volatile	Not Volatile	5.3E+02	N
123-73-9	Crotonaldehyde, trans-				
98-82-8	Cumene	7.5E+02	3.5E+04	3.5E+02	N
135-20-6	Cupferron	Not Volatile	Not Volatile	1.9E-01	C
21725-46-2	Cyanazine				
	Cyanides				
592-01-8	-Calcium Cyanide				
544-92-3	-Copper Cyanide				
57-12-5	-Cyanide (CN-)	1.7E+02	7.0E+01	7.0E-01	N
460-19-5	-Cyanogen				
506-68-3	-Cyanogen Bromide				
506-77-4	-Cyanogen Chloride				
74-90-8	-Hydrogen Cyanide	1.3E+02	7.0E+01	7.0E-01	N
151-50-8	-Potassium Cyanide				
506-61-6	-Potassium Silver Cyanide				
506-64-9	-Silver Cyanide				
143-33-9	-Sodium Cyanide				
E1790664	-Thiocyanates				
463-56-9	~Thiocyanic Acid				
557-21-1	-Zinc Cyanide				
110-82-7	Cyclohexane	8.6E+02	5.3E+05	5.3E+03	N
87-84-3	Cyclohexane, 1,2,3,4,5-pentabromo-6-chloro-				
108-94-1	Cyclohexanone	1.7E+06	6.1E+04	6.1E+02	N
110-83-8	Cyclohexene	4.7E+02	8.8E+04	8.8E+02	N
108-91-8	Cyclohexylamine				
68359-37-5	Cyluthrin				
68085-85-8	Cyhalothrin				
52315-07-8	Cypermethrin				
66215-27-8	Cymazine				
72-54-8	DDD, p,p- (DD)	Not Volatile	Not Volatile	1.8E-01	C
72-55-9	DDE, p,p-	7.4E+01	1.3E+01	1.3E-01	C
50-29-3	DDT	Not Volatile	Not Volatile	1.3E-01	C
75-99-0	Dalapon				
1596-84-5	Daminozide	Not Volatile	Not Volatile	2.4E+00	C
1163-19-5	Decabromodiphenyl ether, 2,2',3,3',4,4',5,5',6,6'				
8065-48-3	Demeton				
103-23-1	D(2-ethylhexyl)adipate				
2303-16-4	Diallate				
333-41-5	Diazinon				

North Carolina Department of Environmental Quality
 Non-Residential Vapor Intrusion Screening Levels
 Not To Be Used for Structures Frequented or Occupied By Children
 June 2021
 (based on May 2021 USEPA Regional Screening Tables)

Note: If a screening level is exceeded and/or more than five compounds with non-cancer effects are present risk should be evaluated using NCEQ Risk Calculator.

CAS #	Chemical Name:	Ground Water Screening Level (GWSL) ug/L	Sub-slab and Exterior Soil Gas Screening Level (SGSL) ug/m ³	Indoor Air Screening Level (IASL) ug/m ³	Does Chemical Have a Non-Cancer Effect?
35367-38-5	Dilubenzuron				
75-37-6	Difluoroethane, 1,1-	4.2E+04	3.5E+06	3.5E+04	N
420-45-1	Difluoropropane, 2,2-	1.3E+03	2.6E+06	2.6E+04	N
94-58-6	Dihydrosafrole	1.9E+03	9.4E+01	9.4E-01	C
108-20-3	Diisopropyl Ether	5.9E+03	6.1E+04	6.1E+02	N
1445-75-6	Diisopropyl Methylphosphonate				
55290-64-7	Dimethipin				
60-51-5	Dimethoate				
119-90-4	Dimethoxybenzidine, 3,3'-				
756-79-6	Dimethyl methylphosphonate				
60-11-7	Dimethylamino azobenzene [p-]	Not Volatile	Not Volatile	9.4E-03	C
21436-96-4	Dimethylaniline HCl, 2,4-				
95-68-1	Dimethylaniline, 2,4-				
121-69-7	Dimethylaniline, N,N-				
119-93-7	Dimethylbenzidine, 3,3'-				
68-12-2	Dimethylformamide	8.7E+06	2.6E+03	2.6E+01	N
57-14-7	Dimethylhydrazine, 1,1-	3.3E+00	1.8E-01	1.8E-03	N
540-73-8	Dimethylhydrazine, 1,2-	2.7E+01	7.7E-03	7.7E-05	C
105-67-9	Dimethylphenol, 2,4-				
576-26-1	Dimethylphenol, 2,6-				
95-65-8	Dimethylphenol, 3,4-				
513-37-1	Dimethylvinylchloride	1.9E+01	9.4E+01	9.4E-01	C
534-52-1	Dinitro-o-cresol, 4,6-				
131-89-5	Dinitro-o-cyclohexyl Phenol, 4,6-				
528-29-0	Dinitrobenzene, 1,2-				
99-65-0	Dinitrobenzene, 1,3-				
100-25-4	Dinitrobenzene, 1,4-				
51-28-5	Dinitrophenol, 2,4-				
E1615210	Dinitrofluorene Mixture, 2,4/2,6-				
121-14-2	Dinitrofluorene, 2,4-	Not Volatile	Not Volatile	1.4E-01	C
606-20-2	Dinitrofluorene, 2,6-				
35572-78-2	Dinitrofluorene, 2-Amino-4,6-				
19406-51-0	Dinitrofluorene, 4-Amino-2,6-				
25321-14-6	Dinitrofluorene, Technical grade				
88-85-7	Dinoseb				
123-91-1	Dioxane, 1,4-	1.2E+04	2.5E+02	2.5E+00	C
	Dioxins				
34465-46-8	~Hexachlorodibenzo-p-dioxin, Mixture	Not Volatile	Not Volatile	9.4E-06	C
1746-01-6	~TCDD, 2,3,7,8-	1.6E-04	3.2E-05	3.2E-07	C
957-51-7	Diphenamid				
101-84-8	Diphenyl Ether	3.1E+01	3.5E+01	3.5E-01	N
127-63-9	Diphenyl Sulfone				

North Carolina Department of Environmental Quality
 Non-Residential Vapor Intrusion Screening Levels
 Not To Be Used for Structures Frequented or Occupied By Children
 June 2021
 (based on May 2021 USEPA Regional Screening Tables)

Note: If a screening level is exceeded and/or more than five compounds with non-cancer effects are present risk should be evaluated using NCDDEQ Risk Calculator.

CAS #	Chemical Name:	Ground Water Screening Level (GWSL) ug/L	Sub-slab and Exterior Soil Gas Screening Level (SGSL) ug/m ³	Indoor Air Screening Level (IASL) ug/m ³	Does Chemical Have a Non-Cancer Effect?
7782-41-4	Fluorine (Soluble Fluoride)	Not Volatile	Not Volatile	1.1E+01	Yes
59756-60-4	Fluindone				
56425-91-3	Flupirtinidol				
85509-19-9	Flustiazole				
66332-96-5	Flutolanil				
69409-94-5	Fluvalinate				
133-07-3	Folpet				
72178-02-0	Fomesafen				
944-22-9	Fonofos				
50-00-0	Formaldehyde	6.8E+04	9.4E+01	9.4E-01	C
64-18-6	Formic Acid	3.8E+04	2.6E+01	2.6E-01	N
39148-24-8	Fosetyl-AL				
	Furans				
132-64-9	~Dibenzofuran				
110-00-9	~Furan				
109-99-9	~Tetrahydrofuran	6.1E+05	1.8E+05	1.8E+03	N
67-45-8	Furazolidone				
98-01-1	Furfural	2.8E+05	4.4E+03	4.4E+01	N
531-82-8	Furum	Not Volatile	Not Volatile	2.9E-02	C
60568-05-0	Furmecycloox	Not Volatile	Not Volatile	1.4E+00	C
77182-82-2	Glufosinate, Ammonium				
111-30-8	Glutaraldehyde	Not Volatile	Not Volatile	7.0E-02	N
765-34-4	Glycidaldehyde	4.2E+04	8.8E+01	8.8E-01	N
1071-83-6	Glyphosate				
113-00-8	Guanidine				
50-01-1	Guanidine Chloride				
506-93-4	Guanidine Nitrate				
69806-40-2	Haloxypol, Methyl				
76-44-8	Hepachlor	7.8E-01	9.4E-01	9.4E-03	C
1024-57-3	Hepachlor Epoxide	5.5E+00	4.7E-01	4.7E-03	C
111-71-7	Hepitana, n-	2.4E+02	2.6E+02	2.6E+00	N
142-82-5	Hepitane, N-	4.3E+00	3.5E+04	3.5E+02	N
87-82-1	Hexabromobenzene				
68631-49-2	Hexabromodiphenyl ether, 2,2',4,4',5,5'-(BDE-153)				
118-74-1	Hexachlorobenzene	3.8E-01	2.7E+00	2.7E-02	C
87-68-3	Hexachlorobutadiene	1.3E+00	5.6E+01	5.6E-01	C
319-84-6	Hexachlorocyclohexane, Alpha-	Not Volatile	Not Volatile	6.8E-03	C
319-85-7	Hexachlorocyclohexane, Beta-	Not Volatile	Not Volatile	2.3E-02	C
58-89-9	Hexachlorocyclohexane, Gamma-(Lindane)	Not Volatile	Not Volatile	4.0E-02	C
608-73-1	Hexachlorocyclohexane, Technical	Not Volatile	Not Volatile	2.4E-02	C

North Carolina Department of Environmental Quality
 Non-Residential Vapor Intrusion Screening Levels
 Not To Be Used for Structures Frequented or Occupied by Children
 June 2021
 (based on May 2021 USEPA Regional Screening Tables)

Note: If a screening level is exceeded and/or more than five compounds with non-cancer effects are present risk should be evaluated using NCEQ Risk Calculator.

CAS #	Chemical Name:	Ground Water Screening Level (GWSL) ug/L	Sub-slab and Exterior Soil Gas Screening Level (SGSL) ug/m ³	Indoor Air Screening Level (IASL) ug/m ³	Does Chemical Have a Non-Cancer Effect?
10025-84-0	Lanthanum Chloride Hexahydrate	THQ=0.2 TCR=1.0E-06	THQ=0.2 TCR=1.0E-06	THQ=0.2 TCR=1.0E-06	
10099-58-8	Lanthanum Chloride, Anhydrous				
10277-43-7	Lanthanum Nitrate Hexahydrate				
7446-27-7	Lead Compounds	Not Volatile	Not Volatile	1.0E+00	C
301-04-2	~Lead Phosphate	Not Volatile	Not Volatile	1.5E-01	C
7439-92-1	~Lead acetate	Not Volatile	Not Volatile	1.5E-01	C
1335-32-6	~Lead subacetate	Not Volatile	Not Volatile	1.1E+00	C
78-00-2	~Tetraethyl Lead	Not Volatile	Not Volatile		
541-25-3	Lewisite				
330-55-2	Linuron				
7439-93-2	Lithium				
94-74-6	MCPA				
94-81-5	MCPB				
93-65-2	MCPP				
121-75-5	Malathion				
108-31-6	Maleic Anhydride	Not Volatile	Not Volatile	6.1E-01	N
123-33-1	Maleic Hydrazide				
109-77-3	Malononitrile				
8018-01-7	Mancozeb				
12427-38-2	Maneb				
7439-96-5	Manganese (Diet)	Not Volatile	Not Volatile	4.4E-02	N
7439-96-5	Manganese (Non-diet)	Not Volatile	Not Volatile	4.4E-02	N
950-10-7	Mepfosfolan				
24307-26-4	Mepiquat Chloride				
149-30-4	Mercapto benzothiazole, 2-				
	Mercury Compounds				
7487-94-7	~Mercury Chloride (and other Mercury salts)	Not Volatile	Not Volatile	2.6E-01	N
7439-97-6	~Mercury (elemental)	7.5E-01	2.6E+01	2.6E-01	N
22967-92-6	~Methyl Mercury				
62-38-4	~Phenylmercuric Acetate				
150-50-5	Merphos				
57837-19-1	Metaxyl				
126-98-7	Methacrylonitrile	2.6E+03	2.6E+03	2.6E+01	N
10265-92-6	Methamidophos				
67-56-1	Methanol	9.4E+07	1.8E+06	1.8E+04	N
950-37-8	Methidathion				
16752-77-5	Methomyl				
99-59-2	Methoxy-5-nitroaniline, 2-				
72-43-5	Methoxychlor				
110-49-6	Methoxyethanol Acetate, 2-	6.9E+04	8.8E+01	8.8E-01	N

North Carolina Department of Environmental Quality
 Non-Residential Vapor Intrusion Screening Levels
 Not To Be Used for Structures Frequented or Occupied by Children
 June 2021
 (based on May 2021 USEPA Regional Screening Tables)

Note: If a screening level is exceeded and/or more than five compounds with non-cancer effects are present risk should be evaluated using NCDEQ Risk Calculator.

CAS #	Chemical Name:	Ground Water Screening Level (GWSL) ug/L	Sub-slab and Exterior Soil Gas Screening Level (SGSL) ug/m ³	Indoor Air Screening Level (IASL) ug/m ³	Does Chemical Have a Non-Cancer Effect?
100-61-8	Monomethylamine				
88671-89-0	Myclobutanol				
74-31-7	N,N'-Diphenyl-1,4-benzenediamine				
300-76-5	Naled				
64742-95-6	Naphtha, High Flash Aromatic (HFAN)	4.9E+03	8.8E+03	8.8E+01	Yes
91-59-8	Naphthylamine, 2-				
15299-99-7	Napamide				
373-02-4	Nickel Acetate	Not Volatile	Not Volatile	1.2E-02	Yes
3333-67-3	Nickel Carbonate	Not Volatile	Not Volatile	1.2E-02	Yes
13463-39-3	Nickel Carbonyl	6.0E-04	1.2E+00	1.2E-02	Yes
12054-48-7	Nickel Hydroxide	Not Volatile	Not Volatile	1.2E-02	Yes
1313-99-1	Nickel Oxide	Not Volatile	Not Volatile	1.8E-02	Yes
E71532	Nickel Refinery Dust	Not Volatile	Not Volatile	1.2E-02	Yes
7440-02-0	Nickel Soluble Salts	Not Volatile	Not Volatile	4.7E-02	Yes
12035-72-2	Nickel Sulfide	Not Volatile	Not Volatile	1.2E-02	Yes
1271-28-9	Nickelocene	Not Volatile	Not Volatile	1.2E-02	Yes
14797-55-8	Nitrate				
E701177	Nitrate + Nitrite (as N)				
14797-65-0	Nitrite				
88-74-4	Nitroamine, 2-	Not Volatile	Not Volatile	4.4E-02	Yes
100-01-6	Nitroamine, 4-	Not Volatile	Not Volatile	5.3E+00	Yes
98-95-3	Nitrobenzene	3.1E+02	3.1E+01	3.1E-01	Yes
9004-70-0	Nitrocellulose				
67-20-9	Nitrofurantoin				
59-87-0	Nitrofurazone	Not Volatile	Not Volatile	3.3E-02	C
55-63-0	Nitroglycerin				
556-88-7	Nitroguanidine				
75-52-5	Nitromethane	1.2E+03	1.4E+02	1.4E+00	Yes
79-46-9	Nitropropane, 2-	4.3E+00	2.1E+00	2.1E-02	Yes
759-73-9	Nitroso-N-ethylurea, N-	Not Volatile	Not Volatile	1.6E-03	C
684-93-5	Nitroso-N-methylurea, N-	Not Volatile	Not Volatile	3.6E-04	C
924-16-3	Nitroso-di-N-butylamine, N-	1.4E+01	7.7E-01	7.7E-03	C
621-64-7	Nitroso-di-N-propylamine, N-	Not Volatile	Not Volatile	6.1E-03	C
1116-54-7	Nitrosodichloroamine, N-	Not Volatile	Not Volatile	1.5E-02	C
55-18-5	Nitrosodimethylamine, N-	Not Volatile	Not Volatile	2.9E-04	C
62-75-9	Nitrosodimethylamine, N-	1.2E+01	8.8E-02	8.8E-04	Yes
86-30-6	Nitrosodiphenylamine, N-	Not Volatile	Not Volatile	4.7E+00	C
10595-95-6	Nitrosomethylamine, N-	3.3E+01	1.9E-01	1.9E-03	C
59-89-2	Nitrosomorpholine [N-]	Not Volatile	Not Volatile	6.5E-03	C
100-75-4	Nitrosopiperidine [N-]	Not Volatile	Not Volatile	4.5E-03	C
930-55-2	Nitrosopyrrolidine, N-	Not Volatile	Not Volatile	2.0E-02	C
99-08-1	Nitrotoluene, m-				

North Carolina Department of Environmental Quality
Non-Residential Vapor Intrusion Screening Levels
Not To Be Used for Structures Frequented or Occupied by Children
 June 2021
 (based on May 2021 USEPA Regional Screening Tables)

Note: If a screening level is exceeded and/or more than five compounds with non-cancer effects are present risk should be evaluated using NCEQ Risk Calculator.

CAS #	Chemical Name:	Ground Water Screening Level (GWSL) ug/L (THQ=0.2 TCR=1.0E-06)	Sub-slab and Exterior Soil Gas Screening Level (SGSL) ug/m ³ (THQ=0.2 TCR=1.0E-06)	Indoor Air Screening Level (IASL) ug/m ³ (THQ=0.2 TCR=1.0E-06)	Does Chemical Have a Non-Cancer Effect?
103-72-0	Phenyl Isothiocyanate				
108-45-2	Phenylethylamine, m-				
95-54-5	Phenylethylamine, o-				
106-50-3	Phenylethylamine, p-				
90-43-7	Phenylphenol, 2-				
298-02-2	Phorate				
75-44-5	Phosgene	3.8E-01	2.6E+01	2.6E-01	N
732-11-6	Phosmet				Yes
	Phosphates, Inorganic				
13776-88-0	~Aluminum metaphosphate				
68333-79-9	~Ammonium polyphosphate				
7790-76-3	~Calcium pyrophosphate				
7783-28-0	~Diammonium phosphate				
7757-93-9	~Dicalcium phosphate				
7782-75-4	~Dimagnesium phosphate				
7758-11-4	~Dipotassium phosphate				
7558-79-4	~Disodium phosphate				
13530-50-2	~Monoaluminum phosphate				
7722-76-1	~Monoammonium phosphate				
7758-23-8	~Monocalcium phosphate				
7757-86-0	~Monomagnesium phosphate				
7778-77-0	~Monopotassium phosphate				
7558-80-7	~Monosodium phosphate				
8017-16-1	~Polyphosphoric acid				
13845-36-8	~Potassium triphosphate				
7758-16-9	~Sodium acid pyrophosphate				
7785-88-8	~Sodium aluminum phosphate (acid)				
10279-59-1	~Sodium aluminum phosphate (anhydrous)				
10305-76-7	~Sodium aluminum phosphate (tetrahydrate)				
10124-56-8	~Sodium hexametaphosphate				
68915-31-1	~Sodium polyphosphate				
7785-84-4	~Sodium trimetaphosphate				
7758-29-4	~Sodium tripolyphosphate				
7320-34-5	~Tetrapotassium phosphate				
7722-88-5	~Tetrasodium pyrophosphate				
15136-87-5	~Trialuminum sodium tetra decahydrogenooctaoctaphosphate (dihydrate)				
7758-87-4	~Tricalcium phosphate				
7757-87-1	~Trimagnesium phosphate				

North Carolina Department of Environmental Quality
 Non-Residential Vapor Intrusion Screening Levels
 Not to Be Used for Structures Frequented or Occupied by Children
 June 2021
 (based on May 2021 USEPA Regional Screening Tables)

Note: If a screening level is exceeded and/or more than five compounds with non-cancer effects are present risk should be evaluated using NCEQ Risk Calculator.

CAS #	Chemical Name:	Ground Water Screening Level (GWSL) ug/L	Sub-slab and Exterior Soil Gas Screening Level (SGSL) ug/m ³	Indoor Air Screening Level (IASL) ug/m ³	Does Chemical Have a Non-Cancer Effect?
31508-00-6	~Pentachlorobiphenyl, 2,3,4,4',5-(PCB 118)	9.5E-01	1.1E+00	1.1E-02	Yes
32598-14-4	~Pentachlorobiphenyl, 2,3,3',4,4'-(PCB 105)	9.6E-01	1.1E+00	1.1E-02	Yes
74472-37-0	~Pentachlorobiphenyl, 2,3,4,4',5-(PCB 114)	3.0E+00	1.1E+00	1.1E-02	Yes
57465-28-8	~Pentachlorobiphenyl, 3,3',4,4',5-(PCB 126)	4.2E-04	3.2E-04	3.2E-06	Yes
1336-36-3	~Polychlorinated Biphenyls (high risk)	1.3E+00	2.2E+00	2.2E-02	C
1336-36-3	~Polychlorinated Biphenyls (low risk)	7.2E+00	1.2E+01	1.2E-01	C
1336-36-3	~Polychlorinated Biphenyls (lowest risk)	3.6E+01	6.1E+01	6.1E-01	C
32598-13-3	~Tetrachlorobiphenyl, 3,3',4,4'-(PCB 77)	Not Volatile	Not Volatile	3.2E-03	C
70362-50-4	~Tetrachlorobiphenyl, 3,4,4',5-(PCB 81)	1.2E-01	1.1E-01	1.1E-03	C
9016-87-9	Polymeric Methylene Diphenyl Diisocyanate (PMDI)	Not Volatile	Not Volatile	5.3E-01	N
83-32-9	~Acenaphthene				
	(PAHs)				
	Polynuclear Aromatic Hydrocarbons				
56-55-3	~Benz[a]anthracene	4.2E+02	2.0E+01	2.0E-01	C
205-82-3	~Benzo[ghi]perylene	Not Volatile	Not Volatile	1.1E-01	C
50-32-8	~Benzo[a]pyrene	Not Volatile	Not Volatile	1.8E-03	N
205-99-2	~Benzo[b]fluoranthene	Not Volatile	Not Volatile	2.0E-01	C
207-08-9	~Benzo[k]fluoranthene	Not Volatile	Not Volatile	2.0E+00	C
91-58-7	~Chloronaphthalene, Beta	Not Volatile	Not Volatile		
218-01-9	~Chrysene	Not Volatile	Not Volatile	2.0E+01	C
53-70-3	~Dibenz[a,h]anthracene	Not Volatile	Not Volatile	2.0E-02	C
192-65-4	~Dibenz[ac]pyrene	Not Volatile	Not Volatile	1.1E-02	C
57-97-6	~Dimethylbenz(a)anthracene, 7,12-	Not Volatile	Not Volatile	1.7E-04	C
206-44-0	~Fluorene				
86-73-7	~Fluorene				
193-39-5	~Indeno[1,2,3-cd]pyrene	Not Volatile	Not Volatile	2.0E-01	C
90-12-0	~Methylnaphthalene, 1-				
91-57-6	~Methylnaphthalene, 2-				
91-20-3	~Naphthalene	2.0E+01	3.6E+01	3.6E-01	C
5783-92-4	~Nitropyrene, 4-	Not Volatile	Not Volatile	1.1E-01	C
129-00-0	~Pyrene				
29420-49-3	Potassium Perfluorobutane Sulfonate				
6747-09-5	Prochloraz				
26399-36-0	Profluralin				

North Carolina Department of Environmental Quality
 Non-Residential Vapor Intrusion Screening Levels
 Not To Be Used for Structures Frequented or Occupied by Children
 June 2021
 (based on May 2021 USEPA Regional Screening Tables)

Note: If a screening level is exceeded and/or more than five compounds with non-cancer effects are present risk should be evaluated using NCDEQ Risk Calculator.

CAS #	Chemical Name:	Ground Water Screening Level (GWSL) ug/L	Sub-slab and Exterior Soil Gas Screening Level (SGSL) ug/m ³	Indoor Air Screening Level (IASL) ug/m ³	Does Chemical Have a Non-Cancer Effect?
961-11-5	Stirofos (Tetrachlorovinphos)				
7440-24-6	Strontium, Stable				
57-24-9	Styrene				
100-42-5	Styrene	7.8E+03	8.8E+04	8.8E+02	Yes
57964-39-3	Styrene-Acrylonitrile (SAN) Trimer (THNA isomer)				
57964-40-6	Styrene-Acrylonitrile (SAN) Trimer (THNP isomer)				
126-33-0	Sulfolane	Not Volatile	Not Volatile	1.8E+00	Yes
80-07-9	Sulfonylbis(4-chlorobenzene), 1,1'-				
7446-11-9	Sulfur Trioxide			8.8E-01	Yes
7664-93-9	Sulfuric Acid	Not Volatile	Not Volatile	8.8E-01	Yes
140-57-8	Sulfurous acid, 2-chloroethyl 2-(4-(1,1-dimethylethyl)phenoxy)-1-methylethyl ester	Not Volatile	Not Volatile	1.7E+00	C
21564-17-0	TCTMB				
34014-18-1	Tebuhuron				
3383-96-8	Temphos				
5902-51-2	Terbacil				
13071-79-9	Terbufos				
886-50-0	Tertbutyn				
540-88-5	Tert-Butyl Acetate	2.7E+02	9.4E+02	9.4E+00	C
5436-43-1	Tetrabromodiphenyl ether, 2,2',4,4'-(BDE-47)				
95-94-3	Tetrachlorobenzene, 1,2,4,5-				
690-20-6	Tetrachloroethane, 1,1,1,2-	1.6E+01	1.7E+02	1.7E+00	C
79-34-5	Tetrachloroethane, 1,1,2,2-	1.4E+01	2.1E+01	2.1E-01	C
127-18-4	Tetrachloroethylene	4.8E+01	3.5E+03	3.5E+01	Yes
58-90-2	Tetrachlorophenol, 2,3,4,6-				
5216-25-1	Tetrachlorotoluene, p-alpha, alpha				
3689-24-5	Tetraethyl Dithiopyrophosphate				
811-97-2	Tetrafluoroethane, 1,1,1,2-	3.4E+04	7.0E+06	7.0E+04	Yes
16853-36-4	Tetramethylphosphoramide, -N,N,N',N'' (TMPA)				
479-45-8	Tetral (Trinitrophenylmethyl)nitramine)				
1314-32-5	Thallic Oxide				
10102-45-1	Thallium (I) Nitrate				
7440-28-0	Thallium (Soluble Salts)				
563-68-8	Thallium Acetate				
6533-73-9	Thallium Carbonate				
7791-12-0	Thallium Chloride				

North Carolina Department of Environmental Quality
 Non-Residential Vapor Intrusion Screening Levels
 Not To Be Used for Structures Frequented or Occupied by Children
 June 2021
 (based on May 2021 USEPA Regional Screening Tables)

Note: If a screening level is exceeded and/or more than five compounds with non-cancer effects are present risk should be evaluated using NCEDEQ Risk Calculator.

CAS #	Chemical Name:	Ground Water Screening Level (GWSL) ug/L	Sub-slab and Exterior Soil Gas Screening Level (SGSL) ug/m ³	Indoor Air Screening Level (IASL) ug/m ³	Does Chemical Have a Non-Cancer Effect?
E1790678	Tributyltin Compounds				
56-35-9	Tributyltin Oxide				
10025-85-1	Trichloramine				
76-13-1	Trichloro-1,1,2-trifluoroethane, 1,1,2-	2.0E+02	4.4E+05	4.4E+03	Yes
76-03-9	Trichloroacetic Acid				
33663-50-2	Trichloroamine HCl, 2,4,6-				
634-93-5	Trichloroaniline, 2,4,6-				
87-61-6	Trichlorobenzene, 1,2,3-				
120-82-1	Trichlorobenzene, 1,2,4-	3.0E+01	1.8E+02	1.8E+00	Yes
71-55-6	Trichloroethane, 1,1,1-	6.2E+03	4.4E+05	4.4E+03	Yes
79-00-5	Trichloroethane, 1,1,2-	5.2E+00	1.8E+01	1.8E-01	Yes
79-01-6	Trichloroethylene	4.4E+00	1.8E+02	1.8E+00	Yes
75-69-4	Trichlorofluoromethane				
95-95-4	Trichlorophenol, 2,4,5-				
88-06-2	Trichlorophenol, 2,4,6-	Not Volatile	Not Volatile	4.0E+00	C
93-76-5	Trichlorophenoxyacetic Acid, 2,4,5-				
93-72-1	Trichlorophenoxypropionic acid, -2,4,5				
598-77-6	Trichloropropane, 1,1,2-				
96-18-4	Trichloropropane, 1,2,3-	1.9E+01	2.6E+01	2.6E-01	Yes
96-19-5	Trichloropropane, 1,2,3-	3.7E-01	2.6E+01	2.6E-01	Yes
1330-78-5	Tricresyl Phosphate (TCP)				
58138-08-2	Tridiphan				
121-44-8	Triethylamine	1.0E+03	6.1E+02	6.1E+00	Yes
112-27-6	Triethylene Glycol				
420-46-2	Trifluoroethane, 1,1,1-	5.6E+02	1.8E+06	1.8E+04	Yes
1582-09-8	Trifluorin				
512-56-1	Trimethyl Phosphate				
526-73-8	Trimethylbenzene, 1,2,3-	2.9E+02	5.3E+03	5.3E+01	Yes
95-36-6	Trimethylbenzene, 1,2,4-	2.1E+02	5.3E+03	5.3E+01	Yes
108-67-8	Trimethylbenzene, 1,3,5-	1.5E+02	5.3E+03	5.3E+01	Yes
25167-70-8	Trimethylpentene, 2,4,4-				
99-35-4	Trinitrobenzene, 1,3,5-				
118-96-7	Trinitrotoluene, 2,4,6-				
791-28-6	Triphenylphosphine Oxide				
13674-87-8	Tris(1,3-Dichloro-2-propyl) Phosphate				
13674-84-5	Tris(1-chloro-2-propyl)phosphate				
126-72-7	Tris(2,3-dibromopropyl)phosphate	2.1E+01	1.9E+00	1.9E-02	C
115-96-8	Tris(2-chloroethyl)phosphate				
78-42-2	Tris(2-ethylhexyl)phosphate				
7440-33-7	Tungsten				
7440-61-1	Uranium	Not Volatile	Not Volatile	3.5E-02	N
51-79-6	Urethane	Not Volatile	Not Volatile	4.2E-02	C