

# Integration of Nafion<sup>Z</sup> sample conditioner into existing FEM ozone monitor

2022 National Ambient Air Monitoring Conference Ozone Session – Wednesday August 24, 2022

> Bryan Bibeau Product Manager Teledyne-API San Diego, CA

### History

- EPA CASTNET program requested approval for addition of Nafion<sup>ż</sup> dryers to ozone analyzers to mitigate interference from ambient water vapor.
- EPA's CASTNET contractor performed testing in Gainesville, FL from October 2016 through 2017.
- Testing with dryer showed improved response
  - -less noise
  - faster response
  - less interference

# History

- CASTNET implementation approved for addition of external Nafion<sup>ż</sup> dryer.
- Requires additional vacuum pump to operate the Nafion<sup>ż</sup> dryer.
- Requires additional real estate in the shelter for mounting components.
- No update to FEM designation on analyzer since it's "outside the box"
- Up to end-user to ensure calibration checks are performed through the dryer



EPA CASTNET Photo

### Interference from air conditioner Summer in Long Island, NY

Site Report -Date&Time : 15/07/2021 00:00 00:00:00 - 12:00:00



— T\_Shelter[C] — Temp[C] — O3[ppb] — O3\_2[ppb]

# TAPI testing – Winter 21/22 – Fort Collins, CO

- Conduct testing in Fort Collins, CO during Winter of 2021/2022
- Utilize TAPI T400 instruments
- Utilize existing TAPI Nafion<sup>ż</sup> dryer product (T265, T200, etc.)
- Follow methodology of CASTNET study



# TAPI testing – Winter 21/22 – Fort Collins, CO



- Analyzer in climate-controlled environment
- Insulated 3/8" O.D. Teflon sample line to maintain ambient temp and RH
- Through-the-probe calibrations
- MODBUS TCP/IP data collection and calibrator control at 5-second rate
- Automated ZSP checks, nightly from 02:45 – 03:15, with 10minute phases at levels 0, 120 and 50 PPB



### Configuration

- Dryer purge flow at 800 cc/min for 1:1 ratio sample/purge
- Shares analyzer sample pump and flow control
- Purge air through analyzer filter (long-life 0.01-micron DFU) to extend dryer life
- Utilizes same components as TAPI T265, US EPA approved FRM ozone analyzer
- Follows methodology described in CASTNET study

## Configuration

- Control analyzer T400 with no Nafion<sup>Ż</sup> dryer
- Test analyzer T400 with Nafion<sup>Ż</sup> dryer
- T703U calibrator for daily checks for both instruments
- Campbell Scientific CR6 DAS for data collection and control
- Vaisala HMP50 for ambient relative humidity and temperature
- Thermocouple for station temperature

## Configuration

- Instruments run in parallel to establish comparison between unmodified instruments.
- Test instrument modified with internal Nafion<sup>ź</sup> dryer and run in parallel to the unmodified control instrument.
- Automated daily checks performed at zero, 70 and 120 ppb using T703U reference photometer/generator.
- All instrument data (including T703U) collected via TCP/IP MODBUS and stored at 1-minute and 60-minute resolutions.

# Design and Operating Specifications

Parameter	Specification		
Tested UV-based O3 analyzer	TAPI T400 FEM (EPA designation EQOA-0992-087)		
Analyzer sample and Dryer purge flow	800 cc/m (use same flow control)		
Nafion Dryer Assembly	TAPI 008140000, PermaPure MD-110-03-25		
Nafion tube length	~ 6'		
Nafion sample air tube ID	1/8"		
Nafion sample air residence time	$\sim 1$ second		
Purge air to sample air flow ratio	1:1		
Purge air inlet filter	Same as sample, FL-50 or 47mm Teflon		
Vacuum pump	Uses analyzer vacuum pump		

#### Factory Pneumatics, T400 base



#### Pneumatics using sample air for dryer



# Pneumatics – NafionŻ' Dryer





#### **Calibration Checks**



- Red bar indicates removal of NafionŻ Dryer from T400 B
  - No calibration performed, i.e. same slopes/offsets throughout this period

### **Calibration Checks**

T4	00A_Erro	ror T400B_Error				
ZERO		PREC	ZERO		PREC	
(ppb)	SPAN (%)	(%)	(ppb)	SPAN (%)	(%)	1
-1.5	-0.7	-3.0	-1.2	-2.0	-3.2	Nafion
-1.0	-1.2	-3.4	0.5	-1.3	-2.1	
-1.8	-1.1	-1.2	0.2	-1.7	-0.2	
-1.1	-2.2	-3.6	-0.3	-1.9	-1.2	
-1.1	-1.2	-1.8	0.6	-1.1	-0.3	
-0.8	-6.5	-7.5	-2.3	-7.9	-9.4	Factory
-0.9	-3.2	-4.4	-1.8	-4.2	-6.7	
-0.8	-2.2	-2.8	-0.2	-2.5	-1.5	
-0.9	-1.4	-1.5	-2.2	-2.7	-4.4	

- Checks failed first day after removing Nafion<sup>Ż</sup> dryer while system equilibrated
- No change in offset/slope before/after

### Linearity Testing

- T400A (Control) vs T400B (with Nafion<sup>TM</sup>)
- Conducted over normal ambient range
- Simultaneous sampling from T703U calibration standard



**Linearity Test** 

#### 3-day Ambient Comparison, Factory



#### 3-day Ambient Comparison, Modified



### Experimental High Dewpoint Testing



#### Dec 24, 2021: Woodsmoke Event

Local Woodsmoke Interference 12/24/2021



### **US EPA Compliance Requirements**

- Submit modification request to the existing US EPA FEM Method Designation
- Request should include written rationale for modification request
- Include specific details on the intended changes
  - Hardware
  - Location
  - Flow rates
  - Pneumatics
- Submit side-by-side data from an existing T400 FEM designated instrument and proposed Nafion<sup>Ż</sup> dryer assembly
- Include specific details on expected service life, recommended cleaning, maintenance and/or replacement intervals

### **Conclusion & Questions**

- Placement of Nafion<sup>Ż</sup> dryer does not significantly affect calibration of instrument as evidenced by automated nightly calibration checks
- Nafion<sup>ż</sup> dryer does not scrub ambient ozone
- Placement of Nafion<sup>ż</sup> dryer does improve response during periods of rapid RH change and/or localized woodsmoke
- Exact interferences around woodsmoke require further investigation
- Expectation is that Nafion<sup>2</sup> will mitigate H<sub>2</sub>0 interference based on previous studies by others

### Updated Designation EQOA-0992-087

Teledyne Advanced Pollution Instrumentation, Model 400E or T400 or N400; Advanced Pollution Instrumentation, Model 400/400A; Teledyne Monitor Labs sensor-e<sup>™</sup> Model TML-10 Ozone Analyzers; or recordum airpointer® system module 801-004000; Automated Equivalent Method: EQOA-0992-087

"Teledyne Advanced Pollution Instrumentation. Model 400E or T400 or N400; Advanced Pollution Instrumentation, Model 400 or 400A; or Teledyne Monitor Labs sensor- $e^{TM}$  Model TML-10 Ozone Analyzer" operated on any full scale range between 0-100 ppb<sup>1</sup> and 0-1000 ppb, with any range mode (Single, Dual, or AutoRange), at any ambient temperature in the range of 5°C to 40°C (0°C -45°C for the N400), and with a TFE filter or a Kynar® DFU. **Models 400E, T400, N400 and TML-10:** operated with a sample flow rate of  $800 \pm 80$  cm<sup>3</sup>/min (measured volumetrically at actual T & P conditions), with the dilution factor set to 1, with Dynamic Zero ON or OFF, with Dynamic Span OFF, with Temp/Press compensation ON, and with or without any of the following options: Internal or external sample pump, Sample/Cal valve option, Internal Zero/Span (IZS), Rack mount with or without slides, analog input option, 4-20 mA isolated current loop output; the NumaView<sup>TM</sup> software.<sup>2</sup> **Models 400E, T400 and N400**: operated with or without the internal sample dryer assembly (TAPI 008140000).

#### References

- Recommendations for Nationwide Approval of NafionŻ Dryers Upstream of UV-Absorption Ozone Analyzers, November 16, 2020:
  - https://cfpub.epa.gov/si/si\_public\_file\_download.cfm?p\_downlo ad\_id=541625&Lab=CEMM
- Use of Nafion Dryers for Ultraviolet (UV) Ozone Analyzers, February 19, 2021: https://www.epa.gov/sites/default/files/2021-03/documents/nafion dryer memo- pdf.pdf





www.teledyne-api.com



Teledyne API



